

Anesthetics

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.


Abstract

This application brief demonstrates analysis of anesthetics.


Introduction

The compounds used in this study are –


Compounds	Sample Concentration
Lidocaine	0.04 mg/mL
Prednisolone	0.01 mg/mL
Naproxen	0.003 mg/mL
Amitriptyline	0.01 mg/mL
Ibuprofen	0.04 mg/mL




Prednisolone




Lidocaine



Naproxen



Amitriptyline



Ibuprofen

Experimental

Conditions

Column:	Xterra MS C ₁₈ 4.6 x 20 mm, 2.5 μm
Part number:	186001889
Mobile phase A:	H ₂ O
Mobile phase B:	ACN

Mobile phase C: 100 mM CH₃COONH₄, pH 5.0

Flow rate: 3.0 mL/min

Injection volume: 20 µL

Sample concentration: As indicated

Temperature: Ambient


Detection: UV @ 238 nm

Instrument: Alliance 2695, 2996 PDA

Gradient

Time (min)	Profile		
	%A	%B	%C
0.0	85	5	10
4.0	20	70	10

Results and Discussion



Featured Products

· [Alliance HPLC System <https://www.waters.com/534293>](https://www.waters.com/534293)

WA20738.009, June 2002



© 2021 Waters Corporation. All Rights Reserved.