


## Cytochrome C Tryptic Digest – pH 10.0, XTerra RP<sub>18</sub>

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---

### Abstract

This application brief highlights the analysis of cytochrome C using XTerra RP<sub>18</sub> columns.

---

## Introduction

Cytochrome C has been studied in this application brief.

---

## Experimental

### HPLC Conditions

|                   |                                                              |
|-------------------|--------------------------------------------------------------|
| Column:           | XTerra RP <sub>18</sub> 4.6 x 50 mm, 3.5 µm (p/n: 186000434) |
| Mobile phase A:   | H <sub>2</sub> O                                             |
| Mobile phase B:   | ACN                                                          |
| Mobile phase C:   | 100 mM HCOOH, pH 10.0                                        |
| Flow rate:        | 0.75 mL/min, 0.2 mL/min to MS                                |
| Injection volume: | 50 µL (25 µg)                                                |
| Temperature:      | Ambient                                                      |
| Detection:        | MS ESI+                                                      |
| Instrument:       | Alliance 2790, Micromass Quattro Ultima                      |

### Gradient


| Time<br>(min) | Profile |    |    |
|---------------|---------|----|----|
|               | %A      | %B | %C |
| 0.0           | 90      | 0  | 10 |
| 45.0          | 50      | 40 | 10 |
| 45.0          | 10      | 80 | 10 |
| 50.0          | 10      | 80 | 10 |

## MS Conditions

|                       |              |
|-----------------------|--------------|
| MS sytem:             | Micromass ZQ |
| Ion source:           | ESI+         |
| Capillary:            | 3.5 kV       |
| Cone:                 | 30 V         |
| Extractor:            | 3.0          |
| Source temp.:         | 150 °C       |
| Desolvation temp.:    | 400 °C       |
| Cone gas flow:        | 60 L/hr      |
| Desolvation gas flow: | 550 L/hr     |

---

## Results and Discussion



---

## Featured Products

Alliance HPLC <<https://www.waters.com/514248>>

WA20738.030, June 2002

© 2021 Waters Corporation. All Rights Reserved.