

Nota applicativa

## Pheniramine

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---


### Abstract

This application brief highlights the analysis of pheniramine using XTerra RP<sub>18</sub> columns.

---

## Introduction

Pheniramine has been analyzed in this study.



## Pheniramine

---

## Experimental

### HPLC Conditions

Column:	XTerra RP <sub>18</sub> 4.6 x 150 mm, 5 μm (p/n: 186000492)
Mobile phase:	At pH 3.0: H <sub>2</sub> O/ACN/100 mM, NH <sub>4</sub> COOH, pH 3.0, 55:35:10 At pH 7.0: H <sub>2</sub> O/ACN/100 mM, NH <sub>4</sub> HCO <sub>3</sub> , pH 7.0, 30:60:10 At pH 10.0: H <sub>2</sub> O/ACN/100 mM, NH <sub>4</sub> HCO <sub>3</sub> , pH 10.0, 60:30:10
Flow rate:	1.0 mL/min

Injection volume: 5  $\mu$ L of 250  $\mu$ g/mL

Temperature: 30  $^{\circ}$ C


Detection: UV @ 265 nm

Instrument: Alliance 2695, 2996 PDA

Mobile phase pH	USP Tailing
3.0	1.19
7.0	1.30
10.0	1.23

---

## Results and Discussion



---

## Featured Products

Alliance HPLC <<https://www.waters.com/514248>>

WA20738.087, June 2002

© 2021 Waters Corporation. All Rights Reserved.