


Note d'application

RP-HPLC Method For Quality Control Of Synthetic Oligonucleotides Using XTerra MS C₁₈ Columns

Waters Corporation



Abstract

This application note demonstrates RP-HPLC method for quality control of synthetic oligonucleotides using XTerra MS C18 Columns.

Introduction

Oligonucleotide synthesis yield

Although state-of-the-art oligonucleotide synthesis is a fast and reliable process, the yield of the target product is limited. The full-length oligonucleotide (N) is typically contaminated with shorter failure products called N-1, N-2, N-3... Oligonucleotide purity decreases with the complexity of synthesis. The typical purity of crude 25mer product is ~75% while the purity of a 60mer product is ~20%. Occasionally the synthesis or post synthesis deprotection result in a product of inferior purity, which may compromise quality of the polymerase chain reaction (PCR) or other molecular biology assays.

Experimental

Preparation of 0.1 M TEAA buffer

Dissolve 5.6 mm glacial acetic acid in ~950 mL of water. While mixing add 13.86 mL of TEA. Adjust pH with diluted acetic acid to ~7 and adjust volume to 1L with water.

Preparation of 16.3 mM TEA - 400 mM HFIP buffer

Dissolve 42.1 mL of HFIP in ~950 mL of water. While mixing vigorously add 2.28 mL of TEA. Adjust volume to 1L with water. The pH of solution should be close to 7.9.

Results and Discussion

XTerra Columns for efficient oligonucleotide separations

XTerra MS C₁₈ Columns are packed with porous 2.5 µm, hybrid particles. The sorbent has extended stability at temperatures and pH's typically used for oligonucleotide separations (50–60 °C; pH 7–9). The column

yields high resolution efficiency required for successful analysis of oligonucleotides (Figure 2). Volatile ion-pairing mobile phases are utilized to separate oligonucleotides according to their length. The resolution performance rivals that seen with capillary gel electrophoresis (Figure 1). With XTerra MS C₁₈ Columns, baseline resolution of N from N-1 oligonucleotides is typically achieved for <30mer. Good resolution can be achieved for larger oligonucleotides (e.g. 60mer or larger: see (Figures 2 & 3)).

Triethylammonium acetate (TEAA) is an inexpensive volatile ion-pairing agent used for routine oligonucleotide analysis (Figure 2) and purification. Alternatively, triethylammonium-hexafluoroisopropanol buffer (TEA-HFIP) can be used for more efficient separation of long oligonucleotides (>45mer, Figure 3), and for applications using mass spectrometry detection.




Figure 1. Separation of 6-30mer oligodeoxythymidines by CGE and RP-HPLC using XTerra MS C₁₈.

HPLC system: Alliance® 2690 Separations Module
Column: XTerra®MS C₁₈, 2.5 µm, 4.6 x 50 mm
Part Number: 186000602
Flow rate: 1 ml/min
Solvent A: 5 % ACN in 0.1 M TEAA, pH 7
Solvent B: 20 % ACN in 0.1 M TEAA, pH 7

Gradient: From 9.3 % B to 44 % B in 20.8 min
Injection volume: 5 µl (1 nmole of oligonucleotide)
Column temp.: 60 °C
Detection: UV @ 260 nm





Figure 2. RP-HPLC quality control of synthetic oligonucleotides. Purity was measured as the % of peak area at UV_{260 nm}.



HPLC system: Waters CapLC® System
Column: XTerra®MS C₁₈, 2.5 µm, 1.0 x 50 mm
Part Number: 186000979
Flow rate: 23.6 µl/min
Solvent A: 16.3 mM TEA, 400 mM HFIP aqueous buffer, pH 7.9
Solvent B: MeOH
Gradient: 0.55 minutes, liner gradient 21–26.5 % B
Injection volume: 0.5 µl
Column temp.: 50 °C
Detection: UV @ 260 nm

Figure 3. RP-HPLC separation of 30–35mer and 57–60mer heterooligonucleotides.

Featured Products

Alliance HPLC System <<https://www.waters.com/534293>>

