

Nota applicativa

## Naproxen – Symmetry C<sub>8</sub>

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---


### Abstract

This application brief highlights the analysis of naproxen using Symmetry C<sub>8</sub>.

---

## Introduction

Compound used in this application brief is naproxen.



### 1. Naproxen

---


## Experimental

### HPLC Method

Column:	Symmetry C <sub>8</sub> , 3.9 x 150 mm, 5 μm
Guard column:	Sentry Guard Column 3.9 mm x 20 mm, 5 μm
Part numbers:	Column - WAT046970, Guard - WAT054250
Mobile phase:	Acetonitrile/water/glacial acetic acid 50491
Flow rate:	1.0 mL/min
Injection volume:	5 μL of 100 μg/mL sample
Detection:	UV @ 254 nm
USP Tailing Factors:	1.115

---

## Results and Discussion



---

## Featured Products

Alliance HPLC <<https://www.waters.com/514248>>

WA31763.110, June 2003

© 2021 Waters Corporation. All Rights Reserved.