

Note d'application

Acetaminophen and Hydrocodone Bitartrate

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.


Abstract

This application brief highlights the analysis of acetaminophen and hydrocodone bitartrate using Symmetry C₁₈ Columns.


Introduction

Compounds used in this study are:

1. Acetaminophen
2. Hydrocodone Bitartrate



1. Acetaminophen




2. Hydrocodone Bitartrate

Experimental

HPLC Method

Column:	Symmetry C ₁₈ , 3.9 x 150 mm, 5 µm (p/n: WAT046980)
Mobile phase:	50 mM Ammonium phosphate, pH 5.8/acetonitrile 85:15
Flow rate:	1.0 mL/min
Injection volume:	20 µL of 775 µg/mL acetaminophen and 8 µg hydrocodone bitartrate sample
Detection:	UV @ 214 nm

Results and Discussion



Featured Products

WA31763.12, June 2003

