# Waters™

응용 자료

# Propranolol in Serum

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

#### Abstract

This application brief highlights the analysis of propranolol in serum using SymmetryShield Columns.

### Introduction

The compounds analyzed in this study are:

- 1. Propranolol
- 2. Butyl paraben (I.S.)

#### 1. Propranolol

## 2. Butyl paraben (I.S.)

# Experimental

#### **HPLC** Method


Column: Symmetry Shield RP8, 3.9 x 150 mm, 5  $\mu$ m

Guard column: Symmetry guard column, 3.9 x 20 mm, 5  $\mu$ m

Part numbers:	Column - WAT200655, Guard - WAT200675	
Mobile phase:	20 mM potassium phosphate, pH 7/methano 38:62	
Flow rate:	1.0 mL/min	
Injection volume:	20 $\mu$ L of reconstituted porcine serum extract	
Temperature:	30 °C	
Detection:	UV @ 254 nm	


# Oasis® HLB Extraction Method

Oasis® HLB 1 cc/30mg Extraction Cartridge Part Number WAT094225



Compound	Concentration µg/mL	% Recovery	%RSD (n=6)
Propranolol	0.800	85.9%	2.1%
	4.00	96.4%	2.4%

## Chromatogram of Serum Extracts: A) Blank B) Spiked Sample



## Featured Products

WA31763.135, June 2003

© 2021 Waters Corporation. All Rights Reserved.