

## Speed Mix - 4.6 x 20 mm Intelligent Speed Separation

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

### Abstract

This application brief demonstrates analysis of speed mix using 4.6 x 20 mm Intelligent Speed Separation


---

Column.


## Introduction

The compounds used in this study are –


1. Theophylline
2. 2-Acetylfuran
3. Acetanilide
4. Acetophenone
5. Propiophenone
6. Butyrophenone
7. Benzophenone
8. Valerophenone
9. Hexanophenone
10. Heptanophenone
11. Octanophenone




**1. Theophylline**




**2. 2-Acetylfuran**




**3. Acetanilide**




**4. Acephenone**




**5. Propiophenone**




**6. Butyrophenone**




**7. Benzophenone**




**8. Valerophenone**



**9. Hexanophenone**



**10. Heptanophenone**



**11. Octanophenone**

## Experimental

### HPLC Method

Column:

Symmetry C<sub>18</sub>, 4.6 x 20 mm, 3.5 μm IS

Symmetry Shield RP<sub>18</sub>, 4.6 x 20 mm, 3.5 μm IS

Part numbers:

Symmetry - 186002090

Symmetry Shield - 186002092

Mobile phase A:

0.1% HCOOH in Water

Mobile phase B: 0.1% HCOOH in ACN

Flow rate: 3.0 mL/min

Injection volume: 10  $\mu$ L


Temperature: 30  $^{\circ}$ C


Detection: UV @ 254 nm

## Gradient

Time (min)	Profile	
	%A	%B
0	100	0
4	0	100

## Results and Discussion





## Featured Products

WA31763.148, June 2003



© 2021 Waters Corporation. All Rights Reserved.