

Tamoxifen – Isolation of Impurities, Transfer from Analytical to Prep

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.


Abstract

This application brief highlights the Tamoxifen impurities isolation and transfer from analytical to prep

using Symmetry and SymmetryPrep Columns.

Introduction

Compound used in this study is Tamoxifen Citrate.





Experimental

HPLC Method

Columns:	Symmetry C ₁₈ , 3.0 x 150 mm, 5 μm (p/n:186000109) SymmetryPrep C ₁₈ , 7.8 x 300 mm, 7 μm (p/n: WAT066235)
Mobile phase:	Acetonitrile/50 mM potassium phosphate buffer, pH 3.0 40:60
Flow rate:	A. 0.4 mL/min B. 5.6 mL/min
Injection volume:	5 mg/mL tamoxifen citrate A. 15 mg B. 101 mg
Detection:	UV @ 254 nm

Results and Discussion



Featured Products

WA31763.163, June 2003

