


應用手冊

# BSA Tryptic Digest - Flow Rate Effects

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

This application brief demonstrates analysis of BSA tryptic digest.

#### Introduction

The compound analysed in this study is BSA Tryptic Digest.

# Experimental

#### **HPLC** Method

Column: Symmetry 300  $C_{18}$ , 4.6 x 50 mm, 5  $\mu m$ 

Part number: WAT106209

Mobile phase A: 0.1% TFA in water

Mobile phase B: 0.1% TFA in acetonitrile

Injection volume: 20  $\mu$ L


Temperature: 35 °C

Detection: UV @ 214 nm

## Gradient

Time (min)	Profile	
	%A	%B
0	100	0
45	70	30

# Results and Discussion



## **Featured Products**

WA31763.41, June 2003

 $\wedge$ 

© 2021 Waters Corporation. All Rights Reserved.