Waters™

Applikationsbericht

Cytochromo o Tryptic Digost

Reproducibility Study		
Waters Corporation		
This is an Application Brief and does not contain a detailed Experimental section.		
Abstract		
This application brief details on cytochron	me c tryptic digest reproducibility using Symmetry Columns.	
Introduction		
The compound used in this study is cytoc	chrome c tryptic digest.	
Experimental		
HPLC Method		
Column:	Symmetry 300 C ₁₈ , 3.9 x 150 mm, 5 μm	

Part number: WAT106154

Mobile phase A: 0.05% TFA in Water

Mobile phase B: 0.1% TFA in Acetonitrile

Flow rate: 1.0 mL/min

Injection volume: 20 µL


Temperature: 35 °C

Detection: UV @ 220 nm

Gradient

Time	Prof	file
(min)	%A	%B
0	100	0
45	72	28

Results and Discussion



Featured Products

WA31763.69, June 2003

© 2022 Waters Corporation. All Rights Reserved.