

Note d'application

## Dihydroergotamine Mesylate

---

Waters Corporation

This is an Application Brief and does not contain a detailed Experimental section.

---

### Abstract


This application brief demonstrates the analysis of dihydroergotamine mesylate using Symmetry Columns.

---

### Introduction

The compound analyzed in this study is dihydroergotamine mesylate.

---



## 1. Dihydroergotamine mesylate

---

### Experimental

#### HPLC Method

Column:	Symmetry C <sub>18</sub> , 3.9 x 150 mm, 5 μm
Part number:	WAT046980
Mobile phase:	50 mM ammonium phosphate, pH 5.8/Acetonitrile 35:65

---


Flow rate:	1.0 mL/min
Injection volume:	10 µL of 100 µg/mL dihydroergotamine mesylate
Detection:	UV @ 281 nm

## USP Tailing Factors

1. 1.1

---

## Results and Discussion



---

## Featured Products

WA31763.74, June 2003

© 2022 Waters Corporation. All Rights Reserved.