

Diltiazem HCL- Isolation of Impurities, Transfer from Analytical to Prep Columns

Waters Corporation


This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This application brief demonstrates the analysis of Diltiazem HCL impurities using Symmetry Columns.

Introduction

The compound analyzed in this study is diltiazem HCL.



1. Diltiazem Hydrochloride

Experimental




HPLC Method

Column:

A. Symmetry C_{18} , 3.9 x 150 mm, 5 μm

	B. Symmetry Prep C ₁₈ , 7.8 x 150 mm, 7 µm
	C. Symmetry Prep C ₁₈ , 19 x 150 mm, 7 µm
Part numbers:	A. WAT046980
	B. WAT066288
	C. WAT066240
Mobile phase:	Acetonitrile/0.1% TFA in water 30:70
Flow rates:	A. 0.7 mL/min B. 2.8 mL/min C. 16.6 mL/min
Injection volume:	Diltiazem: A. 0.5 mg B. 2.0 mg C. 11.9 mg
Detection:	UV @ 280 nm

Results and Discussion



Featured Products

WA31763.76, June 2003

© 2022 Waters Corporation. All Rights Reserved.