

## Analysis of Tricyclic Antidepressants Using XBridge C<sub>18</sub>, pH 7.0

---

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---

### Abstract


This application brief highlights the analysis of tricyclic antidepressants using XBridge C<sub>18</sub> Columns at pH 7.0.

---


## Introduction

Tricyclic antidepressants are analyzed in this application brief.


### Compounds




1. Nortriptyline



3. Amitriptyline



2. Imipramine



4. Clomipramine

---

## Experimental

### Test Conditions


Column:

XBridge C<sub>18</sub>, 5 µm, 4.6 x 150 mm p/n: 186003116

Isocratic Mobile Phase:	20 mM sodium phosphate buffer, pH 7.0/acetonitrile/methanol (30:35:35)
Flow Rate:	1.5 mL/min
Injection Volume:	10 µL
Sample Concentration:	20 µg/mL in water
Temperature:	25 °C
Detection:	UV @ 254 nm
Instrument:	Alliance 2695 with 2996 PDA

---

## Results and Discussion



---

## Featured Products

Alliance HPLC <<https://www.waters.com/514248>>

WA60167, April 2007

