# Waters™



# High Definition Mass Spectrometry as a Tool for Structural Investigation of High *m/z* Ion Species

Iain Campuzano, Kevin Giles

Waters Corporation

This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

Here we present the use of high-efficiency IMS (Triwave technology) on a SYNAPT High Definition Mass Spectrometry (HDMS) System for analysis of high m/z caesium iodide clusters over the m/z range 1,000–20,000. This demonstrates the utility of the SYNAPT HDMS System for the mass measurement of high m/z species, such as in the analysis of non-covalent protein complexes.

### Introduction

Over the past 10 years, interest in high mass non-covalent analysis has increased due to the ability of the current mass spectrometers and electrospray sources to preserve the non-covalent interactions, allowing one to analyze compounds in their native conformation and stoichiometry. The transfer of non-covalently associated complexes from solution to the gas phase using electrospray ionization generally results in the formation of ions possessing

relatively few charges. As a result, these species appear high on the m/z scale, making Time-of-Flight mass spectrometry ideal for their mass analysis. The utility of ion mobility spectrometry (IMS) in probing the structures of relatively large complexes has been highlighted previously.<sup>1</sup>

Here we present the use of high-efficiency IMS (Triwave technology) on a SYNAPT High Definition Mass Spectrometry (HDMS) System for analysis of high m/z caesium iodide clusters over the m/z range 1,000–20,000. This demonstrates the utility of the SYNAPT HDMS System for the mass measurement of high m/z species, such as in the analysis of non-covalent protein complexes.

# Experimental

The instrument used in this study was a SYNAPT HDMS System, which combines high-efficiency ion mobility based measurements and separations with a hybrid quadrupole orthogonal acceleration Time-of-Flight (oa-ToF) mass spectrometer, Figure 1.<sup>2</sup> Samples were introduced with a borosilcate-glass nano electrospray-spray tip and sampled into the vacuum system. The ions pass through a quadrupole mass filter to the enabling Triwave device. Triwave consists of three travelling wave (T-Wave) ion guides. The TRAP T-Wave accumulates ions (with high efficiency), after which these ions are released as discrete packets into the IMS T-Wave, where the ion mobility separation of ions is performed. The TRANSFER T-Wave is used to deliver the ion mobility-separated ions into the oa-ToF analyzer. Each IMS separation was 51 ms long and the ions were released from the TRAP T-wave in 500 µs wide packets. The gas pressure in the TRAP and TRANSFER T-Wave regions was 0.07 mbar (Argon) and the pressure in the IMS T-Wave was 0.5 mbar (Nitrogen). The traveling wave used in the IMS T-Wave for ion mobility separation was operated at a velocity of 250 m/sec. The wave amplitude was ramped from 0 to 30 V over the period of the mobility separation for optimum performance over the large *m/z* range used (*m/z* 1,000 to 32,000).

#### **MS Conditions**

MS system: SYNAPT HDMS System

Ionization mode: nanoESI positive

Capillary voltage: 1000 V

Cone voltage: 150 V

Source temp.: 40 °C

Acquisition range: 1,000 to 32,000 m/z

IMS T-wave ramp: 0 to 3 0V over IMS experiment

IMS T-wave speed: 250 m/sec

IMS pressure: 0.5 mbar (nitrogen)




Figure 1. Schematic of the SYNAPT HDMS System.

# Results and Discussion

Upon MS acquisition of a concentrated caesium iodide solution, intense ion clusters can be observed as high as

m/z 20,000, with each cluster's composition based on the formula Cs(n+1)In. From the mass spectrum generated it is evident that a number of overlapping series, differing in charge state and intensity profile, are present over the m/z scale (Figure 2). However, the HDMS (IMS/MS) analysis clearly illustrates discrete distributions, which are related by their m/z and drift-time, as shown in Figure 3.




Figure 2. Mass spectrum of CsI (m/z 1,000 to 20,000).




Figure 3. HDMS analysis: Drift-time vs m/z plot of CsI (m/z 1,000 to 32,000).

The drift-times of the cluster ions seem to increase monotonically with increasing m/z values for the different charge state species, although with increasing charge state series up to +4, distinct m/z stability regions become clear, as shown in Figure 4.




Figure 4. Charge state distributions of Csl.

# Conclusion

- · HDMS provides greater definition of even relatively simple samples, such as inorganic CsI clusters
- · High mass analysis of CsI clusters up to m/z 20,000 has been shown
- Csl clusters have been successfully separated into five charge states series, demonstrating the ability of HDMS (high-efficiency IMS combined with oa-ToF mass spectrometry) for differentiating high mw species based on their mobility
- Csl can be used to tune, optimize, and calibrate the SYNAPT HDMS System over a wide m/z range prior to high mass acquisitions being carried out

# References

- 1. Ruotolo, Giles, Campuzano, Sandercock, Bateman & Robinson. Science. 310 (2005) 1658–1661.
- 2. Pringle, Giles, Wildgoose, Williams, Slade, Thalassinos, Bateman, Bowers, Scrivens. *Int. J. Mass Spectrom.*, 261 (2007) 1–12.

## Featured Products

TriWave <a href="https://www.waters.com/134663660">https://www.waters.com/134663660</a>

720002632, May 2008

© 2022 Waters Corporation. All Rights Reserved.

Terms of Use Privacy Trademarks Sitemap Careers Cookie Cookie 偏好設定