# Waters™



# Analysis of Food Sugars in MILK Using ACQUITY UPLC BEH Amide Columns

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

This application brief demonstates analysis of food sugars in milk.

## Introduction

Compounds analysed in this application include:

- 1. p-Toluamide
- 2. Fructose
- 3. Glucose
- 4. Sucrose
- 5. Maltose
- 6. Lactose

#### Structures

# Experimental

Flow Rate:


## **Chromatographic Conditions**

Column:	ACQUITY UPLC BEH Amide 2.1 x 100 mm, 1.7 μm
Part Number:	186004801
Mobile Phase A:	80/20 MeCN/H <sub>2</sub> O with 0.2% triethylamine [TEA]
Mobile Phase B:	30/70 MeCN/H <sub>2</sub> O with 0.2% triethylamine [TEA]

0.13 mL/min

Flow Profile:	90% A/10% B (75% MeCN with 0.2% TEA)
Injection Volume:	1.3 μL (PLNO)
Sample Concentration:	Standards at 1 mg/mL each
Sample Diluent:	50/50 MeCN/H <sub>2</sub> O
Column Temperature:	35°C
Strong Needle Wash:	20/80 MeCN/H <sub>2</sub> O (800 μL)
Weak Needle Wash:	75/25 MeCN/H <sub>2</sub> O (500 μL)
Seal Wash:	50/50 MeCN/H <sub>2</sub> O
Instrument:	Waters ACQUITY UPLC with ELSD
ELSD Conditions	
Gain:	200
Pressure:	40 psi
Drift Tube Temperature:	40 °C
Nebulizer:	Cooling
Data Rate:	10 pps
Filter Time Constant:	Normal

#### Results and Discussion



#### Featured Products

- ACQUITY UPLC System <a href="https://www.waters.com/514207">https://www.waters.com/514207</a>
- ACQUITY UPLC ELS Detector <a href="https://www.waters.com/514219">https://www.waters.com/514219</a>

WA60118, September 2009



©2019 Waters Corporation. All Rights Reserved.