# Waters™

# アプリケーションノート

# Analysis of Food Sugars/Saccharides in Beer Using ACQUITY UPLC BEH Amide Columns

**Waters Corporation** 

This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

This application brief highlights the analysis of food sugars/saccharides in beer using ACQUITY UPLC BEH Amide Columns.

# Introduction

#### **Structures**

# Experimental

# **Chromatographic Conditions**

Column: ACQUITY UPLC BEH Amide 2.1 x

100 mm, 1.7  $\mu$ m

Part Number: 186004801

Mobile Phase A: 80/20 MeCN/H<sub>2</sub>O with 0.2%

triethylamine [TEA]

Mobile Phase B: 30/70 MeCN/H<sub>2</sub>O with 0.2%

triethylamine [TEA]

Flow Rate: 0.13 mL/min

Gradient: 10 minute gradient, 80%-50%

MeCN (w/0.2% TEA) with 25 minute re-equilibration

Injection Volume: 1.3 μL (PLNO)

Sample Concentration: Standards at 1mg/mL, beer at

100% (No dilution)

Sample Diluent: 50/50 MeCN/H<sub>2</sub>O

Column Temperature: 35 °C

Strong Needle Wash: 20/80 MeCN/H<sub>2</sub>O (800 µL)

Weak Needle Wash:  $75/25 \text{ MeCN/H}_2\text{O} (500 \text{ }\mu\text{L})$ 

Seal Wash: 50/50 MeCN/H<sub>2</sub>O

Instrument: Waters ACQUITY UPLC with

ELSD

# Gradient

Time	Profile		
(min)	%A	%B	
0.00	90.00	10.00	
10.00	30.00	70.00	
10.01	90.00	10.00	
35.00	90.00	10.00	

# **ELSD Conditions**


Gain: 200

Pressure: 40 psi

Drift Tube Temperature: 40 °C

Nebulizer:	Cooling
Data Rate:	10 pps
Filter Time Constant:	Normal
Results and Discussion	
The compounds analysed in this study are:	
1. p-Toluamide	
2. Fructose	
3. Glucose	
4. Sucrose	
5. Maltose	
6. Lactose	
7. Maltotriose	
8. Maltotetraose	
9. Maltopentaose	
10. Maltohexahose	

11. Maltoheptaose



# **Featured Products**

ACQUITY UPLC ELS Detector <a href="https://www.waters.com/514219">https://www.waters.com/514219</a>

WA60125, October 2009

© 2022 Waters Corporation. All Rights Reserved.				