Waters™



Analysis of Cellulosic Hydrolysates Using ACQUITY UPLC BEH Amide Columns

Waters Corporation

This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This application brief highlights the analysis of cellulosic hydrolysates using ACQUITY UPLC BEH Amide Columns.

Introduction

The compounds analysed in this study are:

- 1. Xylose
- 2. Fructose
- 3. Mannose
- 4. Glucose

- 5. Sucrose
- 6. Cellobiose
- 7. Melezitose
- 8. Raffinose
- 9. Maltotriose
- 10. Maltotetraose
- 11. Maltopentaose
- 12. Maltohexaose
- 13. Maltoheptaose

Experimental

Chromatographic Conditions

Column:

ACQUITY UPLC BEH Amide 2.1 x 100 mm, 1.7 μm

Part Number:	186004801
Mobile Phase A:	80/20 MeCN/H ₂ O with 0.2% triethylamine [TEA]
Mobile Phase B:	30/70 MeCN/H ₂ O with 0.2% triethylamine [TEA]
Flow Rate:	0.12 mL/min
Gradient:	10 minute gradient, 80%-50% MeCN (w/0.2% TEA) with 30 minute re-equilibration
Injection Volume:	1.3 μL (PLNO)
Sample Concentration:	1 mg/mL each
Sample Diluent:	50/50 MeCN/H ₂ O
Column Temperature:	35 °C
Strong Needle Wash:	20/80 MeCN/H ₂ O (800 μL)
Weak Needle Wash:	75/25 MeCN/H ₂ O (500 μL)
Seal Wash:	50/50 MeCN/H ₂ O
Instrument:	Waters ACQUITY UPLC with ELSD


Gradient

Time	Profile	
(min)	%A	%B
0.00	100.00	0.00
10.00	60.00	40.00
10.01	100.00	0.00
40.00	100.00	0.00

ELSD Conditions

Gain:	200
Pressure:	40 psi
Drift Tube Temperature:	40 °C
Nebulizer:	Cooling
Data Rate:	10 pps
Filter Time Constant:	Normal

Results and Discussion



Featured Products

ACQUITY UPLC ELS Detector https://www.waters.com/514219

WA60127, October 2009

© 2022 Waters Corporation. All Rights Reserved.