Waters™

Note d'application

Analysis of Stevia Related Compounds Using ACQUITY UPLC BEH Amide Columns

Waters	Corr	oratio	n
vvalcis	\circ	JOIALIC	/ I I

This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This application brief highlights the analysis of stevia related compounds using ACQUITY UPLC BEH Amide Columns.

Introduction

The compounds used in this study are:

- 1. Steviol
- 2. Stevioside
- 3. Rebaudioside A
- 4. Rebaudioside C

wa60129-f2

Experimental

Chromatographic Conditions

Column:	ACQUITY UPLC BEH Amide 2.1 x 100 mm, 1.7 μm
Part Number:	186004801
Mobile Phase A:	80/20 MeCN/H ₂ O with 0.2% triethylamine [TEA]
Mobile Phase B:	30/70 MeCN/H ₂ O with 0.2% triethylamine [TEA]
Flow Rate:	0.20 mL/min
Injection Volume:	1.3 μL (PLNO)
Sample Concentration:	5 mg/mL
Sample Diluent:	50/50 MeCN/H ₂ O
Column Temperature:	35 °C
Strong Needle Wash:	20/80 MeCN/H ₂ O (800 μL)
Weak Needle Wash:	75/25 MeCN/H ₂ O (500 µL)
Seal Wash:	50/50 MeCN/H ₂ O
Instrument:	Waters ACQUITY UPLC with

ELSD Conditions

Gain: 200

Pressure: 40 psi


Drift Tube Temperature: 40 °C

Nebulizer: Cooling

Data Rate: 10 pps

Filter Time Constant: Normal

Results and Discussion



Featured Products
ACQUITY UPLC ELS Detector https://www.waters.com/514219
WA60128, October 2009
© 2022 Waters Corporation. All Rights Reserved.