## Waters™

Note d'application

# Multi-Residue Analysis of Pesticides in Avocado Using AOAC QuEChERS Method by UPLC-MS/MS

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

#### **Abstract**

This application brief demonstrates multi-residue analysis of pesticides in avocado using AOAC QuEChERS method by UPLC-MS/MS.

### Experimental

#### **Extraction Procedure**

- 1. Add 15 mL 1% acetic acid in acetonitrile into the 50 mL DisQuE extraction tube.
- 2. Add 15 g of homogenized sample into the 50 mL tube.
- 3. Add any internal standards and standard mixture.
- 4. Shake vigorously for 1 minute and centrifuge > 1500 rcf for 5 minute.
- 5. Transfer 1 mL of the acetonitrile extract into the 2 mL clean-up tube containing 50 mg PSA, 150 mg MgSO  $_{4}$ , and 50 mg  $C_{18}$ .
- 6. Shake for 30 seconds and centrifuge >1500 rcf for 1 minute.
- 7. Transfer 100  $\mu$ L of final extract into a 1.5 mL centrifuge tube.
- 8. Add any post-extraction internal standards.
- 9. Dilute as needed with an appropriate buffer or solvent.
- 10. Centrifuge > 16000 rcf for 5 minutes.
- 11. Transfer to autosampler vial.

#### **Test Conditions**

LC Conditions

LC System:

Waters ACQUITY UPLC

#### LC Conditions

System

Column: ACQUITY UPLC BEH C<sub>18</sub>, 2.1 x

100 mm, 1.7  $\mu m$ 

Column Temp: 40 °C

Sample Temp: 4 °C

Flow Rate: 0.3 mL/min.

Mobile Phase A: Water + 0.1% formic acid

Mobile Phase B: Methanol + 0.1% formic acid

Injection Volume: 15  $\mu$ L, Partial loop injection

#### Gradient:

Time	Flow Rate	A%	В%
0.00	0.3	75	25
0.25	0.3	75	25
7.75	0.3	5	100
8.50	0.3	0	100
8.51	0.5	75	25
10.50	0.5	75	25

Time	Flow Rate	A%	В%
11.0	0.3	75	25

#### **MS Conditions**


Instrument: Waters ACQUITY TQ Detector

Ionization: Positive electrospray (ESI+)

Acquisition: Multiple reaction monitoring

(MRM)

## Results and Discussion



Pesticides in Avocados by UPLC-MS/MS

## Featured Products

ACQUITY UPLC System <a href="https://www.waters.com/514207">https://www.waters.com/514207</a>

720003373, March 2010

©2019 Waters Corporation. All Rights Reserved.