Waters[™]

Nota de aplicación

A Definitive Lipidomics Workflow for Human Plasma Utilizing Off-line Enrichment and Class Specific Separation of Phospholipids

Jeremy Netto, Stephen Wong, Federico Torta, Pradeep Narayanaswamy, Mark Ritchie



Waters Corporation, Centre for Life Sciences, National University of Singapore

For research use only. Not for use in diagnostic procedures.

Abstract

Consistent, reliable, and rapid identification and quantification of hundreds of lipid species can now be performed in a single run by the application of this workflow. This application note describes a tandemquadrupole-based workflow that enables a greater understanding of the global lipidome and its association with diseases.

Benefits

Large-scale quantitative and comparative lipidomic studies require the use of simple and high throughput workflows. To answer this need, a tandem quadrupole-based workflow has been developed. The benefits of this targeted lipidomics workflow include the following:

- · Simplified extraction of lipids using Ostro sample preparation chemistry
- · 96-well plate format allows for high throughput extraction with automation technologies
- · BEH HILIC chemistry provides well-defined, predictable class-based separation of lipids
- · ACQUITY UPLC System allows shorter run times per sample
- · Xevo TQ-S MS offers highly sensitive detection and quantification of lipids across a large dynamic range
- TargetLynx processing method provides rapid, automated, and quantitative data analysis of all the lipids of interest in a large sample batch

Introduction

Lipids play many important roles in maintaining homeostasis of living organisms. Lipidomics analyses could further our understanding of mechanisms of disease, including the identification of biomarkers and potential drug targets.

Biofluids such as plasma are typically complex, with large lipid diversity across many orders of concentration. These, together with the chemical complexity of lipids, present demanding analytical challenges ranging from the sample preparation stage to the analytical techniques used to identify and quantify key lipids. Today, many variations of the Bligh and Dyer method are used for total lipid extraction and purification, with equal amounts used for mass spectrometric analysis. Recent advances in lipidomics have made use of developments in chemistries and instrumentation, most notably the use of off-line enrichment or solid-phase sample preparation products^{1,2} and the coupling of UltraPerformance LC with mass spectrometry. However, there is little standardization across platforms and workflows for a complete analysis.

Presented here is a tandem quadrupole-based phospholipid analysis workflow from extraction to separation, identification and quantification of the phospholipids from a single vendor. These commercially available products packaged as a complete solution are provided to ease the strains and increase the productivity of laboratories undertaking longitudinal studies spanning hundreds of lipids over thousands of samples.



Figure 1. Workflow for the targeted analysis of phospholipids in human plasma.

Experimental

LC conditions

System:	ACQUITY UPLC
Column:	ACQUITY UPLC BEH HILIC 2.1 x 100 mm, 1.7 μm
Column temp.:	30 °C
Mobile phase A:	95:5 acetonitrile/water with 10 mM ammonium acetate, pH 8.0
Mobile phase B:	50:50 acetonitrile/water with 10 mM ammonium acetate, pH 8.0
Gradient:	0% to 20% B for 10 min
Flow rate:	500 µL/min
Injection volume:	3.0 µL, partial loop
MS conditions	

MS conditions

Mass spectrometer:	Xevo TQ-S
Ionization mode:	ESI, +/- switching
Capillary voltage:	3.8 kV (+) / 1.9 kV (-)
Desolvation temp.:	450 °C
Desolvation gas:	1000 L/h

Source temp.:	150 °C
Collision cell pressure:	3.6 x 10 ⁻³ mBar

Sample Description

Human plasma samples were obtained from the Centre for Life Sciences, National University of Singapore. The protocol described here follows a recently published application note.³

100 μ L of human plasma was loaded into each well of a Waters Ostro Sample Preparation Plate fitted onto a vacuum manifold. 800 μ L of ethanol was added to each well and mixed thoroughly by aspirating the mixture 10x using a micropipette. A vacuum of approximately 15" Hg was applied to the plate until the solvent was completely drained. This step was repeated with another 800 μ L of ethanol with the total fraction collected labelled as the "flow through."

800 μ L of elution solvent (4.5:4.5:1.0 chloroform/methanol/triethylamine) was added to each well, and the fraction was collected under 15" Hg vacuum as the "eluate." This step was repeated, bringing the total fraction volume to approximately 1600 μ L.

Both the eluate and flow through fractions were dried down under nitrogen, and reconstituted with 200 μ L 1:1 (v/v) chloroform/methanol. 1 μ L of the eluate fraction was combined with 99 μ L of the flow through fraction to give a 1:100 dilution. This combined sample was then injected into the UPLC/MS system.

Lipid class	Polarity	MRM time window (min)	No. of species detected	Cone voltage (V)	Collision energy (V)
Monohexylceramide (MonoHexCer)	+	0 to 2	16	20	30
Phosphatidylglycerol (PG)	H	1 to 3	19	55	45
Dihexylceramide (DiHexCer)	+	2 to 4	16	20	30
Phosphatidylinositol (PI)	-	3 to 5	26	48	30
Phosphatidylethanolamine (PE)	-	4 to 6	33	48	40
Phosphatidylcholine (PC)	+	5 to 7	47	36	30
Lyso-Phosphatidylinositol (LPI)	-	5 to 7	11	48	30
Lyso-Phosphatidylethanolamine (LPE)	1	6 to 8	11	36	24
Sphingomyelin (SM)	+	7 to 9	18	36	24
Lyso-Phosphatidylcholine (LPC)	+	8 to 10	11	42	26

Table 1. Xevo TQ-S MRM method.

Results and Discussion

Lipid extraction from Ostro plates using automated liquid handlers

While advancements in instrument technologies such as UPLC have enabled faster analysis of large numbers of samples, sample preparation time has been commonly viewed as the bottleneck.

When performed manually, lipid extraction using the Bligh and Dyer method takes approximately one hour and the Ostro plate method 1.5 hours. As the Ostro plate is available in a standard 96-well plate design, it can be easily adapted to most automated liquid handlers making the process time for one sample or 96 samples approximately the same, for example, 1.5 hours. Whereas, it took almost two days to manually process 96 samples using the Bligh and Dyer method.

Using an automated liquid handler to process the plasma samples, as shown in Figure 2, the well-to-well reproducibility (%CV) improved compared to manually performing the extraction. This was true for all classes of lipids analyzed, with improvements ranging from 25% (SM) to 60% (PE).

	Manual	Automation
Lipid Class	Overall %CV	Overall %CV
GluCeramides	13.2	9.7
Phosphatidylglycerol (PG)	10.2	7.2
Phosphatidylinositol (PI)	10.2	5.8
Phosphatidylethanolamine (PE)	16.0	6.4
Phosphatidylcholine (PC)	9.9	7.5
Lyso-Phosphatidylinositol (LPI)	23.0	13.0
Lyso-Phosphatidylethanolamine (LPE)	16.0	6.4
Sphingomyelin (SM)	11.9	9.1
Lyso-Phosphatidylcholine (LPC)	14.4	8.4

Table 2. Comparison of well-to-well reproducibility (%CV) of the Ostroplate for manual versus automated sample handling.



Figure 2. Typical screen shot of the Xevo TQ-S MRM method editor.

MRM method setup on Xevo TQ-S

In reversed-phase chromatography of lipids, separation is governed by lipophilicity, alkyl chain length, and degree of saturation for each individual lipid. These result in broad MRM acquisition time windows⁴ that, in turn, negatively affect the instrument's duty cycle, thus hindering accurate quantification.

In the HILIC-UPLC/MS method used in this application note, there was a clear and reproducible separation of the various classes of lipids. This was observed very clearly by the difference in retention times of PCs and SMs, which are normally difficult to identify and quantify using reversed-phase methods.^{5, 6}

By leveraging the reproducible retention times of the lipid classes, MRM acquisition time windows were reduced to two minutes per class, as shown in Table 1. This allowed for the creation of a single MS method to analyze the combined "flow through" and "eluate" fractions in a single run. A total of 215 MRM transitions (+/- polarity) including internal standards were created in this method. Figure 3 shows each lipid transition set up as a single function, for example, monohexosylceramide d18:1/22:0, which will limit the addition or

subtraction of lipids to only those of interest to the operator.



Figure 3. Typical lipid MRM trace using monohexosylceramide d18:1/22:0 as an example.

Data processing and reporting

A complementary data processing method was created using the TargetLynx Application Manager, as shown in Figure 4. The insert shows how easily peak information can be "dragged and dropped" into the data processing method.

Once the processing method had been set with the appropriate transitions and retention times for each lipid, batch processing for any number of samples run under the same conditions described above can be performed. Figure 4 shows a typical TargetLynx results view. Using monohexosylceramide d18:1/22:0 as an example, the application manager automatically integrates both the sample peak and the pre-defined internal standard (IS) peak and calculates the concentration of the lipid in the sample against the known spiked concentration of the IS. A user-defined report can then be printed, or these results can be exported into a number of popular generic formats for further statistical analysis.

			M	lonoHexCer.d18.1/2.	2.0		
Name	Trace	RT	Area	IS Area Norm	alised Peak Area	Calculated Concentration (ng/mL)	
1 02Dec11_MRM_OstroReplicate01_01	784.67 > 264.4	0.61	134496	1140644	11.791	58.96	
2 02Dec11_MRM_OstroReplicate01_02	784.67 > 264.4	0.60	578365	5124017	11.287	56.44	
3 02Dec11_MRM_OstroReplicate01_03	784.67 > 264.4	0.60	550805	4835302	11.391	56.96	
4 02Dec11_MRM_OstroReplicate02_01	784.67 > 264.4	0.60	505959	4842669	10.448	52.24	
5 02Dec11_MRM_OstroReplicate02_02	784.67 > 264.4	0.60	489794	4655512	10.521	52.60	
02Dec11_MRM_OstroReplicate02_03	784.67 > 264.4	0.60	474883	4571429	10.388	51.94	
02Dec11_MRM_OstroReplicate03_01	784.67 > 264.4	0.60	546428	4864058	11.234	56.17	
02Dec11_MRM_OstroReplicate03_02	784.67 > 264.4	0.60	518860	4713043	10.998	54.99	
9 02Dec11_MRM_OstroReplicate03_03	784.67 > 264.4	0.60	550744	4862458	11.326	56.63	
10 00Dac11 MRM OctroRenicola04 01	794 67 - 364 4	n en	570010	4530337	11 044	£0.70	
20ec11_MFM_OstroReplicate01_02 Smooth() combined eluent and flowthrough at 1:100 dilub 00	dn,3x2) on MonoHexCer.d18:1 0.60 578364.50 18688836	/22:0					
220ec11_URRM_OstoReplicate01_02 Smooth/b Combined eluent and flowthrough at 1:100 diluti	4n,3x2) on MonoHexCer.d18.1 0.60 578364.50 19688836 0.49	/22.0	Sample p	oeak			
02Dec 11_MRM_OstroReplicate01_02 Smooth() Combined eluent and flowthrough at 1:100 diluti 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	4n,3x2) on MonoHexCerd18:1 0.60 57354.50 1868836 0.49 4n,3x2) on GluCerd18:1/8 0.62 5124016 12794035	22:0 .0 STD 2	Sample p	oeak 092		- <u>1</u> 1111111	
02Dect 11_MRM_OstroReplicate01_02 Smooth() Combined eluent and flowthrough at 1:100 diluts 0 0 02Dect 11_MRM_OstroReplicate01_02 Smooth() Combined eluent and flowthrough at 1:100 diluts	4n,3x2) on MonoHexCer.d18.1 0.60 575364.50 1969836 0.49 4n,3x2) on GluCer.d18.18 0.62 5124016.1 12794035	122:0 10 STD 10 2 10 2	Sample p	oeak ⁰⁹² Standard pe	eak		

Figure 4. Quantification of monohexosylceramide d18:1/22:0 MRM trace using TargetLynx. Chromatographic peaks are automatically detected and quantified against a standard for a series of samples.

Conclusion

Consistent, reliable, and rapid identification and quantification of hundreds of lipid species can now be performed in a single run by the application of this workflow. The high throughput nature of the workflow utilizing automation technologies and automated data processing and reporting means that large-scale comparative lipidomic studies can be routinely used by laboratories around the world. In addition, the consistency obtained from this standardized platform means that data can be shared and compared across various sites, thereby enabling a greater understanding of the global lipidome and its associations with diseases.

References

- 1. Ferreiro-Vera C *et al*. Comparison of sample preparation approaches for phospholipids profiling in human serum by liquid chromatography-tandem mass spectrometry. *J Chrom A*. 2012;1240: 21-8.
- 2. Pucci V *et al*. A novel strategy for reducing phospholipids-based matrix effect in LC-ESI-MS bioanalysis by means of Hybrid SPE. *J Pharm Biomed Anal.* 2009;50(5): 867-71.
- 3. Ritchie M *et al*. Extraction of Phospholipids from Plasma using Ostro Sample Preparation. Waters Application Note 720004201en. 2012.
- 4. Goh E, Ritchie M. Rapid Lipid Profiling of Serum by Reverse Phase UPLC-Tandem Quadrupole MS. Waters Application Note 720004035en. 2011.
- 5. Mal M, Wong S. A HILIC-Based UPLC-MS Method for the Separation of Lipid Classes from Plasma. Waters Application Note, 720004048en. 2011.
- Ritchie M *et al.* UPLC BEH HILIC: The Preferred Separation Chemistry for Targeted Analysis of Phospholipids. Waters Application Note 720004219en. 2012.

Featured Products

Xevo TQ-S <https://www.waters.com/10160596> ACQUITY UPLC System <https://www.waters.com/514207> TargetLynx <https://www.waters.com/513791>

720004521, December 2012

©2019 Waters Corporation. All Rights Reserved.