Waters™

应用纪要

Rapid Screening of Consumer Products for Disperse Dyes

Jane Cooper, Jeremy Marchand

Waters Corporation This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This application brief describes the simple and rapid screening of disperse dyes in consumer products without sample preparation or chromatographic separation.

Benefits

Xevo TQD with ASAP provides a rapid and easy screening method for disperse dyes.

Introduction

Disperse dyes are used for coloring many synthetic fibers used in consumer products, such as polyester, nylon, synthetic velvets, and PVC.

Disperse dyes are included in the restricted substances lists (RSL) utilized by many companies to fulfill their commitment to protect the consumers of their products, their workers, as well as the

community/environment.

There are legislated and non-legislated regulations and standards that detail many of the potentially sensitizing, carcinogenic, mutagenic, or toxic to reproduction disperse dyes as prohibited including: European Committee for Standardization with regard to toy safety standards (BS EN 71 part 9),¹ Oeko-Tex Standard 1000,² European Union Commission Decision (2009/567/EC),³ and the German Food and Commodities law (LFGB § 30).

This study describes the simple and rapid screening of disperse dyes in consumer products without sample preparation or chromatographic separation.

Qualitative screening methods offer the chemical industry valuable time and cost savings, faster sample turnaround times, and reduced use and associated disposal of non-environmentally friendly solvents and chemicals.

Results and Discussion

Using the Waters Atmospheric Solids Analysis Probe (ASAP) with the Xevo TQD, this study demonstrates the rapid screening of consumer products for disperse dyes by direct analysis under atmospheric pressure ionization (API) conditions, with no sample preparation or chromatographic separation. ASAP can easily and quickly be fitted to the API source in less than 2 min by replacing the current probe and fitting a corona discharge pin. Optimum MRM conditions were developed, as shown in Table 1.

Chemical substance	Cone voltage (V)	Transitions	Collision energy	
Diamaraa Diwa 2	45	297.0 > 235.1*	33	
Disperse Blue 3	45	297.0 > 252.0	21	
Disperse Blue 106	42	336.0 > 147.0*	35	
		336.0 > 178.0	17	
Disperse Brown 1	53	433.0 > 197.1	31	
		433.0 > 357.0*	37	
Disperse Orange 1	49	319.0 > 122.0	22	
		319.0 > 169.0*	26	
Disperse Orange 3	45	243.0 > 92.0*	22	
		243.0 > 122.0	18	

Table 1. Disperse dyes, cone voltages, MRM transitions, and associated collision energy values (*refer to the confirmation transition).

0.5 g of synthetic fabric was spiked with select disperse dyes at a level of 75 μ g/g (many laboratories that base their extraction protocol for disperse dyes on standard method DIN 54231⁴ accept 75 μ g/g as the practical detection limit).

The ASAP assembly is in two parts, as seen in Figure 1, where the outer probe is continually attached to the source throughout usage; whereas, the inner probe can be removed to load the sample without compromising the source conditions. The sample was loaded by rubbing the end of a sealed glass melting point capillary tube over the fabric sample. The glass capillary, attached to the inner ASAP probe, was loaded directly into the source enclosure of the Xevo TQD. The sample was vaporized in a stream of heated nitrogen and ionized by gas phase ion/molecule reactions from a corona discharge needle. A vial of methanol was also placed in the source to act as a chemical modifier, influencing the ionization mechanism to favor the formation of protonated ions. The MRM chromatograms, revealing the desorption profiles generated from ASAP analysis on spiked fabric samples, are shown in Figure 2.



Figure 1. The Xevo Atmospheric Solids Analysis Probe (ASAP).



Figure 2. MRM chromatograms for selected disperse dyes showing the desorption profiles generated from ASAP analysis on A) a spiked fabric sample (equivalent to 75 µg/g), and B) a blank fabric sample.

Conclusion

Waters Xevo TQD with ASAP can be used to provide a rapid, simple screening solution for disperse dyes in consumer products.

No sample preparation or chromatographic separation was required, resulting in a fast run time and reduced overall analysis time, offering businesses valuable productivity and consumables cost savings.

The Xevo TQD offers the ability to acquire data in MRM mode, providing improved selectivity and sensitivity to screen for disperse dyes. The system can also be rapidly switched to UPLC-MS/MS mode if quantification and confirmation of disperse dyes in consumer products is required.

References

- 1. BS EN 71-9:2005+A1:2007 Safety of toys. Organic chemical compounds. Requirements.
- Oeko-Tex Standard 1000, Oeko-Tex International, Ausgabe/Edition 01/2012.[cited 2012 September 20]. Available from: https://www.oekotex.com/xdesk/ximages/470/15540_1000-DEF.pdf
- The Commission of the European Communities. Commission Decision of 9 July 2009 establishing the ecological criteria for the award of the Community Ecolabel for textile products (2009/567/EC). Official Journal of the European Union. L 197/70: 70-86, 9th Jul 2009. [cited 2012 September 20]. Available from: http://eurlex.europa.eu/LexUriServ/LexUriServ.do?uri=OJ:L: 2009:197:0070:0086:EN:PDF
- 4. German Institute for Standardization (DIN). Textiles Detection of dispersed dyestuffs. DIN54231:2005.

Featured Products

Xevo TQD Triple Quadrupole Mass Spectrometry https://www.waters.com/134608730

720004572, January 2013

©2019 Waters Corporation. All Rights Reserved.

Terms of L	Jse	Privacy	Trademarks	Sitemap	Careers	Cookie
Cookie	设置					

沪 ICP 备06003546号-2 京公网安备 31011502007476号