

Matrix Factor Determination for Clopidogrel Using Post-Column Infusion (PCI) Capabilities of UNIFI Software

Jennifer Simeone, Paul D. Rainville

Waters Corporation

Abstract

This work illustrates how the UNIFI Scientific Information System, used with the Xevo TQ-S MS for post-column infusion (PCI), facilitates the determination of matrix factor in bioanalysis studies.

Benefits

UNIFI Software automates the calculation of both analyte and normalized internal standard matrix factor.

Introduction

The reliability of analytical data upon which critical toxicological and efficacy findings are based is an essential part of bioanalysis. LC-MS/MS is the technique of choice in quantitative bioanalysis due to its high selectivity and sensitivity, as well as the time savings afforded by significantly faster chromatographic separations and minimized sample preparation.

LC-MS/MS quantitative analysis is influenced by a phenomenon called ion suppression or matrix effects,

wherein matrix components present in the biological sample influence the response of the analyte under investigation. The need to adequately address matrix effects data during the method development and validation process has been clearly identified.¹⁻³ This information is reported as matrix factor (MF), which is defined as analyte response in the presence of matrix components divided by analyte response in pure solution.

As drug compounds under investigation become increasingly potent, they require lower doses for efficacy and toxicology assessment. This translates to lower limits of quantitation (LLOQ) during bioanalysis, wherein matrix components in the sample can be present at levels that are much higher than the target analyte. In addition, over the course of pre-clinical and clinical trials, very often a number of samples to be analyzed will contain varying degrees of hemolysis arising from erroneous processing of the blood to plasma. Therefore, it is suggested that hemolyzed samples also be considered during method development and validation to assess any potential effects arising from the matrix. For example, the current EMEA guidelines require that in addition to six unique lots of plasma, hemolyzed plasma should also be tested for matrix effects.

The Waters UNIFI Scientific Information System enables the user to easily quantify matrix factor either by post-column infusion or in a spiked experiment. The software is designed to do all necessary calculations and data summaries that a user requires, removing the need to use other software packages such as Microsoft Excel. In this application note, we present a simplified approach for matrix factor determination using the onboard fluidics of the Xevo TQ-S and UNIFI Software.

Experimental

UPLC conditions

LC system:	ACQUITY UPLC
LC column:	ACQUITY UPLC BEH C ₁₈ 1.7 μm, 2.1 x 50 mm
Flow rate:	600 μL/min
Column temp.:	45 °C

Mobile Phase A:	0.1% Formic acid
Mobile Phase B:	Acetonitrile
Gradient:	5% B to 95% B over 2 min

MS conditions

MS system:	Xevo TQ-S
MS/MS parameters:	Transitions: Clopidogrel 322.1 > 212.1 d ₄ -clopidogrel 326.1 > 216.1
Ionization:	Positive ESI
Capillary voltage:	1.00 kV
Collision energies:	16 eV
Cone voltage:	35 V

Data acquisition, processing, and reporting

UNIFI Scientific Information System


Sample Preparation

Three lots of hemolyzed plasma were prepared by adding the appropriate volume of hemolyzed whole blood (human, K₂EDTA) to plasma (human, K₂EDTA) resulting in 5%, 10%, and 15% hemolysis (i.e., 50 µL of hemolyzed blood was combined with 950 µL of plasma to yield 5% hemolyzed plasma). In addition to these three lots, non-hemolyzed blank plasma was also used in the matrix factor evaluation. Each lot of matrix was extracted in replicates of six using a protein precipitation extraction technique where 100 µL of the appropriate matrix was precipitated with 300 µL of methanol, vortex mixed and then centrifuged. Three levels of clopidogrel/ d₄-clopidogrel infusion solutions were prepared in methanol at the following concentrations: 10 ng/mL, 50 ng/mL, and 100 ng/mL.

Results and Discussion

Using UNIFI Software, specific analysis types were defined whereby the method automatically selects specific settings, parameters, and calculations that characterize the analysis type. For example, the post-column infusion MF analysis type is set up such that the software will calculate matrix factor based on the specific sample types entered by the user.

In the sample list, the user inputs the solution injections as “blank,” and the matrix extracts as “matrix spike,” shown in Figure 1. If there are multiple concentrations to be used for MF determination (e.g., low, mid, and high concentrations), these are defined as levels for both the blank and matrix spike sample types.



The screenshot shows the 'Sample list' window in UNIFI software. It contains a table with 14 columns: Item name, Sample type, Sample position, Run time (min), Injection volume (μL), Replicates, Level, Flow path, Fluidics flow rate (μL/min), Reservoir type, Description, Processing options, and Acquisition status. The table lists 20 samples, including standards, blanks, and matrix spikes at various levels (1, 2, 3, 4, 5, 6, 7, 8, 10, 15, 100) and concentrations (5%, 10%, 100% hemolysis).

Item name	Sample type	Sample position	Run time (min)	Injection volume (μL)	Replicates	Level	Flow path	Fluidics flow rate (μL/min)	Reservoir type	Description	Processing options	Acquisition status
03Dec2012_001	Standard	1:F,3	4.00	5.00	1	Level 1	LC	0.5	Off	standard	Quantitation standard	
27Nov2012_002	Standard	1:F,3	4.00	5.00	1	Level 1	LC	0.5	Off	standard	Quantitation standard	
27Nov2012_003	Blank	1:F,1	4.00	5.00	1	Level 1	Combined	5.0	A	blank methanol		
27Nov2012_003	Blank	1:F,1	4.00	5.00	1	Level 1	Combined	5.0	A	blank methanol		
27Nov2012_003	Blank	1:F,1	4.00	5.00	1	Level 1	Combined	5.0	A	blank methanol		
27Nov2012_003	Blank	1:F,1	4.00	5.00	1	Level 1	Combined	5.0	A	blank methanol		
27Nov2012_003	Blank	1:F,1	4.00	5.00	1	Level 1	Combined	5.0	A	blank methanol		
27Nov2012_004	Matrix Spike	1:A,1	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_005	Matrix Spike	1:A,2	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_006	Matrix Spike	1:A,3	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_007	Matrix Spike	1:A,4	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_008	Matrix Spike	1:A,5	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_009	Matrix Spike	1:A,6	4.00	5.00	1	Level 1	Combined	5.0	A	plasma		
27Nov2012_010	Matrix Spike	1:A,7	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_011	Matrix Spike	1:A,8	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_012	Matrix Spike	1:B,1	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_013	Matrix Spike	1:B,2	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_014	Matrix Spike	1:B,3	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_015	Matrix Spike	1:B,4	4.00	5.00	1	Level 1	Combined	5.0	A	5% hemolysis		
27Nov2012_016	Matrix Spike	1:B,5	4.00	5.00	1	Level 1	Combined	5.0	A	10% hemolysis		
27Nov2012_017	Matrix Spike	1:B,6	4.00	5.00	1	Level 1	Combined	5.0	A	10% hemolysis		
27Nov2012_018	Matrix Spike	1:B,7	4.00	5.00	1	Level 1	Combined	5.0	A	10% hemolysis		
27Nov2012_019	Matrix Spike	1:B,8	4.00	5.00	1	Level 1	Combined	5.0	A	10% hemolysis		
27Nov2012_020	Matrix Spike	1:C,1	4.00	5.00	1	Level 1	Combined	5.0	A	10% hemolysis		

Figure 1. Example of a post-column infusion matrix factor sample list.

Once the data has been acquired and processed, the MFs will automatically be calculated based on the summary calculations built into the MF analysis type. There is no longer a need to use additional software, such as Microsoft Excel, to calculate and summarize the MF values.

By simply choosing “matrix factor results” on UNIFI’s Review tab, the calculated matrix factor data is displayed on a per-component basis, shown in Figure 2 with calculated statistics such as mean, standard deviation, and relative standard deviation (or coefficient of variation). In addition, the user can view

chromatograms and summary plots within the same window.




Figure 2. Example results data for calculated matrix factor for 15% hemolyzed plasma prepared at 8.5 pg/mL.

The resulting MFs for all lots of hemolyzed and non-hemolyzed matrix prepared at each concentration level are displayed in Table 1. A matrix factor with a value of less than one indicates suppression, a value greater than one indicates enhancement, and a value of one indicates there is no effect of the matrix on the analyte signal.

Clopidogrel Matrix Factor												
Low Concentration (10 ng/mL)				Mid Concentration (50 ng/mL)				High Concentration (100 ng/mL)				
Plasma	15% hemolysis	10% hemolysis	5% hemolysis	Plasma	15% hemolysis	10% hemolysis	5% hemolysis	Plasma	15% hemolysis	10% hemolysis	5% hemolysis	
0.864	0.821	0.762	0.670	0.722	0.779	0.661	0.669	0.917	0.753	0.898	0.836	
0.834	0.787	0.796	0.685	0.721	0.808	0.889	0.771	0.902	0.846	0.871	0.756	
0.823	0.750	0.776	0.663	0.648	0.751	0.836	0.704	0.821	0.854	0.872	0.890	
0.805	0.815	0.795	0.724	0.745	0.600	0.810	0.667	0.786	0.869	0.886	0.905	
0.764	0.769	0.729	0.754	0.850	0.654	0.441	0.860	0.804	0.788	0.892	0.908	
0.655	0.618	0.725	0.691	0.819	0.650	0.820	0.860	0.902	0.854	0.815	0.885	
Mean	0.791	0.760	0.764	0.698	0.751	0.707	0.743	0.755	0.856	0.827	0.872	0.863
St Dev	0.0742	0.0747	0.0313	0.0348	0.0732	0.0834	0.166	0.0895	0.0578	0.0458	0.0301	0.0587
% CV	9.4	9.8	4.1	5.0	9.7	11.8	22.4	11.9	6.8	5.5	3.5	6.8
% Suppression	20.9	24.0	23.6	30.2	24.9	29.3	25.7	24.5	14.4	17.3	12.8	13.7

Table 1. Matrix factor results for four different lots of matrix.

The data shows that at each concentration level, there is no discernible variation between the different lots of

matrix, therefore the varying degrees of hemolysis are not impacting the produced signal. The obtained MF values for clopidogrel do show some variation between the different concentrations, with suppression values ranging from approximately 12% to 30% across all injections. It is not surprising that suppression is seen in this assay given that the extraction technique used was a protein precipitation, which provides relatively little clean-up to the samples.

In a bioanalytical assay, it is preferable that a deuterated version of the analyte be used as the internal standard (IS) since it will behave in the same manner chromatographically and spectroscopically as the analyte of interest. This includes ion suppression/enhancement since the analyte signal and the IS signal should be impacted in the same way and to the same extent. It is for this reason that matrix factor is often reported as IS normalized matrix factor to take this fact into account. IS normalized matrix factor is defined as the matrix factor of the analyte divided by the matrix factor of the internal standard. Table 2 reports the IS normalized matrix factor for the four lots of matrix.

IS Normalized Matrix Factor												
	Low Concentration				Mid Concentration				High Concentration			
	Plasma	15% hemolysis	10% hemolysis	5% hemolysis	Plasma	15% hemolysis	10% hemolysis	5% hemolysis	Plasma	15% hemolysis	10% hemolysis	5% hemolysis
	0.993	1.12	1.10	1.12	0.987	1.09	1.14	1.12	1.06	1.06	1.09	1.09
	1.02	1.11	1.11	1.11	1.11	1.09	1.06	1.04	1.04	1.10	1.08	1.07
	1.04	1.10	1.10	1.13	1.03	1.25	1.11	1.02	1.06	1.09	1.09	1.10
	1.05	1.09	1.07	1.07	1.02	1.09	1.11	1.13	1.09	1.10	1.07	1.10
	1.07	1.10	1.07	1.09	1.11	1.12	0.846	1.08	1.05	1.06	1.08	1.11
	1.12	1.13	1.10	1.13	1.09	1.09	1.10	1.10	1.09	1.10	1.07	1.09
Mean	1.05	1.11	1.09	1.11	1.06	1.12	1.06	1.08	1.06	1.09	1.08	1.09
St Dev	0.0416	0.0150	0.0175	0.0252	0.0524	0.0657	0.109	0.0452	0.0198	0.0203	0.00906	0.0130
% CV	4.0	1.4	1.6	2.3	4.9	5.9	10.2	4.2	1.9	1.9	0.8	1.2
% Enhancement	4.9	10.9	9.2	10.8	5.9	12.2	6.2	8.2	6.4	8.6	8.0	9.2

Table 2. IS normalized matrix factor results for four different lots of matrix.

As shown, the IS normalized values are all relatively close to one, indicating that the internal standard was suppressed to the same degree as the analyte; thus, when normalized, negligible suppression/enhancement is seen.

Conclusion

- The capabilities of UNIFI Software coupled with the on-board fluidics of the Xevo TQ-S allowed for easy matrix factor determination using post-column infusion.

- UNIFI Software provides users with both the calculated analyte matrix factor as well as the internal standard normalized matrix factor.
- The effect of varying degrees of hemolysis within extracted plasma were examined for clopidogrel, and resulted in no discernible differences.

References

1. Guidance for Industry, Bioanalytical Method Validation, May 2001, <http://www.fda.gov/downloads/Drugs/GuidanceComplianceRegulatoryInformation/Guidances/ucm070107.pdf>
2. Guideline on Bioanalytical Method Validation, July 2011, http://www.ema.europa.eu/docs/en_GB/document_library/Scientific_guideline/2011/08/WC500109686.pdf
3. Eeckhaut A. Van, Lanckmans K. , Sarre S., Smolders I., Michotte Y., *J. Chromatogr. B* 877 (2009) 2198-2207

Featured Products

[ACQUITY UPLC System <https://www.waters.com/514207>](https://www.waters.com/514207)

[Xevo TQ-S <https://www.waters.com/10160596>](https://www.waters.com/10160596)

[UNIFI Scientific Information System <https://www.waters.com/134801648>](https://www.waters.com/134801648)

720004627, March 2013