



Quantitation by High Resolution Mass Spectrometry: Using Target Enhancement and Tof-MRM to Achieve Femtogram-level On-column Sensitivity for Quantitation of Drugs in Human Plasma

Yun Alelyunas, Mark Wrona, Russell J. Mortishire-Smith, Nick Tomczyk, Paul D. Rainville

Waters Corporation



Abstract

In this study, which employs the Waters SYNAPT G2-S mass spectrometer, we determine and compare the quantification attributes of detection limit (LOD), quantification limit (LOQ), and linear dynamic range for three data-acquisition modes: full-scan MS, MS^E and ToF-MRM.

Benefits

- Purely quantitative mode available for routine quantitation
- Simple, enables triple-quadrupole-like workflows on a ToF platform
- Highly selective and sensitive mode of acquisition
- Complementary to full scan and Quan-Qual modes of operation
- Compact file size, for routine quantitation

Introduction

High resolution mass spectrometry (HRMS) has been widely adopted in drug discovery and development organizations. In medicinal chemistry, drug metabolism and pharmacokinetics, and impurity identification, the use of HRMS has revolutionized qualitative screening and structure confirmation. Although triple quadrupole mass spectrometry (MS) remains the gold standard for quantitation, advances in the technology of HRMS platforms have led to improved selectivity, dynamic range, linearity, sensitivity, and a consequent increased usage and interest in HRMS for quantitative studies. In the presence of complex matrix interferences that might prove challenging for triple quadrupole MS approaches, the markedly different nature of HRMS data acquisition can achieve the selectivity and sensitivity for successful quantitation.

Recently, for quantitative analysis, Waters introduced a new targetenhancement mode of data acquisition, ToF-MRM, on the SYNAPT G2-S HDMS and Xevo G2-XS systems. In this mode, a precursor ion is selected by the quadrupole and, optionally, fragmented in the collision cell. The ToF pusher is then synchronized with the m/z of the precursor or a product ion, maximizing the duty cycle for a target m/z range and effecting an increase in response and selectivity.¹ With the addition of the ToF-MRM mode of acquisition, these platforms receive a boost in selectivity and sensitivity, which can be exploited for enhanced HRMS quantitation.

In this study, which employs a SYNAPT G2-S instrument, we determine and compare the quantification attributes of detection limit (LOD), quantification limit (LOQ), and linear dynamic range for three data-acquisition modes: full-scan MS, MS^E and ToF-MRM.

Experimental

LC conditions

LC system:	ACQUITY UPLC I-Class
Column:	ACQUITY BEH C ₁₈ 2.1 x 50 mm, 1.7 µm
Column temp.:	60° C
Sample temp.:	10° C
Injection volume:	5 µL
Flow rate:	0.6 mL/min
Mobile phase A:	water with 0.1% formic acid
Mobile phase B:	90% acetonitrile/ 10% methanol with 0.1% formic acid
Gradient:	2–60% B in 4 minutes, 60–95% in 1 minute, held at 95% B for 1 minute before reverting to initial conditions. The total run time was 7 minutes.

MS conditions

MS system:	SYNAPT G2-S HDMS
------------	------------------

Ionization mode:	ESI+, resolution mode(>20,000 FWHM)
Acquisition range:	50-1000 <i>m/z</i>
Capillary voltage:	1 kV
Cone voltage:	30 V
Source temperature:	120 °C
Desolvation gas temp.:	500 °C
Cone gas flow:	20 L/h
Desolvation gas flow:	800 L/h
Scan time:	0.2 s, continuum

Experiment

MS settings:	precursor mass
MS ^E collision energy (CE) settings:	low CE, 2.0 eV; high CE, ramp 10-30 eV
Tof-MRM settings:	precursor mass (see "Results and Discussion" section for details)

Sample description

Quenched and diluted human plasma was prepared by adding three volumes of acetonitrile to human plasma. After centrifugation, the supernatant was transferred to a new vial and diluted 1:1 with H₂O. The test compounds (alprazolam, verapamil, buspirone, and clopidogrel) were combined as 100 ng/mL for each in a solution containing quenched and diluted human plasma, prepared as stated above. The mixture was then serially diluted 1:1 using quenched and diluted human plasma to achieve a concentration range from 100 ng/mL to ~1 pg/mL.

Method conditions

The analytical LC/MS experiments were performed on a Waters ACQUITY UPLC I-Class System coupled with a SYNAPT G2-S QToF Mass Spectrometer. Waters MassLynx Software was used for data acquisition, and TargetLynx Software was used for data processing.

Results and Discussion

Four compounds – alprazolam, verapamil, buspirone, and clopidogrel (Figure 1) – were quantified as a mixture using the ACQUITY UPLC I-Class System and SYNAPT G2-S HDMS. Data were collected using positive electrospray ionization and three acquisition modes: MS, MS^E and ToF-MRM (resolution mode). The MS mode collects full-scan data over a wide mass range for precursor ions. The MS^E mode collects full-scan data of both precursor and product ions by alternating low CE and high CE scans in the same experiment. This all-the-data-all-the-time mode provides spectral information about target compounds, as well as fragmentation patterns, to aid in structural elucidation. As such, the MS^E approach is frequently used for Quan-Qual screening workflows. For pure quantitation, we acquired data using the ToF-MRM mode with target enhancement. In the ToF-MRM mode, the precursor ion is selected in the quadrupole (MSMS). The ToF pusher at the detector entrance is then synchronized with the targeted masses using the instrument's Target Enhancement function, which significantly increases the duty cycle for ions of interest. ToF-MRM can be used to analyze intact ions (parent > parent, no CE applied) or fragment ions (parent > fragment, CE applied) transitions. ToF-MRM-enhanced transmission mode dramatically enhances ion counts for targeted ions versus generic full-scan modes, which is then coupled with selectivity afforded through quadrupole precursor selection and accurate HRMS monitoring of the target ion. A schematic of the target-enhancement mode of data acquisition on the SYNAPT G2-S is shown in Figure 2.




Figure 1. Structure of compounds used in this study.




Figure 2. A schematic of SYNAPT G2-S illustrating the regions involved in the ToF-MRM target enhancement.

In this study, the same LC conditions, cone voltage, source settings, data acquisition speed and the same data-processing parameters (data smoothing and integration) were used for the data collected in each of the three acquisition modes. Generic LC and MS conditions were used with no compound-specific optimization.

Figure 3 shows the extracted mass chromatogram for each compound, comparing all three modes of data acquisition using the 24 pg/mL sample as an example. Figure 4 shows the measured peak area, comparing the modes at a low concentration of 24 pg/mL and a mid-range concentration of 100 pg/mL. These plots indicate that the responses from ToF-MRM are 2–8 times higher than those of either of the scanning modes of data acquisition. The responses for MS and MS^E modes were of comparable magnitude. Mass spectral differences between fullrange MS and ToF-MRM scan modes are illustrated for buspirone in Figure 5. Compared with MS full-scan data, the ToF-MRM mass spectrum shows a higher signal intensity and less baseline noise, leading to enhanced sensitivity and selectivity. Points across the peak were measured for the ToF-MRM mode of acquisition using representative samples for all four compounds: 49 pg/mL as the low concentration and 12,500 pg/mL levels as the high concentration. All peaks contained between 13 and 19 points (base to base), indicating ToF-MRM data sampling is fast enough for quantitation purposes when coupled with UPLC chromatography.




Figure 3. Overlaid extracted-mass chromatograms at 24 pg/mL concentration using MS, MS^E and ToF-MRM data acquisition.




Figure 4. Plot of peak areas at 24 pg/mL (left) and 100 pg/mL (right), with data acquired using MS, MS^E and ToF-MRM modes.




Figure 5. Mass spectrum of buspirone at 400 pg/mL. MS full scan (top) and ToF-MRM scan(bottom).

Determination of the detection limit and linear dynamic range was based on the regulated bioanalysis method of validation criteria. The limit of detection was determined according to a signal-to-noise ratio of greater than 3. The linear range was determined using values that met the standard bioanalysis criteria of less than 15% deviation across the concentration range. A weighting of $1/X^2$ was used for the calibration curve fitting. Table 1 shows the compiled results for all four compounds using these three data-acquisition

modes.

Results indicate that, in the presence of human plasma, all three modes—particularly the ToF MRM mode—produced an excellent, linear dynamic range. That range, from 3.6 to 3.9 log units, enabled routine quantitation studies. When compared with MS and MS^E ToF-MRM demonstrated an enhancement factor of 4 – 8 in LOD and 4 – 12 in LOQ. For three of the four compounds, the SYNAPT G2-S instrument attained 1.5 pg/mL, or 7.5 femtograms on-column sensitivity, and a signal-to-noise ratio of 10 or greater using data acquired using the ToF-MRM mode. Figure 6 shows the ToF-MRM, extracted-mass chromatograms of the four compounds, including the plasma blank, at LOQ concentration and at LOQ concentration x 4. Table 2 presents a summary of the percent deviation of all data points used in calculating LOD, LOQ, and linear dynamic range.

Compound	Acquisition Type	Linear range [Log] (pg/mL)	R ²	LOD	LOQ (pg/mL)	S/N@LOQ	Data file size (MB)
Alprazolam	ToF-MRM	6-50,000 [3.9]	0.992	6	6	6.7	18
	MS	24-100,000 [3.6]	0.995	12	24	5.6	660
	MS ^E	49-100,000 [3.3]	0.999	49	49	7.0	500
Buspirone	ToF-MRM	<1.5-12,500 [3.9]	0.997	<1.5	1.5	10	18
	MS	6-50,000 [3.9]	0.996	6	6	5.1	660
	MS ^E	6-50,000 [3.9]	0.994	6	6	4.3	500
Clopidogrel	ToF-MRM	6-25,000 [3.6]	0.995	1.5	6	10	18
	MS	12-100,000 [3.9]	0.995	12	12	8.8	660
	MS ^E	24-100,000 [3.6]	0.997	12	24	13	500
Verapamil	ToF-MRM	<1.5-12,500 [3.9]	0.996	<1.5	1.5	11	18
	MS	12-100,000 [3.9]	0.996	12	12	6.9	660
	MS ^E	24-100,000 [3.6]	0.995	6	24	4.1	500

Table 1. Summary of detection limit, quantification limit, linear range, linear dynamic range, and file size for each compound and at each mode of data acquisition. For those compounds with reported top linear range of 100,000 pg/mL, which is the highest concentration prepared in the study, it is possible that the actual linear range could be higher.

The last column in Table 1 is the data file size per injection for each mode of data acquisition. The file size for ToF-MRM is approximately 20 MB, representing a more than 95% reduction compared with MS and MS^E. Such a reduction in file size enables quicker data processing, less demand on data storage, and ease of use for quantitative applications.




Figure 6. ToF-MRM overlaid extracted-mass chromatograms of plasma blank at LOQ concentration and at the LOQ concentration x 4.

No.	Std. Conc (pg/mL)	Amt on Column (femtogram)	Alprazolam %Dev			Buspirone %Dev			Clopidogrel %Dev			Verapamil %Dev		
			ToF-MRM	MS	MS ^E	ToF-MRM	MS	MS ^E	ToF-MRM	MS	MS ^E	ToF-MRM	MS	MS ^E
1	1.53	7.7				6.6						-4.4		
2	3.05	15				-14.9						-5.1		
3	6.10	31	6.3			2.4	-1.1	8.3	5.6			0.5		
4	12.2	61	-13.4			2	-3.7	-12.4	-4.1	2.1		7.7	0.5	
5	24.4	122	3.6	8.2		0.4	8.2	-8	-8.9	-2.1	1.6	1.8	-5.7	-0.4
6	48.8	244	-6.2	-10.6	2.4	0.2	8.6	0.6	-4.5	-0.3	1.7	-2.9	11	6.2
7	97.7	488	-8.5	-9.2	-2.9	-1.5	0.6	-2.4	-7.7	-5.4	-3.3	-6.7	-2.6	-5
8	195	977	-8.2	-1.1	-2	-3	-5	-4.3	-6.9	-5.1	-11.1	-4.4	-1.3	-10.8
9	391	1953	-3.7	-5.7	-1	-0.3	0.3	1.4	-4.3	0.1	-4.6	-2.2	-2.2	-3.8
10	781	3906	-1.6	-4	-6	-0.3	-3	3.7	-1.4	-2.8	-1.7	-0.7	-1.7	-2.1
11	1563	7813	3.5	-0.4	-1	5.2	3.4	6	3.7	2.5	3.2	4.8	3	9.1
12	3125	15625	5.6	3	1.5	9.6	1.6	8.3	5	3.3	2.9	10.1	6.3	7.8
13	6250	31250	10.2	1.7	-0.9	6.3	2.7	8	11	3.4	7.2	9.2	4.8	8.1
14	12500	62500	10.6	1.5	0	-12.7	2.1	3.8	8.4	2.6	1.5	-7.5	3.4	3.5
15	25000	125000	12	5.7	1.2		-0.5	0	4	3.6	2.6		4	4
16	50000	250000	-10.3	1.1	0.3		-14.1	-13		-1.4	-1.2		-5.4	-4.6
17	100000	500000		9.9	8.3					-0.4	1.1		-14.1	-11.9

Table 2. Summary of calculated percent deviation for all compounds and concentrations used in this study.

Conclusion

The powerful modes of data acquisition, flexibility, and the superior acquisition speed of the ToF platform² have aided scientists in their various research activities. Building on this progress, we have shown that, in the presence of human plasma, the ToF-MRM mode of data acquisition can achieve on-column limits of detection that are in the low-femtogram range. It does so, moreover, with an excellent signal-to-noise ratio, fit for purpose linear dynamic range, and with high bioanalytical performance. These excellent quantitation attributes will enable scientists to adopt MRM-like workflows, which have been successful for triple quadrupole MS. Quantitation on HRMS platforms can further close the sensitivity gap between these instrument types and provides a unique opportunity to attain high MRM assay performance through selectivity which is different and complementary to conventional tandem MRM technologies. We expect to witness an extended involvement of HRMS platforms in applications where both quantitative and/or qualitative end points are desired. The mixed capability of qualitative modes to support screening applications combined with focused modes to support throughput and routine assay will further increase the use of HRMS platforms in pharmaceutical applications. ToF-MRM has direct applicability for bioanalysis,

targeted metabolism, stability, and PK/PD profiling studies with low detection limits.

References

1. Waters application note:720004728en "Targeted High Resolution Quantification with ToF-MRM and HD-MRM". <https://www.waters.com/waters/library.htm?cid=511436&lid=134744421&locale=en_US>
2. Waters application note: 720004762en "Effect of MS Scan Speed on UPLC Peak Separation and Metabolite Identification: Time-of-Flight HRMS vs. Orbitrap".

Featured Products

ACQUITY UPLC I-Class PLUS System <<https://www.waters.com/134613317>>

MassLynx Mass Spectrometry Software <<https://www.waters.com/513164>>

TargetLynx <<https://www.waters.com/513791>>

SYNAPT G2-Si High Definition Mass Spectrometry <<https://www.waters.com/134740622>>

720005182, September 2014