

Oasis PRiME HLB Cartridge for Rapid and Effective Cleanup of Avocado, a High Fat Matrix, Prior to APGC-MS/MS Analysis

Michael S. Young, Jeremy C. Shia

Waters Corporation

Abstract

In order to ensure public health and safety, reliable analytical methods are necessary to determine pesticide residue levels in foods. Many of the compounds are well suited for gas chromatography (GC) and are often determined in fruits and vegetables using GC with mass spectrometry (GC-MS).

This application note demonstrates an effective cleanup protocol for pesticide analysis in avocado, a highly fatty matrix. After a modified QuEChERS extraction, cleanup is performed using an Oasis PRiME HLB Cartridge. Analysis is performed using GC coupled with atmospheric pressure ionization mass-spectrometry (APGC-MS/MS).

Benefits

- Efficient, time-saving multi-class/multi-residue methodology
- Simple, rapid, and effective sample cleanup suitable for a diverse range of analytes in a fatty matrix
- Fast, sensitive APGC-MS/MS analysis

Introduction

In recent years, food safety laboratories have adopted new and simplified sample preparation methods designed to reduce analysis time and related costs, as well as to increase throughput. For example, the QuEChERS methods for fruits and vegetables require only minutes for sample preparation and replace prior methods that took hours or days. In this study, this type of simplified sample preparation is applied to pesticide analysis in avocado, a fruit matrix of very high lipid content. A typical avocado contains 10–15% fat and about 1% total phospholipids. In the QuEChERS extraction, significant amounts of the fat and phospholipids are co-extracted along with the target pesticides. The presence of these co-extracted substances, particularly the phospholipids, can lead to chromatographic interference, contamination of the GC injector and column, and contamination of the mass spectrometer itself. To avoid these complications, a cleanup step is recommended prior to the instrumental analysis. This is typically performed using dispersive SPE with mixed sorbents, often with cumbersome multi-step centrifugation. In this study an Oasis PRiME HLB Cartridge was used for a simple pass-through cleanup to effectively remove fats and phospholipids. This method was applied to a number of pesticides registered for use on avocado in various world markets and suitable for GC-MS analysis. The APGC methodology was used for quantitative analysis in this study (APGC-MS/MS).

Experimental

GC conditions

GC system:	Agilent 7890
Column:	Restek Rxi-5 ms, 30 m x 0.25 mm x 0.25 μ m
Flow rate:	1.0 mL/min Helium
Injection volume:	1 μ L (15:1 split)
Temperature program:	80 °C initial, hold for 0.5 min, 12 °C/min to 320 °C

and hold for 8 min

MS conditions

Mass spectrometer:	Xevo TQ-S
Ion mode:	API+ (charge transfer mode)
Corona:	2.8 μ A
Source temperature:	150 °C
Probe temperature:	450 °C
Cone gas:	170 L/Hr
Auxiliary gas:	250 L/Hr
Collision gas:	0.15 mL/min (Ar)
Nebulizer:	4.0 bar
Data management:	MassLynx v4.1

Other instrument parameters are presented in Table 1.

Compound	MRM	Cone (V)	Collision (eV)	RT (min)
Azoxystrobin	403.0>344.2	20	12	20.9
	403.0>329.1	20	32	
Carfentrazone-ethyl	410.9>312.2	20	20	16.1
	410.9>340.2	20	10	
Chlorothalonil	265.9>170.0	20	27	12.0
	265.9>230.9	20	27	
Cypermethrin	163.1>127.0	20	10	19.3*
	163.1>127.0	20	10	
λ -Cyhalothrin	449.0>181.2	20	20	17.8*
	449.0>197.3	20	14	
Cyprodinil	225.1>210.1	20	20	13.9
	225.1>93.1	20	32	
Dichlorvos	184.9>93.0	20	20	6.1
	220.9>109.0	20	15	
Fenpropathrin	349.1>265.2	20	10	17.1
	349.1>210.2	20	20	
Fludioxonil	248.0>127.1	20	25	14.9
	248.0>182.1	20	20	
Folpet	259.9>130.0	20	13	14.3
	294.9>259.9	20	10	
Malathion	173.1>127.1	20	6	13.2
	173.1>99.0	20	10	
Metalaxyl	206.1>132.1	20	20	12.8
	206.1>162.1	20	8	
Oxyfluorfen	361.0>300.1	20	10	15.1
	361.0>252.2	20	30	
Permethrin	183.1>153.0	20	15	18.6*
	183.1>168.0	20	15	
Pyriproxyfen	136.1>78.0	20	20	17.6
	136.1>96.0	20	20	
Simazine	201.1>173.1	20	10	11.2

Avocado is so high in fat, the AOAC QuEChERS method is modified to reduce the sample size from 15 g to 5 g. Weigh 5 g sample into a 50 mL centrifuge tube (for a spiked sample, add the required volume of spiking standard solution). Add 5 mL water and 15 mL 99:1 acetonitrile/acetic acid. Vortex for 30 seconds and shake well for 2 minutes. Add QuEChERS salts (contents of DisQuE pouch for AOAC, p/n 186006812). Shake the tube vigorously by hand for 1 minute and centrifuge at approximately 2500 rcf for 5 minutes. An aliquot of the supernatant extract (top layer) is taken for analysis.

Pass-through SPE Cleanup: Install an Oasis PRiME HLB Cartridge (3 cc, 60 mg, p/n 186008056) on a vacuum manifold. Set to minimal vacuum (~2 in Hg). Pass 0.4 mL of the QuEChERS extract through the cartridge to waste. Install collection vessels. Pass 0.6 mL of the QuEChERS extract and collect. In this study for APGC-MS analysis 200 µL of the collected extract is transferred to a Qsert Vial and analyzed directly. Alternatively, a portion of the collected extract can be evaporated and reconstituted in toluene for splitless GC injection.

Results and Discussion

AOAC QuEChERS extraction

It is important to distinguish any recovery losses resulting from the SPE cleanup from losses resulting from the initial QuEChERS extraction. Therefore, the modified QuEChERS procedure was evaluated for recovery of the target compounds prior to any SPE recovery experiments. All compounds (spiked at 40 ng/g) were recovered at greater than 80% with the exception of folpet (70%), fenpropathrin (65%), and pyriproxyfen (75%).

Oasis PRiME HLB cleanup

SPE cleanup recovery data (see Figure 1) were determined using blank avocado samples obtained using the modified QuEChERS protocol. Blank extracts were spiked at the 9 and 40 µg/kg (ppb) levels and were subjected to the pass-through SPE cleanup protocol. Response for each compound was compared with response obtained from identical blank sample extracts spiked after the SPE cleanup. Only folpet, a thermal and pH labile substance, showed recovery losses greater than 20% resulting from the cleanup protocol. Figure 2 shows that the Oasis PRiME HLB Cartridge cleanup removed greater than 95% of phospholipids from the avocado QuEChERS extract. Also, greater than 90% of chlorophyll and approximately 80% of fat was removed in the cleanup. Cleanup obtained in only seconds with this protocol was comparable to

traditional dispersive SPE (dSPE) cleanups that often require multiple cumbersome centrifugation steps.




Figure 1. SPE cleanup recovery results.




Figure 2. Effective removal of phospholipids from avocado QuEChERS extract with Oasis PRiME HLB cleanup.

APGC-MS/MS

Compared with traditional electron-impact mass spectrometry (EI-MS), the API ionization used for APGC is a softer form of ionization and is often a superior technique for tandem MS. The softer ionization results in less in-source fragmentation and a greater likelihood of obtaining a molecular ion for subsequent fragmentation in the collision cell. An example is shown in Figure 3, a comparison of the API+ and EI+ mass spectra

obtained for λ -cyhalothrin. Note the high abundance of the molecular ion (m/z 449) in the API spectrum compared with the EI spectrum. In this study, two MRM transitions were monitored for determination of λ -cyhalothrin (see Table 1). Each of these transitions resulted from fragmentation of the m/z 449 molecular ion; these transitions would not be possible using EI-MS. Figure 4 shows the APGC-MS/MS determination of the isomers of λ -cyhalothrin spiked at 9 ng/g in avocado after QuEChERS extraction and Oasis PRiME HLB cleanup. Similar low ng/g (ppb) detection limits were observed for all the target compounds in this study.




Figure 3. Comparison of mass-spectra for λ -cyhalothrin; EI+ (left) nearly devoid of any molecular ion but API+ (right) shows significant abundance of the molecular ion at m/z 449.




Figure 4. APGC-MS/MS chromatograms showing λ -cyhalothrin isomers from an avocado sample spiked at 9 ng/g (ppb); MRM transitions were obtained from fragmentation of the molecular ion at m/z 449.

Conclusion

- A modified QuEChERS procedure was effective for extraction of pesticides from avocado prior to APGC-MS/MS analysis.
- Oasis PRiME HLB pass-through cleanup was effective for removal of fats and phospholipids from the QuEChERS extracts.
- Cleanup and recovery obtained using the Oasis PRiME HLB Cartridge was comparable to that obtained from cumbersome and time-consuming multi-step dispersive SPE.
- The Xevo TQ-S Mass Spectrometer operated in APGC-MS/MS mode was effective for low ppb

determination of pesticides in avocado.

Featured Products

Waters Atmospheric Pressure Gas Chromatography (APGC) <<https://www.waters.com/10100362>>

Xevo TQ-S <<https://www.waters.com/10160596>>

MassLynx MS Software <<https://www.waters.com/513662>>

720005816, October 2016