

# An Example of a Particle Selectivity Difference between CORTECS C8 and BEH C8 Columns

Jennifer M. Nguyen, Thomas Swann

Waters Corporation



Abstract

This application note describes rapid separation of samples containing structurally similar compounds. It compares the separation of seven structurally similar catechins on an XBridge BEH  $C_8 XP$  Column, based on fully-porous hybrid particles, and a CORTECS  $C_8$  Column, based on solid-core silica particles.

#### Benefits

- · Increases in selectivity.
- · Increases in efficiency.
- · Up to a 290% increase in resolution.

## Introduction

Structurally similar compounds can be difficult to separate, making the development of rapid chromatographic methods for them very challenging. Changing the column stationary phase is one important tool to provide increased resolution. Using smaller and/or solid-core particles can improve efficiency. However, another source of increased resolution can be selectivity differences inherent in the stationary phase. While it is most common to vary the bonded phase chemistry, selectivity differences may also be achieved by varying the base particles. This application shows an example of this effect by comparing the separation of seven structurally similar catechins on an XBridge BEH C<sub>8</sub> *XP* Column, based on fully-porous hybrid particles, and a CORTECS C<sub>8</sub> Column, based on solid-core silica particles.

## Experimental

A sample containing (-)-epigallocatechin (200  $\mu$ g/mL), (+)-catechin (100  $\mu$ g/mL), (-)-epigallocatechin gallate (100  $\mu$ g/mL), (-)-epicatechin (100  $\mu$ g/mL), (-)-epicatechin gallate (200  $\mu$ g/mL), and (-)-catechin gallate (100  $\mu$ g/mL) was prepared in 1:1 v/v methanol:water.

Instrument:

ACQUITY UPLC H-Class

Data management:

Empower 3 CDS

#### Method conditions

Columns:	XBridge BEH C <sub>8</sub> <i>XP</i> , 2.5 μm, 3.0 x 50 mm (p/n: 186006045)				
	CORTECS C <sub>8</sub> , 2.7 μm, 3.0 x 50 mm (p/n: 186008359)				
Mobile phase A:	Water				
Mobile phase B:	Methanol				
Mobile phase C:	2% Formic acid in water (autoblended to 0.1% formic acid)				
Gradient:	75:20:5 of A:B:C to 45:50:5 of A:B:C in 2.50 min (2.5 μm column) 75:20:5 of A:B:C to 45:50:5 of A:B:C in 2.70 min (2.7 μm column)				
Flow rate:	0.85 mL/min (2.5 μm column) 0.79 mL/min (2.7 μ m column)				
Column temp.:	25 °C				
Detection (UV):	280 nm				
Injection volume:	1 μL				

## **Results and Discussion**

Catechins are a family of flavan-3-ol polyphenolic compounds found in many plants which are associated with a variety of health benefits. A mixture of green tea catechins was prepared and included different isomer groups of catechin and gallocatechin, plus some of their gallate derivatives as shown in Figure 1.



Figure 1. Structures of seven catechins.

This mixture was separated using an XBridge BEH  $C_8 XP$ , 2.5 µm, 3.0 x 50 mm Column (p/n: 186006045) with a rapid method where all compounds elute in under 2 minutes, Figure 2. Components 4 and 5 formed a co-eluting critical pair with these conditions.



Figure 2. Separation of catechins on an XBridge BEH  $C_8$  XP, 2.5  $\mu$ m, 3.0 x 50 mm Column. However, when using a column containing the same  $C_8$  chemistry but based on CORTECS solid-core silica particles, the critical pair was well separated, as shown in Figure 3. As anticipated, a resolution improvement from narrower peaks occurred due to the higher efficiency of the CORTECS Column, but the observed increase was larger than expected.



Figure 3. Separation of catechins on a CORTECS C\_8 2.7  $\mu\text{m}$ , 3.0 x 50 mm Column.

In order to assess factors besides the solid-core efficiency advantage that could influence resolution, relationships between resolution and column-analyte parameters need to be considered. From isocratic chromatography theory, resolution ( $R_s$ ) arises from the combination of column efficiency (plate count, N), analyte selectivity ( $\alpha$ ), and analyte retention (retention factor, k). There are several equations that can be derived to express such relationships but one of the most common is by Purnell, eq 1. The subscripts indicate that, for this equation, the plate count and the retention factor for the *second peak of an analyte pair* are needed to estimate resolution. One can determine N,  $\alpha$ , and k for pairs of analyte peaks in an isocratic chromatogram and utilize such a relationship to parse out the relative contributions to resolution from these three sources.

However, the catechins mixture separation used here is a gradient method. In linear gradient chromatography, the elution strength of the mobile phase is constantly increasing and the gradient peak properties of efficiency, selectivity, and retention are likewise changing. At any given moment, there are instantaneously isocratic values of *N*,  $\alpha$ , and *k* for each peak, but such values are not directly measureable. All that can be observed from a gradient chromatogram are the average values of these parameters which are represented here as  $N_2^*$ ,  $\alpha^*$ , and  $k_2^*$ . Substitution into eq 1 provides the version of the derived Purnell resolution equation for gradient chromatography, eq 2.

$$R_{s} = \left(\frac{\sqrt{N_{2}}}{4}\right) \left(\frac{\alpha - 1}{\alpha}\right) \left(\frac{k_{2}}{1 + k_{2}}\right)$$
<sup>(2)</sup>

$$R_{s} = \left(\frac{\sqrt{N_{2}^{*}}}{4}\right) \left(\frac{\alpha^{*} - 1}{\alpha^{*}}\right) \left(\frac{k_{2}^{*}}{1 + k_{2}^{*}}\right)$$

Equation 2 plus eqs 3, 4, 5, and 6 for the values of efficiency (USP definition), retention factor, selectivity, and measured resolution (USP definition) were employed in this application note to compare the estimated gradient resolution (from eq 2) to the measured gradient resolution (from eq 6). After examining these values, the contributions of each term in eq 2 to the resolution differences observed for the critical pair were then examined.

(3)

(1)

$$N_i^* = 16 \left(\frac{t_{ri}}{W_i}\right)^2$$

For peak *i* with retention time of  $t_{ri}$  and width  $W_i$  at the baseline between tangent lines drawn at half height.

(4)

$$k_i^* = \frac{t_{ri} - t_0}{t_0}$$

For peak *i* at retention time of  $t_{ri}$  and an unretained peak at retention time  $t_0$ .

(5)



For consecutive peaks *m* and *n*.

(6)

$$R_s = 2\left(\frac{t_{rn} - t_{rm}}{W_n + W_m}\right)$$

For consecutive peaks m and n with peak widths W at baseline between tangent lines drawn at half height.

With the CORTECS C<sub>8</sub> Column, a measured  $R_s = 1.82$  was obtained for critical pair components 4 and 5 from the chromatogram in Figure 3 and eq 6. Values of  $N^{*}_{2}$  and  $k^{*}_{2}$  (for the second peak, component 5) plus  $\alpha^{*}$  from this chromatogram are noted in Table 1. With eq 2, the three terms could then be calculated to give the efficiency term,  $N^{*}= 25.8$ , selectivity term,  $\alpha^{*} = 0.115$ , and retention term,  $k^{*} = 0.609$ ; see Table 1. This provided an estimated resolution of  $R_{s} = 1.81$  which closely matches the measured resolution.

Since the XBridge BEH  $C_8 XP$  Column exhibits co-elution for critical pair components 4 and 5, individual injections were performed under the same conditions. These chromatograms are shown in Figure 4. Using eq 6, a measured resolution of  $R_s = 0.47$  was obtained. The CORTECS  $C_8$  Column therefore showed a 290% increase in measured resolution for the critical pair over the XBridge BEH  $C_8 XP$  Column. From these chromatograms, values of  $N_2^*$  and  $k_2^*$  (for the second peak, component 5) plus  $\alpha^*$  were obtained and are shown in Table 1.

Again, using eq 2 gave the calculated efficiency term,  $N^* = 19.9$ , selectivity term,  $\alpha^* = 0.035$ , and retention term,  $k^* = 0.590$ ; see Table 1. This provided an estimated resolution of  $R_s = 0.41$ . This is similar to the measured resolution of  $R_s = 0.47$ .



Figure 4. Individual injections for critical pair components 4 and 5 on an XBridge BEH  $C_8$  XP, 2.5  $\mu$ m, 3.0 x 50 mm Column.

	N <sub>2</sub> *	$N^*$ term $\left(\frac{\sqrt{N_2^*}}{4}\right)$	α*	$\alpha^*$ term $\left(\frac{\alpha^*-1}{\alpha^*}\right)$	<b>k</b> <sup>*</sup> <sub>2</sub>	$\frac{k^* \text{term}}{\left(\frac{k_2^*}{1+k_2^*}\right)}$	Estimated R <sub>s</sub> (eq 2)	Measured R <sub>s</sub> (eq 6)
XBridge BEH C <sub>s</sub> <i>XP</i> 3.0 x 50 mm, 2.5 μm	6,350	19.9	1.04	0.035	1.44	0.590	0.41	0.47
CORTECS C <sub>8</sub> 3.0 x 50 mm, 2.7 μm	10,700	25.8	1.13	0.115	1.56	0.609	1.81	1.82
Percentage change		30%		230%		3%	340%	290%

Table 1. N\*,  $\alpha^*$ , k\* values and terms plus subsequent estimated R<sub>s</sub> values compared to the measured

 $\rm R_{s}$  values for compounds 4 and 5 using XBridge BEH  $\rm C_{8}$  XP and CORTECS  $\rm C_{8}$  Columns.

An Example of a Particle Selectivity Difference between CORTECS C8 and BEH C8 Columns

Finally, after calculating each resolution contribution ( $N^*$ ,  $\alpha^*$ , and  $k^*$  terms), the factor that increases critical pair resolution most between the XBridge BEH  $C_8 XP$  and CORTECS  $C_8$  columns can be determined. The efficiency (  $N^*$ ) term increases by 30%, the selectivity ( $\alpha^*$ ) term by 230% and the retention ( $k^*$ ) term by 3% when switching to the CORTECS  $C_8$  Column. Thus, while it was anticipated that the additional efficiency of the CORTECS Column would help increase the resolution, the selectivity term was found to be far more important in the 290% increase in measured resolution. This separation of structurally similar catechins is a good example of how the difference in selectivity between base particles, not bonded phase chemistries, can significantly benefit the resolution of a critical pair of analytes.

#### Conclusion

Optimizing resolution can be challenging for rapid separations of samples containing structurally similar compounds. Resolution is affected by three factors – efficiency, selectivity, and retention. High efficiency CORTECS  $C_8$  Columns can directly improve resolution through the efficiency factor. However, as seen in this separation of catechins, the selectivity factor may have a larger impact on resolution. The most common way to alter selectivity is by changing the bonded phase chemistry. Alternatively, a column containing the same bonded phase on different base particles can be tried. The separation of catechins illustrates well the potential selectivity difference for two different base particles. By using columns based on CORTECS solid-core silica particles instead of XBridge BEH hybrid particles, different selectivity, and thus higher resolution could be achieved even when both particles had the same  $C_8$  bonding chemistry. This resulted in a better separation of these structurally similar compounds.

#### References

- 1. Flow rates were scaled using the Waters Column Calculator Version 2.0. Download and release notes: http://www.waters.com/waters/support. htm?lid=134891632&type=DWNL
- 2. Foley, J. P. Resolution Equations for Column Chromatography. Analyst. 1991, 116, 1275–1279.
- Snyder, L. R.; Dolan, J. W.; Chap 2. Gradient Elution Fundamentals. *High-Performance Gradient Elution*; John Wiley & Sons Inc.: Hoboken, New Jersey, 2007; p39.

# Featured Products

- ACQUITY UPLC H-Class PLUS System <https://www.waters.com/10138533>
- Empower 3 Chromatography Data Software <a href="https://www.waters.com/10190669">https://www.waters.com/10190669</a>
- Auto-Blend Plus <https://www.waters.com/134623262>

720006038, July 2017

 $\wedge$ 

©2019 Waters Corporation. All Rights Reserved.