

Nota de aplicación

Comparison of Electrospray and UniSpray for Pharmaceutical Compounds

Kerri M. Smith, Paul D. Rainville

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

Abstract

This technical brief shows the comparison of ionization of a selection of pharmaceutical compounds by electrospray (ESI) and impactor ionization (UniSpray) as well as their robustness in human serum.

Benefits

Ionization plays a critical role in analysis by mass spectrometry with multiple ionization techniques available. Here we describe an impactor ionization source and its application to the analysis of pharmaceutical drugs.

Introduction

Impactor ionization is the formation of ions by directing a heated nebulized spray of liquid onto a surface with an applied voltage.¹ The spray is aimed off center and on impact, the ions flow downstream in a path that follows the curvature of the surface (in this case a pin), called the Coandă effect (Figure 1). Commonly used electrospray ionization involves a heated high velocity spray from a charged capillary. These two techniques have similar effects of ionization, producing predominantly $[M+H]^+$ (or $[M-H]^-$) ions, yet their mechanisms appear to be different. This work shows the comparison of ionization of a selection of pharmaceutical compounds by electrospray (ESI) and impactor ionization (UniSpray) as well as their robustness in human serum.




Figure 1. Graphic representation and photo of UniSpray source.

Experimental

LC conditions

LC system:	ACQUITY UPLC I-Class FTN
Vials:	Waters TruView LC-MS Certified, Total Recovery Vial [P/N 186005663CV]
Column:	ACQUITY BEH C ₁₈ , 2.1 x 50 mm, 1.7 µm [P/N 186002350]
Column temp.:	30 °C
Sample temp.:	5 °C
Injection volume:	3 µL
Flow rate:	0.600 mL/min
Mobile phase A:	H ₂ O, 0.1% formic acid
Mobile phase B:	Acetonitrile, 0.1% formic acid
Gradient:	1 to 95 %B over two minutes, hold for 0.5 minutes

MS conditions

MS system:	Xevo TQ-S Mass Spectrometer
Acquisition mode:	MRM
Ionization type:	Electrospray and UniSpray

Ionization mode:	+/-
Capillary voltage:	1 kV/2 kV POS/NEG
Desolvation temperature:	550 °C
Desolvation flow:	1100 L/hr

Results and Discussion

Sample description

Compound libraries with molecular weights ranging from 151 to 824 were purchased from Enzo Life Sciences. The standards were diluted to 100 nM and 1 nM, or 0.1 nM with 30:70:0.1 ACN/water/formic acid for high-throughput compound optimization of $[M+H]^+$ or $[M-H]^-$ by QuanOptimize followed by injection on column. Plasma robustness studies were performed by protein precipitation with ACN at a 3:1 ratio, vortex mixed, and then centrifuged at 16.1 x g for 10 minutes. The supernatant was taken and standards were spiked to 1 nM before injection. An ACQUITY UPLC I-Class System with a BEH C₁₈ Column (2.1 x 50 mm) and mobile phases of water and ACN with 0.1% formic acid was coupled to a Xevo TQ-S Mass Spectrometer equipped with ESI and UniSpray ionization sources. Data was collected in both positive and negative ionization modes and processed using MassLynx v.4.1. All results were calculated on chromatographic peak area.

Tuning the UniSpray Source

The UniSpray source was tuned using 50% mobile phases A and B at analytical flow combined with 10 µL/min of 10 ng/mL verapamil. The capillary protrusion was first optimized for stable spray followed by positional adjustment of the capillary with respect to the pin. The highest signal is seen when the spray is directed to the right of center on the pin, as shown in Figure 1. It should be noted that the following data is from a single tune position. Custom tuning on specific target compounds may yield better sensitivity for compounds that initially showed a poor response compared to ESI.

A compound library consisting of 156 compounds was automatically tuned by flow injection using the MassLynx application manager, QuanOptimize. Following MRM tuning, the compounds were screened using a short chromatographic gradient with data collected in positive as well as negative ionization modes of ESI

and UniSpray. The data were processed in TargetLynx using minimal smoothing. Over 70% of the compounds had greater or comparable peak area response in USI compared to ESI (defined as the ratio of USI to ESI response above 1.0) and are summarized in Figure 2. Plots of USI/ESI response with respect to such chemical properties as pKa and LogP showed no discernible trend (Figure 3). This, with the correlation results in Figure 3C, suggests UniSpray ionization is not discriminatory and behaves similarly to ESI for the compounds in the conditions tested. To test ionization efficiency at lower concentration levels, samples were further diluted to 0.1 nM. Figure 4 shows two examples, amantadine and amlodipine, where increased ionization was observed in Unispray ionization relative to ESI. It is believed that the combination of the droplet size after impact on the pin as well as the Coanda effect and other contributions that help to enhance desolvation aid in the increased ionization and sampling efficiency of UniSpray ionization. For a more detailed discussion, please visit www.waters.com/unispraymechanisms.




Figure 2. Chromatographic peak area ratios USI/ESI showing UniSpray ionization response compared to electrospray.




Figure 3. Plots of USI and ESI chromatographic peak area ratio with respect to pKa (A) and LogP.(B) as well as USI area vs. ESI area (C).




Figure 4. Ionization of amantadine and amlodipine at 0.1 nM comparing ESI (top trace) and USI (bottom trace) in solvent.

To test the robustness of UniSpray ionization for matrix samples, a subset of compounds from the library were spiked into acetonitrile-precipitated human serum. Five vials, each with a different set of compounds, were injected in sequence onto the chromatographic column. Peak areas for injections 6–1800 are plotted and then summarized in Figures 5 and 6 representing nearly six days of continuous operation. The RSD for the 37 compounds tested ranged from 1.9–16. Higher RSD was observed for compounds that were at the lower end of detection.




Figure 5. Summary of compounds tested for robustness in human serum up to 1800 injections.






Figure 6. Representative human serum robustness results for amiloride (top), verapamil (middle), and gabapentin (bottom); Every 5th injection is plotted.

Conclusion

UniSpray is an ionization technique that produces results that are qualitatively similar to those of electrospray, however, increased desolvation and subsequent sampling of ions in the source enhanced the signal of the majority of the compounds in this study. The chromatographic peak area ratio of Unispray ionization to ESI ranged from 0.3 to 11.4 and was an overall average of 2.1 times greater in UniSpray for the 156 compounds tested. Finally, system robustness in human serum was tested over six days of continuous operation with peak area RSDs ranging from 1.9–16%.

References

1. Lubin A, Bajic S, Cabooter D, Augustijns P, Cuyckens F. Atmospheric Pressure Ionization Using a High Voltage Target Compared to Electrospray Ionization. *J Am Soc Mass Spectrom.* 2017;28(2):286-93.
-

Featured Products

ACQUITY UPLC I-Class PLUS System <<https://www.waters.com/134613317>>

Xevo TQ-S <<https://www.waters.com/10160596>>

UniSpray Ion Source <<https://www.waters.com/134891755>>

MassLynx MS Software <<https://www.waters.com/513662>>

TargetLynx <<https://www.waters.com/513791>>

©2019 Waters Corporation. All Rights Reserved.