

Application Note

ACQUITY UPLC I-Class/Xevo TQD IVD System: Analytical Performance for an Alkylating Agent

Waters Corporation

Este é um Resumo de aplicações e, por isso, não inclui uma seção de experimento detalhada.

Para uso em diagnóstico *in vitro*. Não está disponível para venda em todos os países.

Introduction

The Waters ACQUITY UPLC I-Class/Xevo TQD IVD System enables the quantification of organic compounds in human biological liquid matrices.

This document describes a test of the analytical performance of the ACQUITY UPLC I-Class/Xevo TQD IVD System for the analysis of busulfan in plasma.



ACQUITY UPLC I-Class/Xevo TQD IVD System.

Experimental

The ACQUITY UPLC I-Class/Xevo TQD IVD system was controlled by MassLynx IVD (v4.1) and the data processed using the TargetLynx Application Manager. Calibrators and Quality Controls were prepared by spiking commercially available reference material in plasma and the samples were processed using the following conditions:

Sample Preparation Conditions

50 μ L sample was processed with methanol and centrifuged, then subsequently diluted with water prior to analysis.

LC Conditions

Column:	ACQUITY UPLC HSS T3 (IVD) 1.8 μ m, 2.1 mm x 50 mm
Mobile phase A:	2 mM Ammonium acetate+0.1%

	formic acid in water
Mobile phase B:	2 mM Ammonium acetate+0.1% formic acid in methanol
Flow rate:	0.6 mL/min
Gradient:	10% B for 0.5 minutes, 10–98% B over 1.0 minute, 98% B for 0.5 minutes, 10% B for 0.5 minutes

MS Conditions

Resolution:	MS1 (0.75 FWHM), MS2 (0.75 FWHM)
Acquisition mode:	MRM
Polarity:	ESI (+)

Results and Discussion

Performance characteristics of busulfan on the ACQUITY UPLC I-Class/Xevo TQD IVD System are shown in Table 1. A chromatogram illustrating the analytical sensitivity of the busulfan analysis is shown in Figure 1.

Compound	Range (µg/mL)	LLOQ (µg/mL)	%RSD at LLOQ	Total precision	Repeatability
Busulfan	0.025–5	0.02	16.0	≤7.3%	≤5.1%

Table 1. Performance characteristics of busulfan. Range defined by linear fit where $r^2 > 0.99$. LLOQ defined by $S/N (PtP) > 10$ and $\%RSD \leq 20\%$. $\%RSD$ at LLOQ determined through analytical sensitivity experiments performed over five occasions ($n=50$). Total precision and repeatability of QCs performed over five occasions in plasma ($n=25$).




Figure 1. Chromatogram showing the analysis of busulfan using the ACQUITY UPLC I-Class/Xevo TQD IVD System.

Conclusion

The Waters ACQUITY UPLC I-Class/Xevo TQD IVD System has demonstrated the capability to deliver analytical

sensitivity and precision for the analysis of busulfan in plasma.

Disclaimer

The analytical performance data presented here is for illustrative purposes only. Waters does not recommend or suggest analysis of the analytes described herein. These data are intended solely to demonstrate the performance capabilities of the system for analytes representative of those commonly analyzed using liquid chromatography and tandem mass spectrometry. Performance in an individual laboratory may differ due to a number of factors, including laboratory methods, materials used, intra-operator technique, and system conditions. This document does not constitute a warranty of merchantability or fitness for any particular purpose, express or implied, including for the testing of the analytes in this analysis.

Featured Products

[MassTrak ACQUITY UPLC I-Class PLUS/Xevo TQD IVD System </nextgen/br/pt/products/in-vitro-diagnostics-solutions/lc-ms-ms-ivd-systems-portfolio/masstrak-acquity-uplc-i-class-plus-xevo-tqd-ivd-system.html>](#)

[Mass Spectrometry Software to Facilitate LC-MS Analysis in Clinical Laboratories <https://www.waters.com/waters/nav.htm?&cid=134834177>](#)

720006310, August 2018



© 2025 Waters Corporation. All Rights Reserved.

[Termos de uso](#) [Aviso de Privacidade](#) [Marcas comerciais](#) [Carreiras](#) [Avisos jurídicos e de privacidade](#) [Cookies](#) [Preferências de cookies](#)