

A Comparative Study Using Preparative Reverse Phase Liquid Chromatography and Supercritical Fluid Chromatography of Impurities from a Stressed Pharmaceutical Drug Substance

Janet Hammond, Russell J. Mortishire-Smith, Jayne Kirk, Christopher Howson

Waters Corporation, Syngenta UK Ltd Jealott's Hill International Research Centre, BioHub/Waters Open Access Laboratory Alderley Park

Abstract

This application note compares the use of both preparative RPLC and SFC techniques, discussing the scaling process from the analytical to preparative columns while illustrating the efficiency gains when using preparative SFC to isolate the impurities.

Benefits

- ACQUITY UPC² System results in improved resolution and a more desirable elution order of impurities, resulting in higher purity components with maximum yields
- Integration with the ACQUITY QDa Mass Detector maximizes productivity
- The SFC approach enables the purification to be completed in a time of two hours compared to 12.5 hours for the RP approach and used less than 63% of the solvent

- The Investigator SFC System is a versatile, cost effective system for pharmaceutical laboratories routinely performing achiral analysis and small scale purification
 - The workflow is ideal for the pharmaceutical industry where there is increased pressure to reduce of drug development process timelines
 - ACQUITY UPC² System avoids the use of non-volatile solvents that require extensive drying down
-

Introduction

In the pharmaceutical industry, rapid preparative isolation is required for a number of reasons. These range from the use of chromatographic markers for the purpose of structure elucidation, for determination of relative response factors, and for reference standards. Current regulatory requirements on the control of impurities and degradation products in drug substances and drug product, dictate that for impurities at levels greater than 0.1%, unambiguous characterization is required.¹ Modern, sophisticated, spectral hyphenated techniques, such as liquid chromatography mass spectrometry (LC-MS), liquid chromatography mass spectrometry-mass spectrometry (LC-MS/MS), and liquid chromatography nuclear magnetic resonance (LC-NMR), are already being extensively used by the industry, but in many instances, these techniques in isolation are not successful and an extra purification step is required to isolate the impurities for subsequent unambiguous characterization.²

Supercritical fluid chromatography (SFC) is a normal-phase chromatographic technique which found its place early in its development with chiral separations. In recent years, the improved quality of available hardware has led to a wider uptake of the technology for achiral separation applications, and has replaced reverse phase liquid chromatography (RPLC) separation techniques for the resolution of closely related species within the pharmaceutical and life science industries.³

The use of a super critical gas in chromatography, typically carbon dioxide due to the easily achievable conditions required to reach its critical point, (74 bar pressure and >32 °C temperature) offers several key benefits. Once above the critical point the liquid and gas phases are no longer distinct, meaning the super critical phase has solubilizing properties which are similar to a liquid with gas-like diffusivity. This translates chromatographically to lower back-pressures, allowing the use of higher flow rates, resulting in quicker analysis and more efficient chromatographic separations. SFC has also been praised as a 'green' separation technique eliminating the need for organic solvents such as heptane and hexane. Typical co-solvents used in SFC are alcohols (methanol, ethanol, or isopropanol) which are seen as a preferred solvent of choice

because they are recognized by most as green chemistry initiatives.

For preparative and semi-preparative applications, the use of SFC allows the majority of the mobile phase to be a gas under normal temperatures and pressures, meaning that the resulting fractions are lower volume organic solutions. This reduces the time and cost associated with solvent evaporation. The use of SFC for achiral applications also avoids the use of pH modifiers, ion pair reagents, and buffers, all of which can be left as residues in final compounds or may interact with chemically sensitive analytes on concentration and dry down.

Traditionally dominated by reverse phase liquid chromatography (RPLC), there is an increasing trend toward using supercritical fluid chromatography (SFC) to replace RPLC for purifications from the semi-preparative scale up to the kilogram scale.^{4,5}

In this application note we compare the use of both preparative RPLC and SFC techniques, discussing the scaling process from the analytical to preparative columns while illustrating the efficiency gains when using preparative SFC to isolate the impurities.

Experimental

RP UPLC method conditions

System:	ACQUITY UPLC H-Class with ACQUITY UPLC PDA Detector
Column:	BEH C ₁₈ 1.7 µm, 2.1 × 50 mm (P/N 186002350)
Column temp.:	50 °C
Sample temp.:	15 °C
Injection volume:	1 µL
Flow rate:	1 mL/min

Mobile phase A:	H ₂ O
Mobile phase B:	MeCN +0.1% NH ₄ OH
Gradient:	2–98%
CSH C ₁₈ , 1.7 μm, 2.1 × 50 mm, was also screened (P/N 186002140)	

Preparative HPLC conditions

Column:	BEH C ₁₈ 5 μm, 19 × 150 mm (P/N 186008166)
Column temp.:	50 °C
Sample temp.:	15 °C
Injection volume:	2.5 mL
Flow rate:	25 mL/min
Mobile phase A:	H ₂ O
Mobile phase B:	Acetonitrile +0.1% NH ₄ OH
Gradient:	50–60%
Time:	10 minutes

Instrumentation

Waters AutoPurification System: 2545 binary gradient module, 2767 sample manager, system fluidics organizer, 8 × 30 mL flow splitter, 2 × 515 pumps, 2998 photodiode array detector

ACQUITY QDa

Cone voltage: 10 V

Probe temp.: 600 °C

Ionization mode: ESI+

Sampling frequency: 10 Hz

Scan range: 100–600 amu

Wavelength: 254 nm

UPC² method conditions

System: ACQUITY UPC² with ACQUITY UPC² PDA
Detector

Column: TORUS 2-PIC 1.7 µm, 3 mm × 50 mm (P/N
186007600)

Column temp.: 45 °C

Sample temp.: 15 °C

Injection volume: 1 µL

Flow Rate: 2 mL/min

Mobile Phase A: CO₂

Mobile Phase B: Methanol

Gradient: 2–50%

ABPR: 1740 psi

UPC² method conditions

Other columns screened:	Torus DIOL 1.7 µm, 3.0 × 50 mm (P/N 186007609)
	Torus DEA 1.7 µm, 3.0 × 50 mm (P/N 186007618)
	Torus 1-AA 1.7 µm, 3.0 × 50 mm (P/N 186007627)

Preparative SFC conditions

System:	Investigator SFC System
Column:	Torus 2-PIC OBD Prep 5 µm, 10 mm × 150 mm (P/N 186008583)
Column temp.:	45 °C
Sample temp.:	15 °C
Injection volume:	0.5 mL
Flow rate:	15 mL/min
Mobile phase A:	CO ₂
Mobile phase B:	Methanol
Gradient:	2–50%
ABPR:	1850 psi
Time:	4 minutes

Sample preparation

For RP prep 20 mg/mL in 90% water: 10% acetonitrile

For SFC prep 20 mg/mL in methanol

Data management


MassLynx Software v4.1 and ChromScope

Results and Discussion

For this study, the active pharmaceutical ingredient (API) fenofibrate was subjected to accelerated stressed conditions. When the resulting sample was analysed by LC-MS, two new impurities had been generated. RP-UPLC/MS analysis of the stressed material was performed on an ACQUITY UPLC H-Class System and ACQUITY QDa Mass Detector and the resulting chromatogram showed impurity 1 $[M+H]^+$ 333, impurity 2 $[M+H]^+$ 376, and the parent Figure 1A. The two new impurities elute before the main component and are just baseline resolved, limiting the sample loading which can be applied on the column.

1A)

BEH C₁₈ 1.7 μ m, 2.1 x 50 mm



1B)

BEH C₁₈ 5 μ m, 30 x 150 mm




Figure 1. UPLC chromatograms showing parent and impurities 1 and 2 of the stressed sample A) analytical column, B) preparative column.

The comparative ACQUITY UPC² System analysis, because of the different column selectivity, showed the main component eluting before the impurities Figure 2A. This allows for the collection of higher purity components with reduced loss of material. The scaling of analytical methods for preparative analysis is a well-documented process for reversed phase chromatography.⁶ Important factors to consider are:

- Column length to particle size: (L/dp)
- Matching this ratio between analytical and prep to enable for equal 'resolving power'

- Matching linear flow velocities (column ID) and runtime (column length)


Many of the same principles for RPLC are applicable in SFC, but with the addition of one key parameter, density, which needs to be included to provide robust and reproducible scaling in SFC. The density profile of the chromatographic system varies naturally throughout a gradient run, as the proportions of co-solvent change, and because the supercritical fluid is, by its nature, a compressible fluid unlike reversed phase solvents. This affects the thermodynamic interactions in the mobile phase, and thus the retention times of analytes, and it is therefore crucial to match density profiles when transferring SFC methods.

This in turn poses a new challenge, as when moving from an analytical method to preparative method, a number of key properties change which will directly affect the density profile; these include changes in particle size, column length, and column I.D. Density profiles are, therefore, adjusted using the backpressure regulator setting to obtain as near to possible matching average pressures between preparative and analytical conditions. In practical terms, this usually means raising the preparative run back-pressure setting, as the larger column with large particle sizes acts to effectively increase the density within the system.⁶

In this application Waters Torus SFC columns on the analytical scale were initially screened with methanol and isopropanol for 'best hit' separations prior to optimization.

From this screening process, it was evident that the Torus 2-PIC column chromatogram was found to provide the selectivity and resolution required in Figure 2A. The method was transferred to the SFC Investigator System using a Torus 2-PIC 5 μm , 10 mm \times 150 mm Column with methanol co-solvent and the separation is shown in Figure 2B. The scaled up method maintains the ratio between the column length and particle size, which is approximately 30,000 for both columns. The linear velocities between the analytical (3 mm I.D. and length 50 mm) and preparative column (10 mm I.D. and length 150 mm) were also preserved. In order to maintain the mobile phase density scaling from the analytical to preparative system, the average back-pressure regulator (ABPR) was adjusted from 1740 psi to 1850 psi. This analytical to preparative scaling process is summarized in the work flow shown in Figure 3.

2A) Torus 2 PIC, 1.7 μm , 3 x 50 mm



2B) Torus 2 PIC, 5 μm , 10 x 150 mm




Figure 2. UPC² chromatograms showing parent and impurities 1 and 2 of the stressed sample A) analytical column, B) preparative column.




Figure 3. Workflow summary for SFC analytical to preparative scaling process.

The column loading capacity depends on a number of variables which need to be assessed. These variables include the solubility and the resolution of the analytes involved in the separation, as well as the dimensions of the column. In this assessment, the 19 mm internal diameter (I.D.) preparative RP column has almost four times higher capacity than the 10 mm I.D. preparative SFC column. As a consequence, collection of the required amount of each impurity resulted in the RP experiments only requiring two repeat injections of 2.5 mL and the SFC prep requiring ten repeat 0.5 mL injections as seen in Figure 4. The effect of the lower flow rate of 11 mL/min, the peak width of 0.2 min, and a large portion of the eluent being CO₂, the volume of solvent from each SFC run was 4 mL, totaling 40 mL which needed to be reduced to dryness for further

work. Although the RP experiments only needed two repeat injections, a higher flow rate of 25 mL/min, the peak width of 0.5 minutes, and a completely liquid eluant resulted in 60 mL of solvent collected for each peak totaling 120 mL. The composition of the two extracts was different. The SFC extract contained only volatile methanol which was quickly reduced to dryness on a Genevac in less than one hour. In contrast, the RP extract was a mixture of nonvolatile aqueous buffers and acetonitrile which required twelve hours to reduce to dryness. The recoveries of the two impurities were higher from the SFC approach due to the sharper peak shape and the more favorable elution order.

Using the SFC approach enabled the entire purification to be completed in a time of two hours compared to 12.5 hours for the RP approach and used less than 63% of the solvent.

Full structural elucidation of the two impurities was successfully completed using a combination of HRMS and NMR studies (not reported herein).

Total of 100 mgs of forced degraded material put through RP/UPLC and SFC purification approaches

RP – UPLC	SFC
Sample prepared and processed across two repeat injections of 2.5 mL (<30 mins)	Sample prepared and processed across 10 repeat injections of 0.5 mL (<70 mins)
Approximately 60 mL of volume collected per peak per run which required reducing to dryness	Approximately 4 mL of volume collected per peak per run which required reducing to dryness
Genevac time approximately 12 hours	Genevac time approximately <1 hours
3.3 mg of Impurity 1 isolated (72% recovery, >95% purity)	4.1 mg of Impurity 1 isolated (89% recovery, >95% purity)
2.8 mg of Impurity 2 isolated (64% recovery, >84% purity)	3.5 mg of Impurity 2 isolated (82% recovery, >95% purity)
16 x 125 mm TT collection	Bottle collection

Figure 4. Comparison between the SFC and RP purification approach considering solvent volume, time, peak purity, and recoveries.

Conclusion

- The SFC approach reduced the total turnaround time of prep to dry material from 12.5 hours to 2 hours
- There was 63% less solvent used in the SFC purification and the co-solvent methanol was removed more easily than the RP mobile phase solvents
- Having a systematic workflow delivers an efficient, robust and scalable SFC separation from analytical to preparative configurations

References

1. ICH Guidelines Q3A (2R) Impurities in new drug substances.
2. A. Aranyi. Progress in *Pharmaceutical and Biomedical Analysis*, 4 (Identification and Determination of Impurities in Drugs). 2000, 240–251.
3. C.J. Hudella, P.D. McDonald. New Separation Tool for a Broad Range of Analytical Challenges, Waters Corp, *Chromatography Today Nov/Dec* 2012.
4. J. David Pinkston, Dong Wen, Kenneth L. Morand, et al. Comparison of LC/MS and SFC/MS for Screening of a Large and Diverse Library of Pharmaceutically Relevant Compounds, *Anal. Chem.*, 2006, 78 (21), pp 7467–7472.
5. G. Guiochon, A.Tarafder. Fundamental Challenges and Opportunities for Preparative Supercritical Fluid Chromatography, *Journal of Chromatography A* 1218, 2011, 1037–1114.
6. *Practical HPLC Method Development 2nd edn*. John Wiley & Sons, New York, 2008.

Featured Products

ACQUITY QDa Mass Detector <<https://www.waters.com/134761404>>

ACQUITY UPC2 System <<https://www.waters.com/134658367>>

ACQUITY UPLC H-Class PLUS System <<https://www.waters.com/10138533>>

ACQUITY UPLC PDA Detector <<https://www.waters.com/514225>>

Investigator SFC System <<https://www.waters.com/10145751>>

720006327, July 2018

©2019 Waters Corporation. All Rights Reserved.