# Waters™

#### 應用手冊

# Analysis of Amikacin and Kanamycin in Pharmaceutical Products with 3465 Electrochemical Detector

Jinchua Yang, Paul D.Rainville

Waters Corporation

# Abstract

Amikacin (AMI) and kanamycin (KAN) were analyzed by HPAEC-PAD according to the United States Pharmacopoeia (USP) Monographs. The instrument setup included an ACQUITY<sup>™</sup> Arc<sup>™</sup> System and a 3465 Electrochemical Detector (ECD). The instrument control and data acquisition were handled by an Empower<sup>™</sup> 3 Chromatography Data Software (CDS). The analytical performance met the USP system suitability requirements for both analytes in resolution, tailing factor, and repeatability. The 3465 ECD demonstrated an excellent linear response in concentration ranges that are sufficient for the assay of AMI and KAN. Amikacin injection samples were analyzed using this setup with satisfactory results, demonstrating the suitability of the ACQUITY Arc System and the 3465 ECD detector for the assay of amikacin and kanamycin in pharmaceutical products.

## Benefits

- ACQUITY Arc with 3465 ECD provides reliable analysis of amikacin and kanamycin, meeting the USP
  performance requirements
- · 3465 ECD demonstrates an excellent linear response in concentration ranges sufficient for the assay of

amikacin and kanamycin

# Introduction

Amikacin and kanamycin are closely related aminoglycoside antibiotics and are often used for treatment of bacterial infections and tuberculosis. KAN is isolated from the bacterium *Streptomyces kanamyceticus*. AMI is derived from KAN by the acylation of an amino group of KAN. Figure 1 shows the structures of AMI and KAN. The USP has issued AMI and KAN monographs in which High Performance Anion Exchange Chromatography (HPAEC) with Pulsed Amperometric Detection (PAD) was used for the assay.<sup>1–5</sup> This application note describes the HPAEC-PAD assay of AMI and KAN on an ACQUITY Arc System with a 3465 ECD, demonstrating its performance and suitability for the analysis of AMI and KAN in pharmaceutical products.



Figure 1. Structures of amikacin and kanamycin.

# Experimental

## Standards and Reagents

Kanamycin sulfate and amikacin (CRM) were purchased from Sigma-Aldrich, Inc. (St. Louis, MO). Amikacin Injections (250 mg/mL) were obtained from a local provider. Low-carbonate sodium hydroxide (NaOH) solution (50% w/w Certified, Fisher Chemical<sup>™</sup>, carbonate ≤0.10%) was purchased from Fisher Scientific Co. (Pittsburgh, PA). Methanol (for HPLC) was purchased from Honeywell Research Chemicals (Charlotte, NC). Deionized water (>18.2 Mohm·cm) was generated in-house.

## Preparation

Stock standard solutions of KAN and AMI were prepared by dissolving standards in deionized water at 10 mg/mL. A standard mix solution of KAN and AMI at 0.008 mg/mL and 0.020 mg/mL, respectively, were prepared from the stock solutions.

AMI Injection sample solutions were prepared by dilution with deionized water to 0.020 mg/mL AMI (based on the label concentration).

Mobile phases of 115 mM and 134 mM NaOH were prepared by mixing of certain amount of low-carbonate 50% w/w NaOH solution with deionized water to make a 1 liter NaOH solution. The deionized water was degassed prior to be used in the mobile phase preparation. Fresh mobile phase was prepared daily.

# **HPAEC-PAD** Conditions

System:	ACQUITY Arc System (QSM-R and SM FTN-R) with a 3465 ECD
Software:	Empower 3 (3.6.1) Chromatography Data Software
Column:	Hamilton® RCX-10 Analytical Column 250 x 4.6 mm I.D. 7 µm (PEEK) with a RCX-10 guard column
Vial:	PP vial 700 µL volume with cap and preslit septum (p/n:

#### 186005221)

Temp.:	30 °C for column and detection		
	flow cell		
Flow cell:	FlexCell <sup>®</sup> with a gold working		
	electrode, a HyREF reference		
	electrode and a stainless-steel		
	auxiliary electrode, 50 μm		
	spacer (p/n: 700013068).		
Mobile phase:	115 or 134 mM NaOH solution		
	(protected under $N_2$ gas using		
	a solvent stabilization kit (p/n:		
	205001099)		
SM purge/wash solvent:	Deionized water		
Seal wash solvent:	Water:methanol 8:2 (v/v)		
Flow rate:	0.5 ml /min		
now rate.	0.0 mL/mm		
Run time:	30 min		
Injection volume:	10 µL		
Injector loop offline:	0.5 min		
Detection:	PAD		
Potential waveform	See Table 1		
ADF:	0.02 Hz		

Step	Time (s)	Potential (V)	Integration
1	0.00	+0.04	—
2	0.30	+0.04	Begins
3	0.50	+0.04	End
4	0.51	+0.80	—
5	0.70	+0.80	—
6	0.71	-0.80	_
7	0.90	-0.80	_

Table 1. The potential waveform in pulsed amperometric detection.

## Working with High pH Mobile Phases

Additional precautions were taken in working with these mobile phases (pH up to 13.1). The precaution steps include a) PEEK tubing (0.004 inch i.d.) was used to connect the injector valve to the columns and further to the detector flow cell; b) the injector loop was taken offline after injection (at 0.5 minute); and c) deionized water was used to flush the system and columns after daily use. There was no leakage or carry-over observed in the study (about 500 injections of standards and samples).

**Results and Discussion** 

# Detector Flow Cell

The detector flow cell consisting of a gold working electrode, a hydrogen reference electrode (HyREF), and an auxiliary stainless-steel electrode was used in the analysis. The USP Monographs recommend a flow cell with a gold working electrode and a silver-silver chloride reference electrode. The silver-silver chloride reference electrode requires regular maintenance while the HyREF does not. The auxiliary electrode helps to improve the dynamic range of detection. Based on these considerations, the FlexCell with a gold working electrode, a HyREF reference electrode, and a stainless-steel auxiliary electrode was used.

### System Suitability Test

USP Monographs for AMI and KAN specify a set of performance criteria for the assay.<sup>1–5</sup> The assay conditions for both analytes are the same except for the mobile phase. For KAN a mobile phase of 115 mM NaOH solution is recommended and for AMI a 134 mM NaOH solution is recommended. Table 2 shows the USP criteria and the measured results using a standard mix solution of 0.008 mg/mL KAN sulfate and 0.020 mg/mL AMI under the specified mobile phase conditions. Figure 2 shows an overlay of the chromatograms obtained from the standard mix solution in USP specified mobile phases. These results meet all USP system suitability requirements (Table 2).



Figure 2. Overlay of chromatograms from 0.008 mg/mL kanamycin sulfate and 0.020 mg/mL amikacin obtained in USP specified mobile phases (n=4).

Parameter	USP requirement	Kanamycin <sup>1</sup>	Amikacin <sup>2</sup>	
Resolution (KAN/AMI)	≥3	4.9	4.5	
Tailing factor	≤2	1.28	1.89	
RSD peak area (KAN)	≤2%	1.0% <sup>3</sup>		
RSD peak area (AMI)	≤3%		1.6% <sup>3</sup>	

Note: 1) results were obtained in a 115 mM NaOH mobile phase; 2) results were obtained in a 134 mM NaOH mobile phase; 3) n=4.

Table 2. USP system suitability criteria and test results.

## Linearity

USP Monographs for AMI and KAN specify single point (level) calibrations, which require a linear (linear through zero) relationship between the detector's response and the concentration. The detector's response linearity was investigated. Table 3 shows calibration results for AMI and KAN in concentration ranges up to 140% of the target concentrations in the USP specified mobile phases. Data points at six concentration levels were fitted by a least square regression to lines (through zero). The correlation coefficient values for both analytes were 0.997, demonstrating an excellent linear relationship between the peak area and the analyte concentration. An injection volume of 10 µL was used in this study.

Analyte	Target conc. (mg/mL)	Conc. range (mg/mL)	Mobile phase	Equation	Correlation coefficient (R)
AMI	0.020	0-0.0275	134 mM NaOH	$Y=1.05 \times 10^{12} \text{ X}$	0.997
KAN sulfate	0.008	0-0.0111	115 mM NaOH	$Y=1.28 \times 10^{12} \text{ X}$	0.997

Table 3. Calibration results for kanamycin sulfate and amikacin.

## Sample Analysis

Table 4 shows AMI assay results for two AMI Injection samples. These determined concentration results are within the USP acceptance criteria for AMI (90.0%–120.0%). Figure 3 shows the chromatograms of triplicate injections of these samples.



Figure 3. Overlay of chromatograms from analysis of amikacin injection solutions (n=3).

Sample	Labeled conc. (mg/mL)	Determined conc. <sup>1</sup> (mg/mL)	Relative to labeled value	USP acceptance criteria	
AMI Injection A	250	273 (1.8%)	109.2%	00.00/ 100.00/	
AMI Injection B	250	278 (2.8%)	111.2%	90.0%-120.0%	

1): The mean and RSD values from 3 measurements are shown.

Table 4. Analysis results for amikacin injection solutions.

# Conclusion

This work demonstrates that the ACQUITY Arc System with a 3465 Electrochemical Detector under the Empower 3 CDS control provides a reliable and accurate assay of AMI and KAN in pharmaceutical products. The

analytical performance meets the USP performance requirements for the peak resolution, tailing and reproducibility for AMI and KAN.

# References

- United States Pharmacopeia (2022), USP Monographs. Kanamycin Injection, USP43-NF38 2501. Rockville, MD: United States Pharmacopeia.
- United States Pharmacopeia (2022), USP Monographs. Kanamycin Sulfate, USP43-NF38 2502. Rockville, MD: United States Pharmacopeia.
- 3. United States Pharmacopeia (2022), USP Monographs. Amikacin, USP43-NF38 219. Rockville, MD: United States Pharmacopeia.
- United States Pharmacopeia (2022), USP Monographs. Amikacin Sulfate, USP43-NF38 220. Rockville, MD: United States Pharmacopeia.
- United States Pharmacopeia (2022), USP Monographs. Amikacin Sulfate Injection, USP43-NF38 221.
   Rockville, MD: United States Pharmacopeia.

# Featured Products

ACQUITY Arc System <https://www.waters.com/134844390>

Empower Chromatography Data System <a href="https://www.waters.com/10190669">https://www.waters.com/10190669</a>>

3465 Electrochemical Detector for LC Systems <a href="https://www.waters.com/waters/nav.htm?cid=514431">https://www.waters.com/waters/nav.htm?cid=514431</a>

720007647, May 2022

© 2022 Waters Corporation. All Rights Reserved.

Terms of Use	Privacy	Trademarks	Sitemap	Careers	Cookie	Cookie
偏好設定						