

## Analysis of Azithromycin on the Alliance™ iS HPLC System: System Performance under Challenging Method Conditions

---

Anthony Dovell, Paula Hong

Waters Corporation

Ce document est une note d'application et ne contient pas de section détaillée concernant l'expérimentation.

---

### Abstract

Many existing reversed-phase monographs contain a combination of conditions that can impact long term system performance. For example, traditional HPLC gradient methods may use both buffered and organic mobile phases at proportions that the salt may be insoluble, potentially leading to precipitation of the buffer during the course of the analysis. While these methods have been validated and can be performed on a number of different HPLC systems, these conditions can lead to system-to-system variability while still meeting the relevant suitability requirements listed by the USP. The USP monograph for Azithromycin organic impurities is one such example. This application note will discuss the challenges of this method and compare the results obtained across a number of different HPLC systems.

### Benefits

- Alliance iS HPLC System meets system suitability criteria for USP monograph of Azithromycin
-

- Superior retention time precision for high salt mobile phases under HPLC conditions
  - Stable performance over 30 hours under high salt mobile phase conditions
- 

## Introduction

Many regulated legacy monographs require buffers in the mobile phase to ensure reliable performance of the method. For these traditional gradient reversed-phase applications the mobile phase transitions from a buffered, aqueous mobile phase to an organic mobile phase. The proportions of each mobile phase during elution may cause the salt to become insoluble in the water/organic composition, resulting in potential precipitation of the salt. Furthermore, under these conditions there may be variability across different HPLC systems due to design differences. For this application note, the USP monograph for Azithromycin Organic Impurities was chosen as a representative method for these conditions. This specific monograph mobile phases include a weak aqueous mobile phase buffer at a concentration of 12 mM, and a strong mobile phase with acetonitrile/methanol blend. The separation requires a gradient which has a strong elution of 25/75 coupled with a 75% organic mobile phase composition. This monograph was executed on multiple HPLC systems, using replicate injections of standards to investigate system suitability and performance differences across systems.

---

## Experimental

### Sample Description

Azithromycin (p/n: 1A00860), Azithromycin Related Compound F (p/n: 1046089), and Desosaminylazithromycin (p/n: 1046078) were obtained from the U.S. Pharmacopeia (Rockville, MD). An expired Azithromycin standard from the U.S. Pharmacopeia was used to act as a surrogate sample. All solutions were prepared in a diluent of 7:6:7 methanol: acetonitrile: and 1.73 mg/mL monobasic ammonium phosphate (pH 10.0 ± 0.05). Azithromycin standard solution was prepared at 86 µg/mL and the system suitability solution was prepared at 16.5 µg/mL and 27.0 µg/mL of Azithromycin Related Compound F and Desosaminylazithromycin, respectively. The Azithromycin sample solution was prepared at 8.6 mg/mL. All samples were prepared as described in the monograph.

### LC Conditions

---

LC system:	Alliance iS HPLC System
	Vendor X HPLC System
	Vendor Y HPLC System
	Vendor Z HPLC System
Detection:	UV for all systems (210 nm)
Column:	XBridge™ C <sub>18</sub> , 5µm, 4.6 x 250 mm (p/n: 186003117)
Column temperature:	60 °C
Sample temperature:	5 °C
Injection volume:	50 µL
Flow rate:	1.00 mL/min
Mobile phase A:	1.8 mg/mL anhydrous dibasic sodium phosphate (pH 8.9)
Mobile phase B:	3:1 Acetonitrile/Methanol
Seal wash:	90:10 Water/Acetonitrile (where applicable)
Needle wash:	50:50 Water/Mobile Phase B

## Gradient Table

Time (min)	Flow (mL/min)	%A	%B	Curve
Initial	1.00	50	50	6
25	1.00	45	55	6
30	1.00	40	60	6
80	1.00	25	75	6
81	1.00	50	50	6
93	1.00	50	50	6

## Data Management

Chromatography software:

Empower™ 3.7.0

---

## Results and Discussion

### Analysis on Alliance iS HPLC System

For this study, a prescribed analysis and process was used for consistent data quality, including the appropriate number of injections to meet systems suitability. The total run time for the 20 injections was over 31 hours. Each system was thoroughly equilibrated and showed stable system pressure before sample set submission. Data was analyzed and evaluated based on the system suitability requirements including: system suitability standard-azithromycin related compound F/desosaminylzithromycin peak to valley ratio not less than (NLT 1.4) and standard -azithromycin USP Tailing 0.8–1.5. To control system-to-system variability the same HPLC column was used on all systems tested.

The Alliance iS was able to execute the sample set described above, running for more than 31 hours continuously. Evaluation of the system suitability results (Table 1) showed the USP tailing of azithromycin standard met the criteria. For the system suitability sample, the peak to valley ratio for azithromycin related

compound F and desosaminylazithromycin could not be calculated, as the two peaks were baseline resolved with USP resolution of approximately 3.0 (Figure 1). Since peak to valley calculations are only determined for those peaks that are not baseline resolved or co-elute, the system suitability value could not be determined. When looking at non-system suitability requirements, including retention time precision, the Alliance iS HPLC System showed %RSD of less than 0.2% for all analytes for the six replicate injections over an analysis time of nine hours.

Sample	Standard		System suitability			
System	Azithro USP tailing	Peak to Valley Azithro RC F/Desoazi Ratio	USP resolution Azithro RC F/Desosazi	Azithro RC F Retention time RSD	Desosamin Retention time RSD	Azithro retention time RSD
System suitability criteria	0.8–1.5	NLT 1.4	N/A	N/A	N/A	N/A
Alliance iS HPLC System	1.1	>1.4 (*Baseline resolved)	3.04	0.12%	0.10%	0.01%

Table 1. Result for analysis on the Alliance iS HPLC System. All values are average of n=6.





Figure 1. Representative chromatograms of the azithromycin standard solution and system suitability solution using the Alliance iS HPLC.

When analyzing the surrogate azithromycin sample, impurity peaks were detected. Per the USP monograph, the following impurity peaks are discarded: all peaks eluting before azithromycin N-oxide and after 3-deoxyazithromycin (azithromycin B) and peaks with a response less than 0.1 times the response of the azithromycin peak in the Standard solution (0.1%). Based on these criteria, one particular peak of interest met the monograph criteria for integration. After calculating the relative retention time for this peak (0.53 compared to 0.54 monograph value) and overlaying the sample chromatogram with the system suitability standard, the impurity was shown to likely be desosaminylazithromycin.



*Figure 2. Overlay of the surrogate Azithromycin sample (blue) and the System Suitability Standard (black) using the Alliance iS HPLC.*


## Comparison of Performance Across Different HPLC Systems

The monograph was also tested across different HPLC systems to assess the impact of system design on method performance. For the other systems tested, the analysis, as previously described, ran to completion. In general, all systems were able to meet the system suitability criteria. Using the same column on all systems, each system produced comparable USP tailing values for azithromycin in the standard solution. In addition, all systems met the requirement for peak-to-valley of NLT 1.4 between azithromycin related compound F and desosaminylazithromycin.

Given specific characteristics of the method, it was also used to assess retention time precision under challenging conditions. Specifically, during the segment in which the API and some related impurities elute (0–25 min), the mobile phase composition changes by 0.5% per minute, which can result in retention time variability. These retention time shifts become more pronounced the later a compound elutes in the chromatogram. In addition, mobile phase compositions and gradient have the potential to cause buffer precipitation. Specifically mobile phase A consists of a 1.8 mg/mL phosphate buffer while mobile phase B contains a 3:1 mixture of Acetonitrile/Methanol with a gradient composition of up to 75% by the end of the gradient. This high organic

solvent composition coupled with phosphate buffer creates conditions that may cause precipitation of the buffer salt. The precipitation of salt could impact the system in multiple ways: over pressurizing due to clogging of the system or column, or loss of system pressure/prime due to particulates in the pump. The latter may be temporary but can lead to inconsistent retention times from injection-to-injection. Because impurities are identified solely based on retention times, it is critical to have good retention time precise as variations in retention time could lead to the misidentification of impurities in the sample since many of the known impurities closely elute.

When comparing the retention time precision, for the analysis retention time standard deviation of critical compounds, the Alliance iS showed lower %RSD values for all three compounds when compared to the three other HPLC systems. Of the other systems tested, System Z HPLC system showed significantly more retention time variability for all compounds (Figure 3).




*Figure 3. Comparison of retention time %RSD values generated for key compounds Azithromycin Related Compound F, Desosaminylazithromycin, and Azithromycin.*

Upon closer inspection of the data, some abnormalities in system pressure were observed on a few HPLC systems. While each system was able to complete the sample set and meet the system suitability requirements



there was greater retention variability across some systems. The variability varied from system to system. For system Z HPLC, system pressure oscillations were observed in the first injection (black), with increasing variability at higher composition of organic solvent. After these pressure oscillations were observed in the first injection, they continued throughout the sample set. While this system was able to complete the sample set and met both system suitability requirements, the higher retention time %RSD values were observed (>2%). Below we see the overlaid system pressure channels of the first four injections on System Z HPLC System. Data is offset for comparative purposes.



*Figure 4. Top: System pressure overlay of the first four injections from System Z HPLC, offset for comparative purposes. Injections in order are black (blank), green, blue, red. Bottom: Corresponding UV signals for the system suitability injections.*

Another HPLC system tested produced some variability with all %RSD's between 0.2–0.6%. For the first four injections of the sample set from System X, there were minor drop offs in pressure (Figure 5). During the first injection (black) the system pressure fluctuations were typical for gradient analysis up until approximately 30 minutes, at which point some minor pressure dips were observed. Subsequent injections (green, blue, red) produced larger pressure dips. While this system was able to complete the sample set and met both system suitability requirements, the retention time variability may be a result of these momentary drops in pressure.




Figure 5. Top: System pressure overlay of the first four injections from System X HPLC, offset for comparative purposes. Injections in order are black (blank), green, blue, red. Bottom: Corresponding UV signals for the system suitability injections.

In comparison, the Alliance iS showed stable system pressure and retention time precision of less than 0.2% throughout the sample set. Specifically, the Alliance iS produced similar system pressure profiles throughout the sample set with no unusual dips during the gradient (Figure 5). The system also produced good retention time precision over the sample set with low %RSD's <0.2%.




Figure 6. Top: System pressure overlay of the first four injections from the Alliance iS HPLC system, offset for comparative purposes. Injections in order are black (blank), green, blue, red. Bottom: Corresponding UV signals for the system suitability injections.

## Impact of Retention Time Variability of Relative Retention Times


For impurity methods, retention time reproducibility is critical for confirmation or identification of impurities. Given there can be retention time shifts across systems due to gradient delay volumes, many monographs provide RRT for impurities as guidance to assist in identification or confirmation of impurities. While there may be shifts from system-to-system, it is typically expected that RRT would be reproducible from injection-to-injection.

Evaluation of the RRT for the impurity peak showed the impact of retention time variability. The table below summarizes the RRT values for an impurity peak in the azithromycin sample on system Z HPLC and the Alliance iS HPLC over six replicate injections (Table 2). For system Z, the values range from 0.47 to 0.49. While these differences may seem small, there are two impurities specified in the monograph with similar RRT's, and the unknown peak on System Z shows RRTs in between both. Thus, depending on the injection, the peak may be assigned to a different impurity.

Comparatively, the Alliance iS HPLC showed a consistent RRT value across all injections. More interestingly, this impurity would be incorrectly identified using system Z after closer inspection. Because the system suitability standard contains the compound desosaminylazithromycin (DESO), the overlay of the system suitability standard and sample can provide confirmation of this peak. For the Alliance iS the peak at 23.1 min is comparable in retention times and RRT to desosaminylazithromycin, with a specified RRT of 0.54 in the monograph (Table 2). Thus, providing peak confirmation of desosaminylazithromycin on the Alliance iS HPLC. In contrast, when analyzing the data from System Z there is no clear peak in either retention time or RRT indicative of desosaminylazithromycin. Due to the retention time variability, confirmation based on retention time or the RRT of desosaminylazithromycin is challenging and may lead the analyst to the wrong conclusion.

	Injection	Impurity RT (min)	Azithromycin RT (min)	Impurity RRT	Closest preceding impurity RRT	Closest succeeding impurity RRT
System Z HPLC (sample)	1	21.41	45.23	0.47	0.43	0.51
	2	21.82	45.65	0.48		
	3	22.84	46.55	0.49		
	4	21.31	45.42	0.47		
	5	21.94	45.84	0.48		
	6	22.25	45.88	0.48		
System suitability standard (DESO)	Average of n=6 inj.	23.94 (2.65% RSD)			0.52 (Average calculated RRT of Deso in standard)	
Alliance iS HPLC (sample)	1	23.14	43.91	0.53	0.51	0.54 (DESO RRT per monograph)
	2	23.13	43.91	0.53		
	3	23.14	43.91	0.53		
	4	23.15	43.93	0.53		
	5	23.16	43.94	0.53		
	6	23.15	43.92	0.53		
System suitability standard (DESO)	Average of n=6 inj.	23.11 (0.097% RSD)			0.53 (Average calculated RRT of DESO in standard)	

*Table 2. Tabulated relative retention times of an impurity in the surrogate azithromycin sample. Based on RT and RRT comparisons on the Alliance iS HPLC, the peak identity is likely Desosaminylazithromycin (DESO).*



*Figure 7. Overlay of the surrogate Azithromycin replicate sample injections using System Z HPLC system. Impurities shift into the retention time windows of other impurities, leading to difficulties processing.*

## Conclusion

The USP provides monographs for the purpose of providing pharmaceutical quality assurance. In the case of HPLC monographs, system parameters and system suitability requirements are detailed in order to generate complaint and reliable data. While many HPLC systems from various vendors can meet the minimum system requirements needed to execute an assay, it is not uncommon to see differences in data sets generated using varying instruments. While all systems tested showed acceptable system suitability results, system performance varied from system to system. In this study, the Alliance iS HPLC System was shown to be the most robust and reproducible system in comparison to the others evaluated. On the Alliance iS HPLC System, the system suitability requirements were easily met and showed the lowest retention time variability from injection to injection, resulting in impurity identification confidence.

---

## References

1. United States Pharmacopeia (2022) USP Monographs Azithromycin Organic Impurities, USP-NF. *Rockville, MD*:USP DOI:[https://doi.org/10.31003/USPNF\\_M6740\\_05\\_01](https://doi.org/10.31003/USPNF_M6740_05_01)

---

## Featured Products

Alliance iS HPLC System <

<https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/alliance-is-hplc-system.html>>

ACQUITY UPLC Tunable UV Detector <<https://www.waters.com/514228>>

Empower Chromatography Data System <<https://www.waters.com/10190669>>

720008116, November 2023



© 2024 Waters Corporation. All Rights Reserved.

[Conditions d'utilisation](#) [Politique de confidentialité](#) [Marques](#) [Carrières](#) [Mentions légales](#) et [déclaration de confidentialité](#) [Cookies](#) [Préférences de cookies](#)