

Analysis of Tricyclic Antidepressant Drugs in Plasma for Clinical Research

Vallerie Muckoya, Stephen Balloch, Gareth Hammond, Lisa J. Calton

Waters Corporation

For research use only. Not for use in diagnostic procedures.

Abstract

Chromatography and mass detection provide analytical selectivity, allowing for the quantification of 15 TCAs across a broad spectrum of polarities. Additionally, the method involves rapid and cost-effective sample preparation, utilizing only a small sample volume.

Benefits

- Analytical selectivity afforded by chromatography and mass detection
 - Quantification of 15 TCAs that span a wide range of polarities
 - Quick and inexpensive sample preparation, utilizing a small volume of sample
-

Introduction

The quantitative analysis of tricyclic antidepressants in plasma, is vital in order to undertake pharmacokinetic

studies and monitor therapy efficiently.¹

Here we describe a clinical research method utilizing protein precipitation for the extraction of 15 tricyclic antidepressants from plasma. Chromatographic separation was performed using an ACQUITY™ UPLC I-Class FL System, using a XSelect™ Premier HSS C₁₈ Column (2.1 x 100 mm, 2.5 µm). The detector was a Xevo™ TQ-S micro Mass Spectrometer operating in positive electrospray ionization mode (Figure 1).



Figure 1. The ACQUITY UPLC I-Class FL System and Xevo TQ S-micro Mass Spectrometer.

Experimental

Sample Preparation

Plasma calibrators and quality control materials were prepared in house using pooled human plasma supplied by

BioIVT (West Sussex, UK). Concentrated stock solutions were prepared from certified powders and solutions supplied by Cambridge Bioscience (Cambridgeshire, UK), Sigma-Aldrich (Dorset, UK) and Toronto Research Chemicals (Ontario, Canada). Stable labelled internal standards were supplied by ALSACHIM (Illkirch-Graffenstaden, France), Sigma-Aldrich (Dorset, UK) and Toronto Research Chemicals (Ontario, Canada). The calibration ranges and in-house quality control concentrations for the analytes are as shown in Table 1.

Analytes	Calibration range (ng/mL)	QC concentrations (ng/mL)
Amitriptyline, clomipramine, clozapine, desipramine, doxepine, imipramine, maprotiline, norclomipramine, norclozapine, nordoxepin, normaprotiline, nortrimipramine, nortriptyline, protriptyline, trimipramine	10–500	25, 75 and 400
Normaprotiline and trimipramine	20–1000	50, 150 and 800
Clozapine and norclozapine	50–2500	125, 375 and 2000

Table 1. Calibration and in-house QC concentration range.

Sample Extraction

To a 50 μ L of sample in a microcentrifuge tube, 150 μ L of internal standard in acetonitrile was added. The concentrations of internal standards are detailed in Table 2.

Internal standard	Concentration (ng/mL)
Nortriptyline- ² H ₃	70
Protriptyline- ² H ₃	87
Normaprotiline- ² H ₄	237
Nordoxepin- ² H ₃	115
Desipramine- ² H ₃	80
Norclozapine- ² H ₈	3059
Amitriptyline- ² H ₃	100
Maprotiline- ² H ₅	77
Doxepin- ² H ₃	107
Nortrimipramine- ¹³ C ₂ ² H ₃	83
Imipramine- ² H ₆	107
Trimipramine- ² H ₃	182
Norclomipramine- ² H ₃	175
Clomipramine- ² H ₃	88
Clozapine- ² H ₄	686

Table 2. Internal standard concentrations.

Tubes were placed on a multi-tube vortex mixer at 1500 rpm for three minutes, then centrifuged for two minutes at 16, 100 g. 25 µL of supernatant was transferred to a 1 mL 96 well plate and 475 µL water added. The plate was capped and vortexed for two minutes at 1500 rpm prior to analysis.

UPLC Conditions

System:	ACQUITY UPLC I-Class FL System
Needle:	30 µL

Column:	XSelect Premier HSS T3 Column; 2.5 µm, 2.1 x 100 mm (p/n: 186009831)
Mobile phase A:	Water + 5 mM ammonium formate + 0.1% formic acid
Mobile phase B:	Methanol + 5 mM ammonium formate + 0.1% formic acid
Needle wash solvent:	80% aqueous methanol + 0.1% formic acid
Purge solvent:	Water: Methanol 40:60 v:v
Seal wash:	20% aqueous methanol
Column temperature:	45 °C (precolumn heater active)
Injection volume:	20 µL

Gradient Table

Time (min)	Flow rate (mL/min)	%A	%B	Curve
0	0.50	40	60	Initial
1.50	0.50	40	60	6
3.00	0.50	0	100	6
3.14	0.50	0	100	11
3.15	0.50	40	60	11

Table 3. Chromatographic elution timetable.

Run time: 4.0 minutes (4.5 minutes injection-to-injection)

MS Conditions

System:	Xevo TQ-S micro
Resolution:	MS1 (0.7 FWHM) MS2 (0.7 FWHM)
Acquisition mode:	Multiple Reaction Monitoring (MRM) (see Table 3 for details)
Polarity:	ESI + ionization
Capillary:	0.5 kV
Source temperature:	150 °C
Desolvation temperature:	600 °C
Cone gas:	150 L/hr
Desolvation gas flow:	1000 L/hr
Inter-scan delay:	automatic
Polarity/Mode switch inter-scan delay:	0.020 seconds
Inter-channel delay:	automatic

Data Management

Chromatography software:	MassLynx™ v4.2 with TargetLynx™ Application Manager.
--------------------------	--

Analyte	Polarity	Precursor ion (m/z)	Product ion (m/z)	Cone voltage (V)	Collision (V)	Dwell (s)
Nortriptyline (Quan)	ESI+	264.1	105.1	30	20	0.012
Nortriptyline (Qual)	ESI+	264.1	233.1	30	14	0.012
Nortriptyline- ² H ₃ (ISTD)	ESI+	267.2	105.1	30	20	0.012
Protriptyline (Quan)	ESI+	264.1	155.1	36	22	0.012
Protriptyline (Qual)	ESI+	264.1	233.2	36	18	0.012
Protriptyline- ² H ₃ (ISTD)	ESI+	267.1	155.1	36	22	0.012
Normaprotiline (Quan)	ESI+	264.1	169.1	32	18	0.012
Normaprotiline (Qual)	ESI+	264.1	131.1	32	22	0.012
Normaprotiline- ² H ₄ (ISTD)	ESI+	268.1	173.1	32	18	0.012
Nordoxepin (Quan)	ESI+	266.1	107.1	32	22	0.012
Nordoxepin (Qual)	ESI+	266.1	44.1	32	20	0.012
Nordoxepin- ² H ₃ (ISTD)	ESI+	269.1	107.1	32	22	0.012
Desipramine (Quan)	ESI+	267.1	72.1	26	18	0.012
Desipramine (Qual)	ESI+	267.1	208.1	26	24	0.012
Desipramine- ² H ₃ (ISTD)	ESI+	270.1	75.1	26	18	0.012
Norclozapine (Quan)	ESI+	313.1	192.1	46	38	0.012
Norclozapine (Qual)	ESI+	313.1	270.1	46	24	0.012
Norclozapine- ² H ₄ (ISTD)	ESI+	321.2	193.1	46	38	0.012
Amitriptyline (Quan)	ESI+	278.1	105.1	32	22	0.012
Amitriptyline (Qual)	ESI+	278.1	84.1	32	22	0.012
Amitriptyline- ² H ₃ (ISTD)	ESI+	281.1	105.1	32	22	0.012
Maprotiline (Quan)	ESI+	278.1	250.1	36	18	0.012
Maprotiline (Qual)	ESI+	278.1	219.1	36	24	0.012
Maprotiline- ² H ₅ (ISTD)	ESI+	283.2	255.1	36	18	0.012
Doxepin (Quan)	ESI+	280.1	107.1	38	26	0.012
Doxepin (Qual)	ESI+	280.1	84.1	38	26	0.012
Doxepin- ² H ₃ (ISTD)	ESI+	283.1	107.1	38	26	0.012
Nortrimipramine (Quan)	ESI+	281.1	44.1	28	28	0.012
Nortrimipramine (Qual)	ESI+	281.1	86.1	28	16	0.012
Nortrimipramine ¹³ C ² H ₃ (ISTD)	ESI+	285.3	48.1	28	28	0.012
Imipramine (Quan)	ESI+	281.2	58.1	24	28	0.012
Imipramine (Qual)	ESI+	281.2	86.1	24	16	0.012
Imipramine- ² H ₆ (ISTD)	ESI+	287.2	64.1	24	28	0.012
Trimipramine (Quan)	ESI+	295.1	100.1	30	18	0.012
Trimipramine (Qual)	ESI+	295.1	58.1	30	26	0.012
Trimipramine- ² H ₃ (ISTD)	ESI+	298.1	103.1	30	18	0.012
Norclomipramine (Quan)	ESI+	301.1	72.1	28	16	0.012
Norclomipramine (Qual)	ESI+	301.1	44.1	28	32	0.012
Norclomipramine- ² H ₃ (ISTD)	ESI+	304	75.1	28	16	0.012
Clomipramine (Quan)	ESI+	315.1	86.1	30	18	0.012
Clomipramine (Qual)	ESI+	315.1	58.1	30	34	0.012
Clomipramine- ² H ₃ (ISTD)	ESI+	318.2	89.1	30	18	0.012
Clozapine (Quan)	ESI+	327.1	270.1	38	20	0.012
Clozapine (Qual)	ESI+	327.1	192.1	38	44	0.012
Clozapine- ² H ₄ (ISTD)	ESI+	331.1	273.1	38	20	0.012

Table 4. Guideline MRM parameters for analytes and internal standards used in this study.

Results and Discussion

Chromatographic separation was achieved for isobaric compounds (nortriptyline and protriptyline) and interfering qualifier transitions (281.2>86.1) of imipramine and nortrimipramine. This ensured selective quantification of these analytes. An example chromatogram is shown in Figure 2 of a pooled plasma mid-level calibrator sample.




Figure 2. Chromatogram of plasma sample showing the analysis of tricyclic antidepressant drugs on XSelect Premier HSS T3 Column.

There was no significant carryover observed from high concentration plasma sample into subsequent blank injection. The high concentration sample contained 500 ng/mL of amitriptyline, clomipramine, desipramine, doxepin, imipramine, maprotiline, norclomipramine, nordoxepin, nortrimipramine, nortriptyline, and protriptyline, 1000 ng/mL of normaprotiline, and trimipramine, and 2500 ng/mL of clozapine, and norclozapine. A dilution of 1:5 of the high concentration sample was performed that gave a mean % bias of < 10% for all analytes.

No significant interference was observed when the TCAs were analysed with endogenous compounds (albumin, bilirubin, cholesterol, creatinine, triglycerides, and uric acid) spiked at low and high concentrations. This was assessed by determining the recovery (n=3) from low and high pooled plasma samples at low (QC₁) and high (QC₃) concentrations. The recoveries for the low and high pools were within 15%, except for amitriptyline,

clomipramine, clozapine, imipramine, norclomipramine, nortriptyline, protriptyline, and trimipramine which exhibited some interference (17%) from triglycerides at low concentration samples.

Analytical sensitivity studies were performed by extracting and quantifying ten replicates of four levels of low concentration samples prepared in plasma, over five days (n=40). Results obtained demonstrated that the method would allow for precise quantification ($\leq 20\%$ CV, $\leq 15\%$ bias) at the concentrations shown below in Table 5.

Analyte	LLMI (ng/mL)	Precision (% CV)	Bias (%)
Nortriptyline	5	9.6	-0.1
Protriptyline	5	10	4
Normaprotiline	10	10	2.5
Nordoxepin	5	5.8	4.1
Desipramine	5	5	2.4
Norclozapine	25	5.1	3.1
Amitriptyline	5	7	6.3
Maprotiline	5	8	3.2
Doxepin	5	7.4	6.9
Nortrimipramine	5	6.9	1.2
Imipramine	5	6.8	1.7
Trimipramine	10	4.7	2.2
Norclomipramine	5	6.7	6.4
Clomipramine	5	6.1	0.6
Clozapine	25	7.2	1.8

Table 5. Analytical sensitivity summary (LLMI is lower limit of the measuring interval).

Total precision was determined by extracting and quantifying five replicates of three concentration levels of plasma pools over five separate days (n=25). Repeatability was assessed by analysing five replicates at each QC level. The results presented in Figure 3 demonstrated the total precision and repeatability assessed at the three concentration levels (25, 75, and 400 ng/mL) for all analytes, with the exception of normaprotiline, and trimipramine at 50, 150, and 800 ng/mL, and clozapine, norclozapine assessed at 125, 375, and 2000 ng/mL, was $\leq 8.0\%$ CV.




Figure 3. Total precision and repeatability summary.

Matrix effects investigations were performed using donor human plasma samples ($n=6$) at low (QC_1) and high (QC_3) levels and evaluated as a percentage of extracted solvent samples spiked to equivalent concentrations. As shown in Table 6, the matrix factor calculations using analyte: internal standard response ratio, demonstrated that the internal standard compensated for any signal enhancement or suppression observed.

Analyte	Spiked concentration (ng/mL)	Matrix factor based on peak area mean (range)	Matrix factor based on response mean (range)
Nortriptyline	25	0.93 (0.90–0.95)	0.99 (0.96–1.05)
	400	0.99 (0.97–1.00)	0.99 (0.97–1.03)
Protriptyline	25	0.93 (0.91–0.95)	1.00 (0.96–1.07)
	400	0.97 (0.96–0.97)	0.98 (0.95–1.01)
Normaprotiline	50	0.90 (0.89–0.92)	1.01 (0.95–1.07)
	800	0.97 (0.95–0.98)	0.97 (0.95–0.99)
Nordoxepin	25	0.92 (0.90–0.94)	0.99 (0.96–1.04)
	400	0.95 (0.94–0.96)	0.97 (0.94–0.99)
Desipramine	25	0.93 (0.90–0.95)	1.00 (0.97–1.06)
	400	0.98 (0.97–0.98)	0.99 (0.95–1.01)
Norclozapine	125	0.91 (0.89–0.92)	0.99 (0.94–1.03)
	2000	0.95 (0.94–0.96)	0.97 (0.92–1.01)
Amitriptyline	25	0.92 (0.88–0.95)	0.99 (0.96–1.04)
	400	0.96 (0.93–0.97)	0.97 (0.94–0.99)
Maprotiline	25	0.92 (0.89–0.94)	0.98 (0.94–1.03)
	400	0.97 (0.96–0.97)	0.97 (0.94–1.01)
Doxepin	25	0.91 (0.88–0.93)	0.99 (0.95–1.03)
	400	0.94 (0.92–0.95)	0.97 (0.94–1.00)
Nortrimipramine	25	0.92 (0.90–0.95)	0.98 (0.95–1.03)
	400	0.96 (0.94–0.97)	0.98 (0.95–1.00)
Imipramine	25	0.93 (0.90–0.95)	1.00 (0.97–1.04)
	400	0.96 (0.94–0.97)	0.99 (0.95–1.01)
Trimipramine	50	0.92 (0.90–0.95)	1.00 (0.96–1.04)
	800	0.98 (0.97–0.99)	0.99 (0.96–1.02)
Norclomipramine	25	0.93 (0.92–0.95)	1.00 (0.97–1.05)
	400	0.96 (0.94–0.98)	0.99 (0.96–1.00)
Clomipramine	25	0.94 (0.91–0.97)	0.99 (0.96–1.06)
	400	0.96 (0.94–0.98)	0.98 (0.94–1.00)
Clozapine	125	0.91 (0.89–0.93)	0.99 (0.95–1.03)
	2000	0.95 (0.94–0.96)	0.97 (0.95–0.99)

Table 6. Quantitative matrix factor mean (range) based on both peak area and analyte: internal standard response ratio.

The method demonstrated good linearity for all analytes as shown in the Table 7 below. This was evaluated when different ratios of low and high concentration pools of the analytes were combined and analyzed. In addition, calibration lines exhibited coefficient of determination (r^2) of >0.995 for all analytes.

Analyte	Linear range (ng/mL)
Nortriptyline	10–576
Protriptyline	10–576
Normaprotiline	10–1300
Nordoxepin	10–650
Desipramine	5–650
Norclozapine	25–2881
Amitriptyline	5–650
Maprotiline	5–576
Doxepin	5–650
Nortrimipramine	5–576
Imipramine	5–650
Trimipramine	10–1300
Norclomipramine	5–650
Clomipramine	10–500
Clozapine	25–2497

Table 7. Linearity range summary of TCAs.

Accuracy was evaluated using external quality assurance (EQA) serum samples provided by LGC (Greater London, UK) for all analytes with the exception of protriptyline which was not included in the scheme. The results obtained were compared to the LC-MS method mean for the samples. Bland-Altman agreement (Figure 4) provided a mean method bias of <14.75% demonstrating very good agreement with the EQA LC-MS mean values for the analytes evaluated except for norclomipramine. A summary of results is shown in Table 8.

Analyte	Scheme range (ng/mL)	Number of samples analyzed	Mean % bias from scheme LC-MS
Nortriptyline	28–670	13	2.4
Normaprotiline	9.8–294.83	8	12.5
Nordoxepin	7–470.9	6	0.7
Desipramine	34–648	12	2.4
Norclozapine	122–3500	11	14.7
Amitriptyline	26–522	13	7.7
Maprotiline	22.19–545.13	8	7.6
Doxepin	15–565	6	1.4
Nortrimipramine	8–560	6	6.8
Imipramine	29–550.2	12	4.5
Trimipramine	13–570	6	9.5
Clomipramine	23–631	14	3.7
Clozapine	104–3000	11	11.8

Table 8. Accuracy summary.




Figure 4. Comparison of Waters LC-MS/MS method to EQA scheme MS method mean for using Bland-Altman fit for a) Nordoxepin, b) Nortrimipramine, c) Desipramine, d) Clomipramine.

Conclusion

- The analysis of 15 tricyclic antidepressants with different polarities was achieved by a selective and robust UHPLC-MS/MS clinical research method and used the XSelect Premier HSS Technology to separate isobaric compounds
- The sample preparation and analysis was simple and inexpensive, requiring only 50 µL of plasma and taking less than 4.5 minutes per injection
- The method showed excellent precision, no significant carryover or matrix effects, and accuracy was confirmed by the analysis of EQA samples

References

1. Gillman PK, Tricyclic antidepressant pharmacology and therapeutic drug interactions updated, *British Journal of Pharmacology* (2007) 151, 737–748.

Featured Products

[ACQUITY UPLC I-Class PLUS System <https://www.waters.com/134613317>](https://www.waters.com/134613317)

[Xevo TQ-S micro Triple Quadrupole Mass Spectrometry <https://www.waters.com/134798856>](https://www.waters.com/134798856)

[MassLynx MS Software <https://www.waters.com/513662>](https://www.waters.com/513662)

[MassLynx Quantitation Applications <https://www.waters.com/513791>](https://www.waters.com/513791)

720008285, March 2024



© 2024 Waters Corporation. All Rights Reserved.

[Conditions d'utilisation](#) [Politique de confidentialité](#) [Marques](#) [Carrières](#) [Mentions légales et déclaration de confidentialité](#) [Cookies](#) [Préférences de cookies](#)