

Nota de aplicación

## Expediting Method Development for Oligonucleotide Impurity Analysis Using the ACQUITY™ QDa™ II Mass Detector

---

Xiangsha Du, Robert E. Birdsall, Karen Nyholm

Waters Corporation, United States

*Published on May 02, 2025*

---

### Abstract

Therapeutic oligonucleotides such as antisense oligos, siRNAs, aptamers, and mRNA, continue to grow in clinical relevance, creating a need for highly sensitive and reliable analytical tools for impurity profiling. During the synthetic manufacturing process, incomplete coupling can occur, resulting in truncated species or shortmers. Although traditional LC-UV workflows are commonly used for routine analysis, they often fall short in delivering the mass specificity required to confidently identify closely eluting impurities, such as n-1, n-2, or other sequence variants.

In this study, it is demonstrated how users can effectively separate shortmer impurities from the parent full length product during a ion-pairing concentration and gradient delivery using the ACQUITY™ Premier™ QSM LC System. To streamline method development, the Waters ACQUITY QDa II Mass Detector, a compact mass detector, was utilized to confirm impurity identity in mass profiling, for more confidence in assessing oligonucleotide quality.

## Benefits

- Quaternary solvent blending speeds up oligonucleotide method development by enabling seamless, real-time optimization of ion-pairing conditions, with no downtime
- The ACQUITY QDa II Mass Detector provides identification by mass as well as quantitation of impurities associated with oligonucleotides, a critical step when UV detection alone is insufficient

---

## Introduction

The primary structures of oligonucleotides consist of a chain of nucleotide bases - adenine (A), thymine (T), cytosine (C), guanine (G), or uracil (U), connected by a sugar and phosphate diester backbone (Figure 1A). As with other biopharmaceutical drug products, it is essential to monitor product/process related impurities to support patient safety and therapeutic efficacy. Although the synthesis process is well controlled, impurities such as shortmers (n-x) and longmers (n+x) commonly occur in oligonucleotide therapeutics due to incomplete protection/deprotection or polishing steps during the synthesis process.<sup>1</sup> Furthermore, chemical modifications such as phosphorothioate (PS) substitution, which replaces an oxygen with a sulfur on the phosphate group to improve delivery and resistance to nuclease degradation – can increase the complexity of the impurity profile by introducing diastereomers and additional synthesis impurities.

Ion-pairing reversed-phase liquid chromatography (IP-RPLC) is the most widely used technique for oligonucleotide separation.<sup>3</sup> It leverages alkyl-amines to form ion pairs with the negatively charged phosphate backbone, allowing resolution based on oligo length. However, n-1 and n+1 variants, differing by a single nucleotide from the full-length product (FLP), are particularly challenging to resolve and detect via UV alone, due to their similar physicochemical properties and potential co-elution. While LC-UV detection is a mainstay in quality control (QC) for its robustness and sensitivity, developing effective separation methods can be time-consuming requiring iterative optimization of ion-pairing agents, mobile phase composition, and gradient conditions.

To help accelerate this process, the Waters ACQUITY Premier LC System with quaternary solvent blending capability enables rapid method optimization. When paired with the ACQUITY QDa II Mass Detector, analysts can integrate an inline, compact MS solution for deeper insights. Analysts can gain orthogonal mass-based insights to complement UV detection, streamlining impurity identification and method development.<sup>4,5</sup>

This study demonstrates how this LC-UV/MS setup enables faster, more confident optimization for oligonucleotide impurity analysis. Two representative full-length products, FLP-PO (phosphodiester) and FLP-PS (phosphorothioate) (Figure 1B), were used to showcase the platform's capabilities.




Figure 1. A) Primary structure of oligonucleotides. B) Sequence of 25 mer full-length products with phosphate (FLP-PO) and phosphorothioate (FLP-PS) backbones and their corresponding n-1 shortmer.



Figure 2. The ACQUITY™ QDa™ II Mass Detector.

---

## Experimental

LC-MS grade water, methanol and acetonitrile solvents were purchased from Honeywell®. Dibutylamine (DBA) and triethylamine (TEA) were purchased from Sigma-Aldrich®. 1,1,1,3,3,3-Hexafluoro-2-propanol (HFIP) was purchased from Honeywell-Fluka®.

FLP-PO and FLP-PS and their relative (n-1) shortmer samples, were purchased from integrated DNA Technologies® (IDT). FLP-PO and FLP-PS were prepared at a concentration of 0.1 mg/mL in water. (n-1) shortmer was spiked in at 10%.

### LC System Conditions

LC system:	ACQUITY Premier System (QSM variant)
Detection:	TUV, Ti FC, $\lambda=260$ nm, 5 mm flow cell
Column:	ACQUITY™ Premier™ Oligonucleotide BEH™ C <sub>18</sub> Column, 130 Å, 2.5 $\mu$ m, 4.6 x 150 mm (p/n: 186009903)
Column temperature:	70 °C
Sample temperature:	10 °C
Injection volume:	5 $\mu$ L
Flow rate:	0.200 mL/min
Mobile phase:	A: water B: methanol/acetonitrile = 50/50 C: 20 mM DBA/150 mM HFIP (pH 8.5) or 40 mM TEA/300 mM HFIP (pH 8.5)

Chromatography software:

Empower™ 3.8.1

## Gradient Table

### PO backbone

Time (min)	Flow (mL/min)	%A	%B	%C	%D	Curve
initial	0.650	67.5	20.0	12.5	0.0	initial
2.00	0.650	67.5	20.0	12.5	0.0	6
25.00	0.650	53.5	34.0	12.5	0.0	6
26.00	0.650	10.0	90.0	0.0	0.0	6
37.00	0.650	10.0	90.0	0.0	0.0	6
38.00	0.650	67.5	20.0	12.5	0.0	6
48.00	0.650	67.5	20.0	12.5	0.0	6

### PS backbone

Time (min)	Flow (mL/min)	%A	%B	%C	%D	Curve
initial	0.650	54.5	8.0	37.5	0.0	initial
2.00	0.650	54.5	8.0	37.5	0.0	6
25.00	0.650	51.5	11.0	37.5	0.0	6
26.00	0.650	10.0	90.0	0.0	0.0	6
37.00	0.650	10.0	90.0	0.0	0.0	6
38.00	0.650	54.5	8.0	37.5	0.0	6
48.00	0.650	54.5	8.0	37.5	0.0	6

## MS Detector Settings

MS detector:

ACQUITY QDa II Mass Detector

Scan mode:

Negative Electrospray (ES-)

Data collection mode:	Continuum
Scan rate:	5 Hz
Scan range:	350–1500 <i>m/z</i>
Probe temperature:	600 °C
Capillary voltage:	1.5 kV
Cone voltage:	20 V

## Selected Ion Recording (SIR)

PO sample	<i>m/z</i>
Full length	1230.8
n-T	1180.1
n-C	1182.6
n-A	1178.6
n-G	1175.9
PS sample	<i>m/z</i>
Full length	1295.1
n-T	1241.5
n-C	1244.2
n-A	1240.2
n-G	1237.5
PO impurity	1292.3

---

## Results and Discussion

---

LC-UV/MS methods are frequently used for comprehensive oligonucleotide analysis, however method optimization can be time-consuming and labor-intensive because both the chromatography and MS require optimization with ion-pairing selection and gradient to facilitate adequate separation of impurities while maintaining good MS-response for identification. In this study, method development was streamlined by integrating LC-UV/MS workflow using a quaternary pump system configuration and a compact mass detector.

## Optimization for Phosphodiester Oligonucleotides (FLP-PO)

As shown in Figure 3, the Empower 3 Software consolidates instrument control for all modules within the instrument method settings. Users have access to all four mobile phase reservoirs of the ACQUITY Premier QSM System allowing for quaternary solvent blending for enhanced flexibility and efficiency in method optimization. To demonstrate the utility of quaternary solvent blending, water and organic solvents (ACN/MeOH = 50/50) were configured in mobile phases A and B. A concentrated ion-pairing stock solution of 20 mM DBA: 150 mM HFIP (pH 8.5) was prepared as mobile phase C. Mass detection was performed using the ACQUITY QDa II Mass Detector (Figure 3B), which offers a full scan channel for analytes up to 1500  $m/z$ . Additionally, it can monitor associated impurity ions using SIR for increased specificity and sensitivity. In this study, users can understand how the lower charge state  $[M-6H]^{-6}$   $m/z$  ions were selected for monitoring the corresponding FLP and n-1 impurities to minimize ion overlap for increased accuracy.




Figure 3. Empower™ instrument method: A) QSM method settings, B) QDa™ II method settings.

## Impact of Ion-Pairing Concentration

A mixture of FLP-PO oligonucleotides with shortmer (n-T) was spiked at 10% for optimization of detector response and separation efficiency. A consistent 20-34% organic gradient was used to evaluate DBA concentrations using a single mobile phase preparation by altering the composition percent (%) of reservoir line C. As shown in Figure 3 where line C at 12.5% represents 2.5 mM of ion-pairing in the mobile phase. Using this approach to screen IP concentration, as illustrated in Figure 4A, DBA concentrations between 1 mM and 10 mM achieved baseline resolution with a USP resolution greater than 1.5. Resolution was observed to increase with elevated DBA concentrations (Figure 4B), however, the ion pairing reagent impacted the ionization efficiency of oligonucleotides, leading to significant suppression of the MS signal as shown in DBA concentrations at 5 mM and above (Figure 4C). Therefore, a concentration of 2.5 mM DBA emerged as the optimal balance - showing sufficient resolution while minimizing MS signal suppression.




Figure 4. A) DBA/HFIP concentration screening for the separation of FLP-PO from the n-T shortmer. Resolution and signal over noise ratio from UV spectra (B) and peak area of SIR channels (C) with the increasing of DBA concentration.

For quality control and regulatory compliance, it is crucial to assess the robustness of the optimized method. In this instance, mass load impact on peak resolution was evaluated for the optimized method. As shown in Figure 5A and 5B, the method scaled linearly and was able to maintain the desired peak resolution with increasing mass load. This demonstrates that the optimized method is reliable and can be consistently reproduced across development and manufacturing processes, ensuring best practices. For thorough identification of shortmer

impurities, the optimized method was verified using a combination of FLP-PO oligonucleotides spiked with all possible n-1 sequences, including n-A, n-C, n-G, and n-T. As shown in Figure 5C, the ACQUITY QDa II Mass Detector, which is configured in-line, enables quick confirmation that the individual n-1 species are completely resolved from the FLP.




Figure 5. A) Linear fitting of FLP-PO and n-T shortmer peak area with increased mass load according to SIR channels. B) Resolution and %area of n-T with increase mass load from UV spectra. C) UV and SIR spectra of all n-1 spiked into FLP-PO oligonucleotides.

## Impurity Profiling of FLP-PO with LC-UV/MS

After confirming the method's accuracy using spike-in standards, the optimized method was applied to the full-length product (FLP) to analyze intrinsic impurities. The non-spike-in FLP-PO profile was analyzed using 2.5 mM DBA (Figure 6). The SIR peaks of each n-1 impurity were integrated, yielding an overall area percentage of 4.10%, consistent with the 4.19% impurity peak integrated from the UV spectrum (Figure 6A, B). The consistency between SIRs and UV spectra highlights the transferability of this robust, optimized method to routine QC environment analysis.

In addition to SIR monitoring, a full scan of the analyte was acquired from 350 to 1500  $m/z$ . Extraction of total ion chromatography (TIC) and the comparison of the charge table in Figure 6C to the extracted spectrum shown in






Figure 7. A) TEA/HFIP concentration screening for the separation of FLP-PS from n-T and PO impurities. Resolution and signal-to-noise ratio from UV spectra (B) and peak area of SIR channels (C) with the increasing of TEA concentration.

\*A temperature pre-screening was performed at 45 °C, 60 °C, and 70 °C, with 70 °C brings the best separation. All data presented in this application note were collected at 70 °C.

As indicated in Figure 8, a gradient for 8–11% offered optimal resolution of PO impurities without overly extending analysis time and lowering overall sensitivity through band broadening. Longmer (n+1) impurities and the acrylonitrile impurities were confirmed to be separated from the FLP under increased mass load (Figure 8B). The optimized conditions were similarly evaluated using a spike-in panel of n-1 impurities with the ACQUITY QDa II Mass Detector in-line. As shown in Figure 9, the impurity levels are in good agreement between the UV and SIR channels in the spike-in sample. Mass detection enhances specificity, revealing that the PO impurity causes co-elution and shouldering in the UV response. This detailed understanding allows for more accurate PO quantification and aids future method improvement.




Figure 8. A) Gradient screening for the separation of FLP-PS from major impurities. B) Impurities profiling with increased mass load.




Figure 9. UV and SIR spectra of all n-1 spiked into FLP-PS oligonucleotides.

## Impurity Profiling of FLP-PS with LC-UV/MS

This is shown in the analysis of the stock FLP-PS sample. In Figure 10A, the UV spectrum reveals a broad impurity peak (14.56% of the total area). With the increased specificity of the ACQUITY QDa II Mass Detector in SIR mode, individual impurities are distinguished for more accurate assessment. Figure 10B shows that the peak is mainly comprised of oxidized PO impurity (11.15%) and n-C shortmer (2.55%), with other impurities at negligible levels. Full scan data (Figure 10D inset) confirms no ion overlap between the PO impurity (1291.9  $m/z$ ) and FLP ion (1294.7  $m/z$ ), as well as for n-C shortmer. This data demonstrates the ACQUITY QDa II Mass Detector's ability to identify PO impurities that would be obscured in UV detection, improving impurity quantification. The integration of quaternary solvent blending and MS offers an efficient method development platform for consistently monitoring oligonucleotide impurities in manufacturing environments.




Figure 10. A, B) UV and SIR spectra of FLP-PS oligonucleotides. C) Charge states of FLP-PS, n-1 shortmer PS and PO impurities. D) TIC extraction from 17.6 to 18.4 min.

## Conclusion

In summary, this integrated LC-UV/MS workflow tackles the common challenges users may face in oligonucleotide method development - particularly the need for faster optimization, confident analyte identification, and robust impurity profiling. The ACQUITY Premier LC System with quaternary solvent capabilities, coupled with the ACQUITY QDa II Mass Detector, streamlines method development by enabling both UV and mass-based detection, reducing setup time and analytical complexity.

With extended mass range and sensitivity, the workflow described in this application note enables more comprehensive analysis of larger oligonucleotide species and related impurities, helping scientists meet stringent regulatory and QC requirements more efficiently. Additionally, this approach addresses a growing need for scalable, reproducible workflows that can be readily adopted in routine quality control environments.

## References

1. N. M. El Zahar, N. Magdy, A. M. El-Kosasy, M. G. Bartlett. Chromatographic approaches for the characterization and quality control of therapeutic oligonucleotide impurities. *Biomedical Chromatography*. 2018, 32, e4088.
2. F. Eckstein. Phosphorothioates, essential components of therapeutic oligonucleotides. *Nucleic Acid Ther.* 2014, 24 (6), 374–387.
3. S. M. McCarthy, M. Gilar. Hexylammonium Acetate as an Ion-Pairing Agent for IP-RP LC Analysis of Oligonucleotides. Waters Application Note. [720003361](#). 2016.
4. R.E. Birdsall, X. Du, K. Nyholm. Extending the Analytics of Biopharmaceutical QA/QC Labs with the ACQUITY™ QDa™ II Mass Detector. Waters Application Note. [720008386](#). 2024.
5. P. Bigos, X. Du, R.E. Birdsall, K. Nyholm. Advancing Sensitivity and Efficiency in Released N-Linked Glycan Analysis With the ACQUITY™ QDa™ II Mass Detector. Waters Application Note. [720008631](#). 2024.
6. Z. Kadlecová, K. Kalíková, E. Eva Tesařová, M. Gilar. Phosphorothioate oligonucleotides separation in ion-pairing reversed-phase liquid chromatography: effect of ion-pairing system. *J. Chromatogr. A*. 2022, 1676, 463201.
7. K.J. Fountain, M. Gilar, J.C. Gebler. Analysis of native and chemically modified oligonucleotides by tandem ion-pair reversed-phase high-performance liquid chromatography/electrospray ionization mass spectrometry. *Rapid Commun. Mass Spectrom.* 2003, 17, 646–653.

Honeywell and Honeywell-Fluka are registered trademarks of Honeywell International Inc. <

<https://www.honeywell.com/global/terms-and-conditions>>

Sigma-Aldrich is a registered trademark of Merck KGaA, Darmstadt, Germany. <

[https://www.sigmaaldrich.com/global/life-science/legal/trademarks?srsId=AfmBOopQyJaPjNgJwWdfWqrZyrM2\\_yyL\\_VIJCzuFdDmjltm-yJ4o8A6o](https://www.sigmaaldrich.com/global/life-science/legal/trademarks?srsId=AfmBOopQyJaPjNgJwWdfWqrZyrM2_yyL_VIJCzuFdDmjltm-yJ4o8A6o)>

Integrated DNA Technologies is a registered trademark of Integrated DNA Technologies, Inc. <

<https://www.idtdna.com/page/about>>

---

## Featured Products

---

[ACQUITY Premier System <](#)

<https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/acquity-premier-system.html>>

[Empower Chromatography Data Software \(CDS\) <](#)

<https://www.waters.com/nextgen/global/products/informatics-and-software/chromatography-software/empower-software-solutions/empower-cds.html>>

[ACQUITY QDa II Mass Detector <<https://www.waters.com/nextgen/global/products/mass-spectrometry/mass-spectrometry-systems/acquity-qda-ii-mass-detector.html>](#)>

[ACQUITY UPLC and ACQUITY Premier Tunable UV Detectors <](#)

<https://www.waters.com/nextgen/global/products/chromatography/chromatography-detectors/acquity-uplc-and-acquity-premier-tunable-uv-detectors.html>>

720008796, April 2025



© 2025 Waters Corporation. All Rights Reserved.

[Condiciones de uso](#) [Aviso de privacidad de Waters](#) [Marcas comerciales](#) [Empleo](#) [Avisos legales y de privacidad](#) [Cookies](#) [Preferencias de cookies](#)