

Application Note

UHPLC-MS/MS Analysis of Amino Acids in Dried Blood Spots using Waters Kairos Amino Acid Kit for Clinical Research

Rachel McBrinn, Gareth Hammond, Padhraic Rossiter, Lisa J. Calton

Waters Corporation, United States

Published on June 05, 2025

Contact Sales

For research use only. Not for use in diagnostic procedures.

Abstract

This study details the quantification of 42 amino acids in dried blood spot (DBS) matrices using the Kairos™ Amino Acid Kit and AccQ•Tag™ Ultra “3X” Derivatizing Reagent. Analysis was performed with an ACQUITY™ UPLC™ I-Class PLUS System and a Xevo™ TQ-S micro Triple Quadrupole Mass Spectrometer. Method performance was evaluated using Kairos Calibrators and Quality Controls, external Centers for Disease Control and Prevention (CDC) DBS materials, and in-house Calibrator and Quality Control DBS materials.

The AccQ•Tag Ultra “3X” Reagent enabled accurate and precise quantification of amino acids in DBS samples,

with a run time of 9 minutes per sample. This method provides clinical researchers with a simple and rapid analysis technique, eliminating the need for mobile-phase buffers or ion-pair reagents.

Benefits

- Analytical chromatographic method providing separation of isobaric species in <10 minutes
- 350 µL plate format
- Buffers and ion-pairing agents are not required

Introduction

Quantifying amino acids from DBS using optical detection typically requires complex mixtures of buffers and/or ion-pair reagents, along with lengthy analytical run times. This study demonstrates the utility of the RUO Kairos Amino Acid Kit for analyzing 42 amino acids in DBS using LC-MS-based technologies.

For comprehensive instructions on using the Kairos Amino Acid Kit (p/n: [176004379](https://www.waters.com/nextgen/global/shop/application-kits/176004379-kairos-amino-acid-500-manual-first-time.html) < <https://www.waters.com/nextgen/global/shop/application-kits/176004379-kairos-amino-acid-500-manual-first-time.html>>), refer to the Care and Use Manual (p/n: [720006448](https://help.waters.com/help/global/support/library-details.html?documentid=720006448EN) < <https://help.waters.com/help/global/support/library-details.html?documentid=720006448EN>>), which details sample preparation, analytical methodology, and performance characteristics for solvent Calibrators and Quality Controls. This study summarizes the method performance of Waters in-house DBS Quality Controls prepared with the Kairos Amino Acid Kit and analyzed using the ACQUITY UPLC I-Class PLUS System and Xevo TQ-S micro Mass Spectrometer.

Both Waters in-house and external CDC DBS materials were extracted and incubated with the Kairos Internal Standard Solution. Extracts were transferred to a 96-well collection plate containing borate buffer and derivatized using the AccQ•Tag Ultra “3X” Derivatizing Reagent. A 2 µL sample extract was injected and quantified using the Kairos Amino Acid Kit Calibrators.

UHPLC-MS/MS analysis successfully quantified 37 amino acids extracted from a single DBS for clinical research, with the remaining 5 reported for semi-quantitative purposes only. Chromatographic separation was performed using an ACQUITY UPLC I-Class PLUS System and a CORTECS™ UPLC C₁₈ Column, followed by detection on a Xevo TQ-S micro Mass Spectrometer (Figure 1).



Figure 1. The ACQUITY UPLC I-Class PLUS System with Xevo TQ-S micro Mass Spectrometer.

Experimental

Sample Description

Reagent Kit

Refer to Kairos Amino Acid Kit (p/n: 176004379 <<https://www.waters.com/nextgen/global/shop/application-kits/176004379-kairos-amino-acid-500-manual-first-time.html>>) and Care and Use Manual (p/n: 720006448 <<https://help.waters.com/help/global/support/library-details.html?documentid=720006448EN>>) for reconstitution of Kairos Amino Acid Kit Calibrator and Quality Control Sets, Kairos Amino Acid Kit Reagents, Borate Buffer, and AccQ•Tag Ultra “3X” Derivatization Reagent.

The Kairos Internal Standard (IS) was reconstituted using 2 mL of 80:20 MeOH:H₂O (80:20 methanol/water v/v) and mixed at room temperature for 10 minutes, ensuring all material was fully dissolved. The contents of the vial were transferred to a volumetric flask and made up to 50 mL using 80:20 MeOH/H₂O.

Calibrators and QC Sample Preparation

Calibrator and QC samples were prepared fresh on the day of analysis.

- Step 1. Add 75 μ L of Kairos IS to a 1.5 mL microfuge
- Step 2. To this add 20 μ L of Calibrator/QC sample
- Step 3. Vortex for 5 seconds
- Step 4. Leave samples aside until required (Step 5 below)

DBS Sample Preparation

DBS samples were extracted fresh on the day of analysis.

- Step 1. Punch a 3.2 mm DBS disc into 350 μ L collection plate
- Step 2. Add 75 μ L of Kairos IS to disc
- Step 3. Seal plate with foil adhesive and incubate at room temperature for 20 minutes, 500 rpm
- Step 4. Add 70 μ L of borate buffer containing 0.5 M NaOH to a 350 μ L collection plate
- Step 5. Transfer 10 μ L of incubated DBS extract/Calibrator/QCs into borate buffer*
- Step 6. Pipette to mix
- Step 7. Vortex AccQ•Tag Ultra "3X" Reagent and add 20 μ L to each sample
- Step 8. Seal plate with silicone cap mat
- Step 9. Allow plate to rest at room temperature for 1 minute
- Step 10. Incubate sample plate for 10 minutes, 55 °C, 500rpm
- Step 11. Inject 2 μ L on UHPLC-MS/MS System

*Note: *On transfer, the DBS sample extracts are translucent and light yellow in the borate solution, Calibrators and QCs extracts are colorless.*

Method Conditions

The LC Conditions, Gradient Table, MS Conditions and MRM Parameters are detailed in the Kairos Amino Acid Kit Care and Use Manual (p/n: 720006448 <<https://help.waters.com/help/global/support/library-details.html?documentid=720006448EN>>).

Data Management and Processing

MS software:	MassLynx™ v4.2 Software (SCN 1042)
Informatics:	Informatics: TargetLynx™ XS v4.2 Application Manager waters_connect™ Software with QUAN Review Application
Data processing:	TargetLynx Software (User factor of 6.45 must be applied for quantitation of DBS samples only).

Results and Discussion

Kairos Amino Acid Kit Performance

No significant carryover was observed from the Kairos Amino Acid Kit Calibrator 6 into subsequent blank samples for all amino acids (detector response $\leq 20\%$ of LOQ) and no functionally significant carryover was observed on in-house prepared QC High DBS extracted samples into subsequent Endogenous DBS extracts.

Matrix effects were assessed on six independent sources of contrived DBS endogenous samples incubated in triplicate using the Kairos IS Solution. The IS peak area was compared to IS peak area not exposed to DBS matrix no significant ion enhancement or ($< \pm 15\%$ compared to control, %CV $< 10\%$) was observed, except for histidine.

Linearity was assessed using the Kairos Amino Acid Kit Calibrators on one day (n=4). Regression analysis demonstrated a linear fit using 1/x weighting across the measuring range listed for the 42 amino acids (Appendix Table 4). For each amino acid, the coefficient of determination (R^2) was > 0.99 and the %imprecision for each calibrator was $\leq 15\%$, except for at the lower limit of the measuring interval (LLMI) $\leq 20\%$.

Analytical sensitivity was assessed by analyzing samples prepared in 0.1 M HCl over the range of 0.5–20 μM , replicates of ten per level across five analytical runs. The results are presented in Appendix Table 3. The LOQ for

the amino acids analyzed were less than Calibrator 1 measured values except for histidine (9 μM), anserine (20 μM), 3-methyl histidine (10 μM), 1-methyl-histidine (10 μM), and argininosuccinic acid (20 μM).

Kairos Amino Acid Kit QC solvent samples (QC LOW and QC HIGH) precision and accuracy were assessed in replicates of five over five analytical runs. Average within day, total precision performance (%CV) and % Accuracy to assigned values were calculated for each analyte. The Kairos QC within-day precision was ≤ 11.5 %CV and total precision was ≤ 14.0 %CV except for anserine, 3-methyl-histidine and argininosuccinic acid which exceeded precision goals of ≤ 15 %CV. The %Accuracy to QC measured values were within $\pm 15\%$ of the assigned values. See Appendix Table 5 for the full summary.

In-House DBS Amino Acid Precision, Accuracy, and Recovery

In-House DBS QCs were prepared (55% Hematocrit) at 5 concentrations (QCs A-E) covering the Kairos Amino Acid Kit measuring range. The in-house DBS QCs were assessed for precision, accuracy, and recovery performance, 5 replicates per QC across five analytical runs ($n=25$). The average within-day precision and total precision was ≤ 15 %CV except for histidine, anserine and 3-methyl-histidine (total precision ≤ 18.6 %CV). %Accuracy to assigned values for all amino acids were $< 15\%$.

Recovery was assessed on QC B-E by deducting the endogenous QC A calculated concentrations and results compared to the spiked amino acid amount. Each amino acid recovered within 40–140%, the average recovery ranged from 65.5–135.2%. See Table 1 for the full summary of the data.

A single 3.2 mm DBS punch contains approximately 3.1 μL of blood and a user factor of 6.45 is applied in TargetLynx Software to correct the measurements of amino acids from extracted DBS when quantified using the solvent Kairos Amino Acid Calibrators as described above.

Analyte	In-house DBS				
	Measuring range assessed (μM)	Within-day precision (%CV)	Total precision (%CV)	%Accuracy to assigned values	% Recovery
Alanine	123.3–1171.1	5.1	5.5	100.4	107.1
Glycine	160.6–1745.6	5.0	5.7	102.7	115.3
Isoleucine	16.2–918.6	5.4	6.8	98.9	97.2
Leucine	13.6–1023.8	6.3	6.6	99.1	102.6
Phenylalanine	25.9–689.2	4.8	5.2	98.0	99.1
Valine	14.7–839	6.0	6.7	100.0	104.1
Serine	50.7–993.3	6.1	7.2	100.0	103.4
Threonine	31.4–696.9	4.9	5.7	101.4	96.1
Tyrosine	12.4–850.7	4.8	6.4	97.9	97.9
Glutamine	130.2–1919.8	4.5	5.3	102.8	121.2
Arginine	76.7–716.3	8.2	10.0	99.8	98.0
Lysine	34.0–748.5	5.1	6.4	104.7	99.7
Proline	36.7–1044.3	5.2	6.3	102.4	106.2
Aspartic acid	83.6–787.7	4.7	7.3	100.9	100.2
Glutamic acid	128.6–1120.5	4.7	5.4	102.6	108.8
Histidine	35.4–764.6	9.6	15.8	98.2	78.9
Asparagine	27.2–777.5	4.3	5.5	100.6	107.8
Methionine	5.6–674.9	5.4	7.0	100.7	93.4
Tryptophan	8.4–839.9	3.7	4.3	101.2	102.7
Hydroxylysine	57.0–455.7	4.7	8.4	104.8	68.4
Sarcosine	5.8–728.2	5.2	10.6	97.6	104.2
Beta alanine	13.3–623.2	5.9	9.0	99.3	107.6
α-aminobutyric acid	10.7–847.1	5.5	12.8	113.3	117.6
β-aminoisobutyric acid	19.7–745.9	4.6	8.8	95.2	91.4
γ-aminobutyric acid	12.5–1033.2	4.5	11.5	98.5	135.2
Aminoadipic acid	156.6–1179.4	4.9	14.1	106.8	115.1
Kynurenine	73.4–584.2	5.0	6.3	98.7	68.5
Homocitrulline	83.5–613.9	4.9	5.9	96.1	97.4
Citrulline	11.5–785.8	7.5	13.7	104.4	108.7
Taurine	174.9–690.7	10.4	11.8	102.8	122.1
Allo-isoleucine	86.8–724.6	4.9	5.9	97.0	99.0
Phosphoethanolamine	98.7–777.9	7.4	11.6	107.2	89.6
Glycyl proline	88.0–662.6	5.1	7.9	93.2	107.6
Anserine	47.8–498.3	10.7	18.6	96.3	65.6
L-Ornithine	40.3–565.9	6.2	9.4	105.7	67.3
Ethanolamine	109.7–775.3	6.5	7.5	98.4	105.7
Hydroxyproline	133.8–916.1	5.7	7.4	103.9	122.6
Cystathionine	47.4–261.5	5.0	8.8	97.9	78.2
3-Methyl-histidine	76.1–672.2	11.9	15.7	105.4	83.7
1-Methyl-histidine	70.9–565.9	9.7	10.7	98.6	78.7
Carnosine	84.7–609.9	7.2	10.5	103.7	112.7
Argininosuccinic acid	58.8–316.2	8.0	13.3	108.8	96.2

Table 1. Precision, accuracy, and recovery of in-house DBS QC materials using the Kairos Amino Acid Kit.

To evaluate method performance, external DBS material from the CDC were assessed for precision and accuracy (5 replicates over 5 analytical runs, n=25) to CDC reported values. Within-day precision ≤6.8%, total precision ≤14.4% and %Accuracy to CDC measurement concentrations were within ±15%. See Table 2 for the full summary.

CDC Quality control material			
Analyte	Within-day precision (%CV)	Total precision (%CV)	%Accuracy to CDC reported
Alanine	3.6	4.3	113.1
Glycine	3.8	4.5	104.7
Isoleucine	3.9	4.1	111.2
Leucine	3.5	3.8	101.7
Phenylalanine	3.4	3.8	96.1
Valine	3.5	4.2	107.9
Tyrosine	3.4	3.8	108.0
Arginine	6.8	8.2	96.8
Methionine	3.2	5.0	108.1
Allo-Isoleucine	4.0	5.1	94.4
Ornithine	6.1	9.9	87.8
Citrulline	4.4	14.4	99.6

Table 2. Precision and Accuracy of 12 Amino Acids from CDC QC material using the Kairos Amino Acid Kit.

Figure 2 shows chromatographic traces of isobaric leucines (leucine, isoleucine, and allo-isoleucine) in the external DBS material covering 4 QC levels, Endogenous (Red trace), 100 μ M (Green trace), 200 μ M (Purple trace), and 400 μ M (Black Trace). Figure 3 shows the Total Ion Chromatogram (TIC) for the 12 amino acids present in the external DBS material.




Figure 2. Detection and separation of isobaric Leucines in CDC QC material. Red trace = endogenous DBS; Green, Purple and Black traces represent 100, 200, and 400 μ M of Leucines spiked DBS, respectively.




Figure 3. TIC of 12 Amino Acids present in CDC QC Material.

Conclusion

The Kairos Amino Acid Kit enables clinical researchers to achieve reliable results with a single, flexible kit that

exhibits no significant carryover or matrix effects, and demonstrates good linearity, analytical sensitivity, precision, and accuracy. This method allows for the quantification of 42 amino acids in a single DBS in under 10 minutes using the ACQUITY UPLC I-Class System and Xevo TQ-S micro Mass Spectrometer.

The method eliminates the need for mobile-phase buffers or ion-pair reagents. With the derivatized samples remaining stable. The chromatographic conditions effectively separate isobaric amino acids, ensuring confident peak identification.

The clinical research method demonstrated good precision, accuracy, and recovery for 37 amino acids assessed in the in-house DBS material. The remaining five analytes; anserine, histidine, 1 and 3-methyl histidine and argininosuccinic acid are included; however, they did not fulfill all method performance criteria.

References

Appendix

Analyte	LOQ (µM)	LOQ DBS (µM) (x6.45)	Precision (%CV)	Accuracy (%Bias)	Signal-to-noise (PtP)
Alanine	0.6	4.1	9.5%	-1.7%	18.7
Glycine	1.0	6.4	13.3%	-1.9%	179.3
Isoleucine	0.5	3.4	3.9%	-3.2%	192.9
Leucine	0.5	3.1	3.3%	-1.6%	119.1
Phenylalanine	0.5	3.0	3.2%	-0.9%	250.9
Valine	0.5	3.3	6.5%	-0.9%	58.8
Serine	0.7	4.3	18.8%	-2.4%	101.6
Threonine	0.7	4.4	16.2%	-0.6%	32.4
Tyrosine	1	6.9	5.9%	11.2%	177.8
Glutamine	0.7	4.5	6.1%	0.7%	18.9
Arginine	4	24.1	13.0%	12.9%	36650.7
Lysine	0.5	3.2	14.2%	0.0%	11.9
Proline	0.7	4.2	12.4%	-7.9%	25.6
Aspartic acid	2	11.1	18.0%	-14.5%	81.7
Glutamic acid	0.6	4.1	4.4%	-2.4%	16.6
Histidine	9	60.7	19.0%	1.9%	204.1
Asparagine	0.6	3.9	5.4%	0.0%	386.1
Methionine	0.5	3.4	9.3%	9.3%	555.8
Tryptophan	0.5	3.4	2.2%	2.2%	516.6
Hydroxylysine	0.5	3.5	8.7%	-2.9%	30.3
Sarcosine	0.6	3.7	7.1%	-1.8%	21.3
Beta alanine	0.6	4.1	10.1%	-5.2%	110.5
α-Aminobutyric acid	0.6	3.6	3.7%	-1.1%	82210.8
α-Aminoadipic acid	0.6	4.1	5.5%	-1.5%	25.1
Kynurenine	0.4	2.5	18.4%	-0.6%	17758.3
Homocitrulline	0.4	2.9	12.5%	0.1%	26293.9
Taurine	2	10.6	15.5%	6.6%	87891.3
Allo- Isoleucine	0.5	3.2	4.6%	-3.1%	343.4
Phosphoethanolamine	0.6	3.9	15.3%	3.0%	16675.8
Anserine	20	129.6	14.6%	-6.4%	41059.4
γ-Aminobutyric acid	0.8	5.4	5.9%	-1.9%	65577.9
β-Aminobutyric acid	0.8	4.9	7.3%	0.2%	8673.5
Citrulline	0.6	3.8	13.4%	1.1%	13.8
Glycyl proline	0.6	3.6	12.8%	-4.5%	13122.3
Ornithine	2.0	12.9	17.9%	-2.3%	18826.5
Ethanolamine	0.5	3.1	11.1%	3.3%	151987.6
Hydroxyproline	0.7	4.2	10.5%	-2.1%	46.3
Cystathionine	2	13.9	12.9%	-3.4%	14.7
3-Methylhistidine	10	64.5	18.7%	-2.1%	31636.5
1-Methylhistidine	10	62.0	17.6%	11.9%	27036.9
Carnosine	3	22.1	14.9%	7.8%	13125.4
Argininosuccinic acid	20	131.5	14.6%	-1.2%	12055.7

Table 3. Analytical Sensitivity of the Kairos Amino Acid Kit in solution using the ACQUITY UPLC I-Class PLUS System and Xevo TQ-S micro Mass Spectrometer.

Kairos amino acid calibrators assigned values versus measured					
Analyte	r ²	Slope	Measuring range (μM)	Imprecision (%CV)	Max non-Linearity (%ADL)
Alanine	1.000	0.992	4.0–4019.0	≤5.5	8.8
Glycine	1.000	0.991	3.4–1036.9	≤7.2	8.2
Isoleucine	1.000	0.983	4.6–1012.2	≤1.4	15.1
Leucine	1.000	1.004	4.7–3983.2	≤0.9	11.9
Phenylalanine	1.000	0.994	4.4–3959.8	≤1.1	10.9
Valine	1.000	0.982	4.7–3963.0	≤4.5	7.9
Serine	1.000	1.007	4.9–968.2	≤5.1	4.6
Threonine	1.000	0.997	4.7–3601.8	≤4.0	7.5
Tyrosine	1.000	0.998	4.6–3767.7	≤3.4	4.4
Glutamine	1.000	0.977	3.8–1084.8	≤4.6	8.8
Arginine	0.996	0.993	3.8–1113.4	≤17.8*	-15.2
Lysine	1.000	1.010	3.5–928.8	≤11.3	-12.4
Proline	0.999	0.988	3.4–1007.3	≤12.7	-16.8
Aspartic acid	1.000	1.008	3.6–972.2	≤14.2	6.6
Glutamic acid	0.999	1.027	3.1–991.6	≤3.9	9.5
Histidine	0.999	0.986	5.1–499.0	≤19.5*	-10.9
Asparagine	1.000	0.996	3.9–1044.9	≤4.2	-5.3
Methionine	1.000	0.999	5.1–1104.6	≤4.5	7.9
Tryptophan	1.000	1.001	3.9–980.3	≤0.9	5.8
Hydroxylysine	0.999	1.022	2.1–905.2	≤7.5	-6.8
Sarcosine	0.997	0.998	3.8–903.9	≤4.6	15.9
Beta alanine	0.999	1.023	3.1–823.2	≤3.1	-9.3
α-aminobutyric acid	0.999	0.980	5–1112.28	≤4.5	-10.5
β-aminoisobutyric acid	0.997	1.052	5–500	≤3.3	9.8
γ-aminobutyric acid	0.998	1.029	5–1000	≤4.4	-12.3
Aminoadipic acid	0.999	1.025	3–897.2	≤5.6	7.3
Kynurenine	0.999	0.981	4.5–464.5	≤5.6	12.1
Homocitrulline	0.999	0.974	4.3–1074	≤3.1	7.8
Citrulline	0.999	1.007	5–1000	≤8.0	6.2
Taurine	0.998	1.0473	2.7–867.6	≤8.7	17.5
Allo-isoleucine	1.000	0.9925	4.6–1094.7	≤1.6	8.3
Phosphoethanolamine	0.999	1.0076	5.2–1026.3	≤5.2	3.3
Glycyl proline	0.999	1.0196	5–1000	≤11.3	5.1
Anserine	0.999	0.9878	15.2–980.5	≤20.4*	-13.2
L-Ornithine	0.993	1.0902	5–500	≤6.6	18.9
Ethanolamine	1.000	0.9891	5–1000	≤6.6	14.5
Hydroxyproline	0.998	1.0139	5–1000	≤6.2	8.7
Cystathionine	0.998	1.0358	5–500	≤10.0	-14.3
3-Methyl-histidine	0.992	1.0403	5–1000	≤14.4	10.2
1-Methyl-histidine	0.999	0.9349	20–1000	≤21.1*	8.4
Carnosine	0.991	1.1097	5–1000	≤12.3	18.1
Argininosuccinic acid	1.000	0.8641	20–1000	≤9.8	-8.0

Table 4. Linearity of the Kairos Amino Acid Kit in solution.

*Denotes maximum %CV at LLMI.

Analyte	QC Low kairos quality controls		QC High kairos quality controls		Average %Accuracy to assigned
	Within-day precision (%CV)	Total precision (%CV)	Within-day precision (%CV)	Total precision (%CV)	
Alanine	1.7	4.4	1.9	6.3	100.6
Glycine	2.6	6.1	2.1	6.4	97.2
Isoleucine	1.2	5.4	4.4	7.9	107.5
Leucine	0.8	1.2	0.7	4.9	110.2
Phenylalanine	0.6	1.8	0.3	4.5	105.3
Valine	1.4	1.8	1.5	4.9	108.4
Serine	2.1	3.2	2.0	5.3	101.5
Threonine	1.5	2.3	2.4	5.9	93.4
Tyrosine	1.6	3.3	1.2	4.2	98.3
Glutamine	2.4	3.3	1.8	3.8	98.6
Arginine	4.4	7.0	3.6	5.9	105.4
Lysine	2.4	2.4	1.9	5.9	93.9
Proline	2.1	4.1	3.0	9.7	87.0
Aspartic acid	2.5	3.6	1.5	6.7	94.1
Glutamic acid	1.7	5.4	1.7	6.5	114.5
Histidine	7.9	13.1	4.9	9.8	100.5
Asparagine	2.7	2.8	2.3	4.2	103.1
Methionine	1.8	1.8	1.1	5.9	99.7
Tryptophan	0.7	1.3	0.5	4.7	93.9
Hydroxylysine	2.8	5.5	1.5	11.2	105.6
Sarcosine	2.8	6.3	2.2	3.9	97.4
Beta alanine	2.8	4.9	2.2	5.9	90.2
α -aminobutyric acid	3.1	7.6	2.7	5.7	118.1
β -aminoisobutyric acid	4.1	5.5	4.1	7.1	114.5
γ -aminobutyric acid	3.4	3.8	3.6	5.4	99.3
Aminoadipic Acid	2.5	11.9	1.8	14.0	114.5
Kynurenine	2.5	5.2	3.6	8.5	96.6
Homocitrulline	2.8	3.7	2.3	7.1	103.1
Citrulline	2.4	7.4	4.4	13.3	99.4
Taurine	2.5	7.0	4.0	4.0	118.3
Allo-isoleucine	1.3	3.4	2.1	4.9	112.1
Phosphoethanolamine	4.8	5.6	4.1	5.8	105.1
Glycyl proline	3.2	7.2	2.6	11.3	97.6
Anserine	11.5	14.9	7.6	17.4	104.8
L-Ornithine	5.0	5.4	7.2	9.4	90.9
Ethanolamine	2.2	3.0	2.2	5.6	104.6
Hydroxyproline	3.8	4.8	3.9	10.5	99.7
Cystathionine	4.1	4.3	6.2	9.3	90.9
3-Methyl-histidine	9.8	16.6	6.5	13.9	100.4
1-Methyl-histidine	11.5	13.5	6.7	11.1	94.4
Carnosine	7.8	9.5	4.3	6.9	101.4
Argininosuccinic acid	9.2	31.3	8.5	33.0	100.2

Table 5. Precision and Accuracy of the Kairos Amino Acid Kit QCs in solution.

Kairos, AccQ•Tag, ACQUITY, UPLC, Xevo, CORTECS, MassLynx, TargetLynx, and waters_connect are trademarks

of Waters Technologies Corporation.

Featured Products

[ACQUITY UPLC I-Class PLUS System <](#)

<https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/acquity-uplc-i-class-plus-system.html>>

[Xevo TQ-S micro Triple Quadrupole Mass Spectrometer <](#)

<https://www.waters.com/nextgen/global/products/mass-spectrometry/mass-spectrometry-systems/xevo-tq-s-micro.html>>

[Kairos Amino Acid Kit <](#)<https://www.waters.com/nextgen/global/products/application-kits/kairos-amino-acid-kit.html>>

[ACQUITY Premier Columns <](#)<https://www.waters.com/nextgen/global/products/columns/acquity-premier-columns.html>>

[MassLynx Mass Spectrometry Software <](#)<https://www.waters.com/nextgen/global/products/informatics-and-software/mass-spectrometry-software/masslynx-mass-spectrometry-software.html>>

[waters_connect for Quantitation <](#)https://www.waters.com/nextgen/global/products/informatics-and-software/waters_connect-for-quantitation.html>

720008827, May 2025



© 2025 Waters Corporation. All Rights Reserved.

[Termos de uso](#) [Aviso de Privacidade](#) [Marcas comerciais](#) [Carreiras](#) [Avisos jurídicos e de privacidade](#) [Cookies](#) [Preferências de cookies](#)