

Application Note

Analysis of Organic Carbonates in Lithium-Ion Battery Electrolyte by High-Performance Liquid Chromatography (HPLC) with Mass Detection (MS)

Margaret Maziarz, Stephanie Harden, Paul Rainville

Waters Corporation, United States

Published on January 15, 2026

Abstract

Lithium-ion battery electrolyte is a solution composed of a lithium salt dissolved in organic carbonates. The degradation of these carbonates is a key factor affecting battery efficiency and performance. This application note presents an HPLC-MS method for the analysis of common organic carbonates and carbonate-based aging products in electrolyte of the lithium-ion battery. The five carbonate solvents selected for the study included ethylene carbonate (EC), propylene carbonate (PC), dimethyl carbonate (DMC), ethyl methyl carbonate (EMC) and diethyl carbonate (DEC). Two aging products were tested, including dimethyl 2,5-dioxahexanedioate (DMDOHC) and diethyl 2,5-dioxahexanedioate (DEDOHC). Chromatographic separation was performed on an XSelect™ Premier HSS PFP Column, operated under reversed-phase conditions, with analysis carried out on an Arc™ Premier System coupled to an ACQUITY™ QDa™ II Mass Detector. The quantification limits for the carbonate solvents and aging products achieved using positive electrospray ionization and selected ion recording (SIR) acquisition mode ranged from 0.05 to 1 ppm and 0.00075 to 0.0025 ppm, respectively. Additionally, the analysis

of lithium hexafluorophosphate (LiPF₆)-based electrolyte solution was demonstrated.

Benefits

- Sensitive method for low-level detection and quantification of carbonate solvents (0.05–1 ppm) and carbonate-based aging products (0.00075–0.0025 ppm) in lithium-ion battery electrolyte solutions using the ACQUITY QDa II Mass Detector with SIR acquisition mode.
- Robust chromatographic separation using the XSelect Premier HSS PFP Column under 7 minutes analysis time, operated under reversed-phase conditions.

Introduction

The electrolyte is one of the main components in the lithium-ion battery.^{1,2} It is a conductive solution that transports lithium ions from the cathode to the anode during the charging process. During discharge, the lithium ions flow from the anode to the cathode. The electrolyte solution is composed of lithium salt dissolved in a mixture of organic solvents.^{1,2} The organic solvents are typically a mixture of cyclic and linear carbonates, including EC, PC, DMC, EMC, and DEC.² Lithium hexafluorophosphate (LiPF₆) is commonly used as the conducting lithium salt,² providing the necessary ionic conductivity for efficient battery operation.

The LiPF₆ salt in carbonate-based systems can exhibit chemical and thermal instability, which can impact the performance of the lithium-ion batteries.²⁻⁴ Electrolyte solutions can also degrade, forming aging products including DMDOHC and DEDOHC.² Various techniques are employed to investigate quality and properties of the electrolytes. Techniques such as scanning calorimetry (DSC) and thermogravimetric analysis (TGA) offer a comprehensive analysis of an electrolyte's thermal properties, while rheology measures viscosity of the conductivity solutions.¹ Additionally, chromatographic methods are used to measure the quantity of carbonate solvents in electrolyte solutions.²⁻⁴ Gas chromatography mass spectrometry (GC/MS) and liquid chromatography mass spectrometry (LC/MS) are common techniques used to investigate carbonate-based electrolytes.³

In this application note, an HPLC method with mass detection was developed for the analysis of five common carbonate solvents and two carbonate-based aging products (Table 1). The method employed an ACQUITY QDa II Mass Detector coupled with an Arc Premier System. Method performance characteristics including limit of detection (LOD), limit of quantification (LOQ), linearity and system suitability were demonstrated. Additionally,

the developed method was applied for the analysis of LiPF₆-based electrolyte solution.

Experimental

Materials

Methanol (Optima LC/MS) was purchased from Fisher Chemicals (catalog number A456-4). Formic acid (LC/MS) was purchased from Fisher Chemicals (catalog number 85178).

Organic carbonate solutions, carbonate-based aging products, and battery grade lithium hexafluorophosphate electrolyte were obtained from Sigma-Aldrich.

Sample Description

Standard Solutions

Individual stock solutions containing each carbonate solution and carbonate-based aging products were prepared at 10,000 ppm in methanol and used to make a mixture standard solution. Mixture standard solution was serially diluted with methanol to prepare LOD, LOQ, and linearity solutions.

Concentration in ppm represents a volume of analytes to the volume of solution (v/v).

Lithium Hexafluorophosphate Sample

Commercially available LiPF₆ salt electrolyte solution in EC/DMC/DEC=1:1:1 (v/v/v) at 1.0 M was diluted with methanol at 1/5000 (v/v).

Method Conditions

System:	Arc Premier System, Column Manager with active pre-heating, PDA and the ACQUITY QDa II Mass Detector
Column:	XSelect Premier HSS PFP Column, 2.5 µm, 4.6 mm x 100 mm (Waters, p/n: 186010051)

Column temperature:	40 °C
Sample temperature:	10 °C
Flow rate:	1.0 mL/min
Injection volume:	5.0 µL
Vials:	LC-MS Maximum Recovery Vials 2 mL volume, p/n: 600000670CV
Mobile phase:	A: 0.1% formic acid in water B: 0.1% formic acid in methanol
Wash solvents:	Purge/Sample Wash: 60:40 water/methanol Seal Wash: 90:10 water/acetonitrile
Separation:	Described in the gradient table below

Gradient Table

Time (min)	Flow (mL/min)	%A	%B	Curve
Initial	1.0	95.0	5.0	6
7.00	1.0	5.0	95.0	6
8.00	1.0	5.0	95.0	6
8.10	1.0	95.0	5.0	6
12.00	1.0	95.0	5.0	6

MS Conditions

MS system:	ACQUITY QDa II Mass Detector
Ionization mode:	Electrospray ionization in positive mode (ESI+)
Acquisition:	MS Scan: 35–300 Daltons SIR mode for quantitative analysis, as described in Table 1
Probe temperature:	600 °C
Capillary voltage:	0.5 kV
Data:	Centroid

Data Management

Chromatography data software (CDS):	Empower™ 3.6.1
-------------------------------------	----------------


Compound	Name	MM *	Ion	Selected Ion Recording (SIR)	Acquisition time (min)	Cone voltage (V)
Ethyl carbonate	EC	88.02	[M+H] ⁺	89.0	1.75–2.75	10
Propylene carbonate	PC	102.03	[M+H] ⁺	103.0	2.85–3.85	10
Dimethyl carbonate	DMC	90.08	[M+H] ⁺	91.0	3.40–4.40	10
Ethyl methyl carbonate	EMC	104.05	[M+H] ⁺	105.1	4.50–5.50	5
Dimethyl 2,5-Dioxahexanedioate	DMDOHC	178.05	[M+Na] ⁺	201.0	5.00–6.00	10
Diethyl carbonate	DEC	118.13	[M+H] ⁺	119.0	5.40–6.40	10
Diethyl 2,5-Dioxahexanedioate	DEDOHC	206.08	[M+Na] ⁺	229.1	6.10–7.0 0	10

*Table 1. Carbonate solvents and aging degradation products. Acquisition settings for ACQUITY QDa II Mass Detector. * MM: monoisotopic mass (Da).*

Results and Discussion

Chromatographic columns with a wide range of selectivity were explored during this study to ensure separation between carbonate solvents and carbonate-based aging products. The selected columns (XSelect Premier CSH™ C₁₈, XSelect Premier CSH Phenyl Hexyl, XSelect Premier HSS T3, XSelect Premier HSS PFP) were tested with acetonitrile and methanol solvents. The study showed that XSelect Premier HSS PFP Column with methanol provided the best retentivity and separation for all analytes (Figure 1). The method successfully separated carbonate solutions and carbonate-based aging products in less than a 7-minute gradient time. The MS total ion chromatogram (TIC) data was collected across the mass range of 35–300 Da using electrospray ionization in positive (ESI+) mode (Figure 1A). The mass spectral data facilitated quick and accurate identification of the components. The carbonate solvents formed $[M + H]^+$ molecular ions, while the selected aging products formed sodium adducts, $[M + Na]^+$. These ions were measured using a selected ion recording (SIR) acquisition mode, which determines intensity for a single ion of interest (Figure 1B).

A. MS TIC Data



B. MS SIR Data




Figure 1. Separation of carbonate solvents and carbonate-based aging products run on XSelect Premier HSS PFP Column. MS TIC with mass spectral data (A) and data with SIR acquisition mode (B). Sample contains: EC/PC at 100 ppm, DMC/EMC/DEC at 1,000 ppm, DMDOHC/DEDOHC at 10 ppm.

Optimization of Ionization Parameters

Ionization parameters, such as ionization mode, probe temperature and cone voltage are often optimized to enhance the MS signal and improve detection limits for target analytes. In this study, the impact of the ACQUITY QDa II Mass Detector ionization parameters on the sensitivity of carbonates solvents were evaluated. Probe temperature, cone voltage, and capillary voltage were systematically tested and examples of probe temperature and cone voltage optimization for EC, PC, DMC, and EMC are shown in Figures 2 and 3, with data acquired using SIR acquisition mode. A probe temperature of 600 °C produced the highest signal for carbonate solvents. Optimal cone voltages were 10 volts for EC, PC, DMC, and DMC, and 5 volts for EMC.




Figure 2. Effect of probe temperature on the MS signal of carbonate solvents, measured in SIR mode with the ACQUITY QDa II Mass Detector.




Figure 3. Effect of cone voltage on the MS signal of carbonate solvents, measured in SIR mode with the ACQUITY QDa II Mass Detector.

LOD and LOQ

LOD and LOQ were determined using the signal-to-noise (S/N) criteria of 3:1 and 10:1, respectively. The S/N values were calculated using a noise centered on the peak region in the blank injection with the half height multiplier for noise region set to five, as recommended by the United States Pharmacopeia (USP).⁵

Representative chromatograms of the LOQ solutions prepared in methanol are shown in Figure 4. Data from six

replicate injections was evaluated to establish and verify performance at the LOD and LOQ levels (Table 2). Excellent performance at the LOQ levels was achieved with the relative standard deviation (RSD) of the peak areas $\leq 6.81\%$. No internal standard was used in this work to correct for data variability.




Figure 4. Representative chromatograms of carbonate solvents and carbonate-based aging products at the LOQ level, acquired in SIR mode using the ACQUITY QDa II Mass Detector.

Name	SIR	LOD Conc. (ppm)	LOD S/N	LOD % RSD of peak areas	LOQ Conc (ppm)	LOQ S/N	LOQ % RSD of peak areas
EC	89.0	0.05	5	5.80	0.1	14	5.14
PC	103.1	0.02	6	4.70	0.05	13	4.47
DMC	91.0	0.5	5	7.14	1.0	10	4.20
EMC	105.1	0.5	5	11.23	1.0	10	5.33
DMDOHC	201.1	0.001	4	6.98	0.0025	10	4.23
DEC	119.0	0.2	6	10.04	0.5	12	4.98
DEDOHC	229.1	0.00025	4	7.20	0.00075	12	6.81

Table 2. LOD and LOQ results for carbonate solvents and carbonate-based aging products determined from 6 replicate injections acquired in SIR mode.

System Suitability

A system suitability test was performed to verify performance of the system and method. Data from 6 replicate injections of the standards solution containing carbonate solvents was used to determine %RSD of peak areas and retention times (Figure 5). The values for RSDs of peak areas were $\leq 1.7\%$.




Figure 5. System suitability results for carbonate solvents acquired in SIR mode. Chromatogram: EC/PC at 10 ppm and DMC/DEC/EMC at 100 ppm (A). Individual results are shown for EC (B), PC (C), DMC (D), EMC (E) and DEC (F).

Linearity

Method linearity for quantifying carbonate solvents over a concentration range of 0.05–50 ppm and 0.5–500 ppm for EP/PC and for DMC/EMC/DEC, respectively, was evaluated. The calibration curves exhibited a linear relationship between MS responses and concentration with correlation coefficients of $R^2 > 0.998$ using a linear fit (Table 3).

Name	SIR (<i>m/z</i>)	Linearity range (ppm)	Fit type	Weighing	Linearity fit (<i>R</i> ²)
EC	89.0	0.05–50	Linear	None	0.9986
PC	103.1	0.05–50	Linear	None	0.9981
DMC	91.0	0.5–500	Linear	None	0.9988
EMC	105.1	0.5–500	Linear	1/ <i>x</i>	0.9984
DEC	119.0	0.5–500	Linear	1/ <i>x</i>	0.9982

Table 3. Method linearity results for the carbonate solvents.

Electrolyte Solution

The developed method was applied for the analysis of commercially available electrolyte containing LiPF₆ salt in EC/DMC/DEC = 1:1:1 (v/v/v) at a 1.0 M concentration. Sample was diluted at 1/5000 (v/v) in methanol. The analysis confirmed the presence of carbonate solvents EC, DMC, and DEC, as well as the trace-level of a carbonate aging product DEDOHC (Figure 6).




Figure 6. Analysis of electrolyte containing LiPF₆ salt in EC/DMC/DEC = 1:1:1 (v/v/v). Standard solution at 10 ppm of EC/PC/DMC/EMC/DEC and 1 ppm of DMDOHC/DEDOHC for reference (A). Sample showed presence of EC/DMC/DEC and DEDOHC (B). SIR data acquired on the ACQUITY QDa II Mass Detector.

Conclusion

A sensitive and reliable HPLC-MS method using the Arc Premier System coupled with the ACQUITY QDa II Mass Detector was developed for the analysis of carbonate solvents and carbonate-based aging products in

lithium-ion battery electrolyte. This reversed-phase LC method successfully separated all five carbonate solvents (EC, PC, DMC, EMC, DEC) and two aging products (DMDOHC/DEDOHC) on an XSelect Premier HSS PFP Column. Low-level quantification limits were achieved using SIR acquisition mode ranging from 0.05 to 1 ppm and 0.00075–0.0025 ppm for carbonate solvents and aging products, respectively. The method was demonstrated to be applicable for the analysis of lithium-ion battery electrolyte.

References

1. Waters TA Instruments, [Lithium-Ion Battery Material Testing <https://www.tainstruments.com/applications/battery-material-testing/>](https://www.tainstruments.com/applications/battery-material-testing/) .
 2. Schultz C, Vedder S, Streipert B, Winter M, Nowak S. Quantitative investigation of the decomposition of organic lithium-Ion Battery Electrolytes with LC-MS/MS. *RSC Advances*, 2017, 7:27853–28762.
 3. Bhalkikar A, Marin CM, Cheung CL, Method development for Separating Organic Carbonates by Ion-Moderated High-Performance Liquid Chromatography. *Journal of Separation Science*, 2016, 39: 4484–4491.
 4. Fang C, Tran TH, Zhao Y, Liu G. Electrolyte Decomposition and Solid Electrolyte Interphase Revealed by Mass Spectrometry. *Electrochimica Acta*, 2021, 399: 139362.
 5. USP General Chapter <621>, Chromatography. United States Pharmacopoeia USP-NF 2025 Issue 3, Official 01-Dec-2024.
-

Featured Products

Arc Premier System <<https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/arc-premier-system.html>>

ACQUITY QDa II Mass Detector <<https://www.waters.com/nextgen/global/products/mass-spectrometry/mass-spectrometry-systems/acquity-qda-ii-mass-detector.html>>

Empower Chromatography Data Software (CDS) <

<https://www.waters.com/nextgen/global/products/informatics-and-software/chromatography->

[software/empower-software-solutions/empower-cds.html>](#)

720009211, January 2026



© 2026 Waters Corporation. All Rights Reserved.

[Termos de uso](#) [Aviso de Privacidade](#) [Marcas comerciais](#) [Carreiras](#) [Avisos jurídicos e de privacidade](#) [Cookies](#) [Preferências de cookies](#)