

Application Note

waters_connect™-Enabled LipidQuan™ Workflow: A Tandem Quadrupole Method for Rapid Quantification of Bioactive Lipid in Plasma

Aarav Batra, Robert Plumb, Nikunj Tanna

Waters Corporation, United States

Published on September 4, 2026

For research use only. Not for use in diagnostic procedures.

Main

For research use only. Not for use in diagnostic procedures.

Abstract

Targeted lipidomics using tandem quadrupole mass spectrometry provides accurate quantitative information changes in biofluids and tissues lipid concentrations due to disease, environmental exposure, genetics etc. LipidQuan is a hydrophilic interaction liquid chromatography with tandem mass spectrometry (HILIC-MS/MS) method for the quantification of over 400 bioactive lipids in biofluids. This application note describes the acquisition and analysis of lipid concentration data using Xevo™ TQ Absolute XR Tandem Quadrupole Mass Spectrometer and waters_connect for Quantitation Software.

Benefits

- Quantitative method for the measurement of bioactive lipids in plasma to support translational biomarker studies
- Over 440 lipids measured in one simple, rapid (8-minute) HILIC method for polar and non-polar lipids
- Simple, intuitive data processing with exception focused review in waters_connect for Quantitation Software
- Multiple lipid classes including mono-, di-, triglycerides, FFA's, cholesterol esters, ceramides, hexosyl ceramides, PG, PC, SM, LPC, LPE, PS, PA, PI, LPA, and LPI in a single LC-MS/MS method

Introduction

Lipids perform a key role in multiple biological processes including cell signaling, energy storage, cell membranes creation, hormone level regulation, vital organs protection, and facilitate absorption of fat-soluble vitamins. The quantification of lipid in biofluids provides information on system health, pathophysiological processes, and disease staging.^{1,2} LipidQuan is a validated method for the rapid quantification of over 400 bioactive lipids in biofluids, e.g., plasma and tissues³ via HILIC-MS/MS operated in multiple reaction monitoring (MRM) mode. This application note describes the acquisition and data analysis bioactive lipid in rat plasma using ACQUITY™ Premier UPLC™ System coupled to Xevo TQ Absolute XR Tandem Quadrupole Mass Spectrometer and waters_connect Quantitation Software.

Experimental

Sample Preparation

A calibration line was prepared using the Avanti EquiSPLASH™ mix over the range of 0.334-167 ng/mL in protein precipitation solution. Rat plasma was prepared by protein precipitation with a pre-cooled acetonitrile:isopropanol (ACN:IPA) (2:1) at 4 °C (1:5, plasma:ACN/IPA) containing EquiSPLASH at 200 ng/mL. Samples were vortex-mixed for 1 minute and centrifuged at maximum of 10,300 g for 6 minutes at 4 °C before transferring the supernatant to glass vials for LC-MS analysis.

LC-MS/MS

The samples were analyzed using via a rapid (8-minute) HILIC-MS/MS method using an ACQUITY Premier UPLC System coupled to a Xevo TQ Absolute XR Tandem Quadrupole Mass Spectrometer operated in MRM mode with either +ve or -ve ESI ionization. The MRM precursor and product ion pair transitions, cone voltages and collision energies are described in Monjuma, et al.³ A total of 431 transitions were acquired in +ve ESI and 446 transitions in -ve ESI. LC-MS/MS peak integration and lipid quantification was performed using the MS Quan Application in waters_connect for Quantitation Software.

LC Conditions

System:	ACQUITY Premier UPLC System
Column:	ACQUITY BEH™ Amide Column 1.7 µm, 2.1 × 100 mm
Column temperature:	45 °C
Flow rate:	0.6 mL/min
Mobile phase A:	95:5 Acetonitrile/water + 10 mM ammonium acetate
Mobile phase B:	50:50 Acetonitrile/water + 10 mM ammonium acetate
Gradient:	0.1% to 20.0% B for 2 minutes, then 20% to 80% B for 3 minutes followed by 3 minutes re-equilibration
Run time:	8 minutes
Injection volume:	1 µL

MS Conditions

System:	Xevo TQ Absolute XR Tandem Quadrupole Mass Spectrometer
Ionization mode:	ESI (+/-)
Capillary voltage:	2.8 kV (+)/1.9 kV (-)
Acquisition mode:	MRM
Source temperature:	120 °C
Desolvation temperature:	500 °C
Cone gas flow:	150 L/hr
Desolvation flow:	1000 L/hr
Nebulizer gas:	7 bar
Ion guide offset 1:	3 V
Ion Guide offset 2:	0.3 V

Data Management

Chromatography software:	waters_connect Software
MS software:	waters_connect Software
Data processing:	MS Quan Application

Results and Discussion

Quantitative lipidomics can provide insight into relative and quantitative changes in lipid abundance due to phenotypic variation, e.g., age, gender, disease state, and environmental exposure. Analysis of large data sets require a high-throughput method. LipidQuan Workflow facilitates the analysis of polar and non-polar lipid classes biofluids by employing a HILIC-based lipid class separation and MRM MS quantification.^{3,4} The LipidQuan Methodology contains a total of 2,041 MRM transitions for 16 polar and non-polar lipid classes. The lipids were chromatographed using a rapid (8 minute) HILIC separation which facilitates classed-based separation based on the polar head groups. This facilitated classed-based quantification using the EquiSPASH mix (Avanti Research) as it contains stable isotopically labeled (SIL) for 13 lipid classes. Representative chromatograms of the +ve & -ve ESI analysis of control rat plasma are shown in Figure 1.




Figure 1. Positive (A) and negative (B) ion chromatograms of control rat plasma.

A)

Peak Detection

Showing 7 of 7 records

No.	Analyte Name	Group	Linked Internal Standard	Calibration Reference Compound	Expected Retention Time (min)	Add ...	Data Type	Precursor Mass (m/z)	Product Mass 1 (m/z)	Included Transitions	Wavelength (nm)
1	15:0-18:1 (d7) PE		None	None	1.41		MRM	709.50	288.30 (Quan)		
2	15:0-18:1 (d7) PG		None	None	0.81		MRM	740.60	288.30 (Quan)		
3	15:0-18:1 (d7) PI		None	None	2.11		MRM	828.60	288.30 (Quan)		
4	15:0-18:1 (d7) PS		None	None	2.03		MRM	753.54	288.30 (Quan)		

B)

Calibration Levels

No.	Analyte Name	Units	Reporting Limit	Concentration Factor	Level 1	Level 2	Level 3	Level 4	Level 5
1	15:0-18:1 (d7) DG	ng/mL			167	83.5	33.4	16.7	8.35
2	15:0-18:1 (d7)-15:0 TG	ng/mL			167	83.5	33.4	16.7	8.35
3	18:1 (d7) Chol Ester	ng/mL							
4	18:1 (d7) LPC	ng/mL							
5	18:1 (d7) LPE	ng/mL							
6	18:1 (d7) MG	ng/mL							
7	18:1 (d7) SM	ng/mL							

C)


Analyte Reference Compounds

Showing 265 of 265 records


No.	Analyte Name	Group	Linked Internal Standard	Calibration Reference Compound	Expected Retention Time (min)	Add ...	Data Type	Precursor Mass (m/z)	Product Mass 1 (m/z)	Product Mass 2 (m/z)	Product Mass 3 (m/z)
245	Stearoyl-L-carnitine		None	Stearoyl-L-carnitine	1.78		MRM	428.42	84.99 (Quan)		
246	Stearoyl-L-carnitine, C4OC		None	Stearoyl-L-carnitine	4.40		MRM	281.95	84.99 (Quan)		
247	Tetradecanoyl-L-carnitine		None	Stearoyl-L-carnitine	1.86		MRM	371.40	84.99 (Quan)		
248	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.36		MRM	824.77	101.05 (Quan)		
249	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.40		MRM	822.78	94.04 (Quan)	94.04	
250	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.56		MRM	852.80	101.05 (Quan)	101.05	579.53 (Quan)
251	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.40		MRM	850.79	101.05 (Quan)	101.05	
252	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.96		MRM	880.85	101.05 (Quan)	101.05	
253	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.40		MRM	878.82	101.05 (Quan)	101.05	379.53
254	TG16:0/16:0/16:0		None	15:0-18:1 (d7)-15:0 TG	0.39		MRM	876.80	101.05 (Quan)	101.05	379.53

Figure 3. MS Quan Application analysis method for +ve ESI A) Analyte peak detection parameters, B) Calibration levels for SIL EquiSPLASH mix, C) Analyte reference compounds for individual lipids.


18:1 (d7) LPE ES- $R^2 = 0.9974$ $Y = 1.3e5 \cdot X - 9.18e3$




18:1 (d7) LPE ES- $R^2 = 0.9974$ $Y = 1.3e5 \cdot X - 9.18e3$




Quality Controls




Quality Controls




15:0-18:1 (d7) PE $R^2 = 0.9985$ $Y = 2.78e4 \cdot X - 1.22e3$



18:1 (d7) LPC $R^2 = 0.9986$ $Y = 5.71e5 \cdot X + 2.58e5$ 3



Quality Controls



Quality Controls




Figure 4. Representative calibration of lipids 18:1 (d7) LPE, 18:1 (d9) SM, 15:0 18:1 (d7) PE, and 18:1 (d7) LPC in rat plasma extract over the concentration range 0.334–167 ng/mL.

Data Analysis

The large number of lipids monitored (>440), combined with the complex nature of lipids, the HILIC separation mechanism, and number of lipid isomers isobaric results in the close elution of lipid species. These factors make baseline allocation review a complex and time-consuming process. MS Quan Application offers a rapid approach to data analysis allowing the scientist to review the data in one simple interface, as shown in Figure 5. Using a review by exception approach, allows scientist to quickly check calibration, QCs, and blanks for the whole data

set or for individual compounds (Figure 5).




Figure 5. MS Quan Application dashboard and peak integration review.




Figure 6. Peak integration review and baseline modification.

Conclusions

The generation of accurate quantitative measurements of lipids provides insight into lipid dysregulation as a result of pathophysiological process, genetics, environmental exposure, and other factors. LipidQuan is a validated HILIC-MS/MS-based method for the quantification of over 400 bioactive lipids in biofluids and tissues. The Xevo TQ Absolute XR Tandem Quadrupole Mass Spectrometer combined with waters_connect

Software has been employed to streamline the acquisition and data analysis lipid concentrations in rat plasma. The high-speed MRM acquisition capability of the Xevo TQ Absolute XR Tandem Quadrupole Mass Spectrometer allowed for the collection of over 430 MRM transitions in just 5 minutes while the robustness of the StepWave XR Ion Guide ensured robust performance for the duration of the study. Data acquisition, peak integration, and lipid quantification was simplified by waters_connect for Quantitation Software and the MS Quan Application.

References

1. Perrotti, F.; et al. Advances in Lipidomics for Cancer Biomarkers Discovery. *Intl. J. Mol. Sci.* 2016, 17, 1992. DOI: [10.3390/ijms17121992](https://doi.org/10.3390/ijms17121992) <<https://doi.org/10.3390/ijms17121992>> .
2. Havulinna, A. S.; et al. Circulating Ceramides Predict Cardiovascular Outcomes in the Population-Based FINRISK 2002 Cohort. *Arterioscler. Thromb. Vasc. Biol.* 2016, 36, 2424–2430. DOI: [10.1161/ATVBAHA.116.307497](https://doi.org/10.1161/ATVBAHA.116.307497) <<https://doi.org/10.1161/ATVBAHA.116.307497>> .
3. Munjoma, N.; et al. High Throughput LC-MS Platform for Large Scale Screening of Bioactive Polar Lipids in Human Plasma and Serum. *J. Proteome Res.* 2022, 21, 2596–2608. DOI: [10.1021/acs.jproteome.2c00297](https://doi.org/10.1021/acs.jproteome.2c00297) <<https://doi.org/10.1021/acs.jproteome.2c00297>> .
4. Foddy, H.; et al. A 30,000 Injection Assessment of Xevo TQ Absolute XR Mass Spectrometer Quantitative Stability and Uptime for Analysis of Naltrexone for Clinical Research. Waters Application Note. [720009342](https://www.waters.com/nextgen/global/library/application-notes/2026/a-30000-injection-assessment-of-xevo-tq-absolute-xr-mass-spectrometer-quantitative-stability-and-uptime-for-analysis-of-naltrexone-for-clinical-research.html) <<https://www.waters.com/nextgen/global/library/application-notes/2026/a-30000-injection-assessment-of-xevo-tq-absolute-xr-mass-spectrometer-quantitative-stability-and-uptime-for-analysis-of-naltrexone-for-clinical-research.html>> . May, 2026.

Featured Products

Xevo TQ Absolute XR <<https://www.waters.com/nextgen/global/products/mass-spectrometry/mass-spectrometry-systems/xevo-tq-absolute.html>>

ACQUITY Premier System <

<https://www.waters.com/nextgen/global/products/chromatography/chromatography-systems/acquity-premier-system.html>>

waters_connect <<https://www.waters.com/nextgen/global/products/informatics-and->

[software/waters_connect/waters-connect-for-biopharmaceuticals.html>](#)

720009559, September 2026



© 2026 Waters Corporation. All Rights Reserved.

[Terms of Use](#) [Privacy Notice](#) [Trademarks](#) [Careers](#) [Legal and Privacy Notices](#) [Cookies](#)
[Cookie Preferences](#) [Do Not Sell My Personal Information](#)