

ACQUITY UPLC I-Class PLUS System (SM-FTN-I)

The Waters™ ACQUITY™ UPLC™ I-Class PLUS System's holistic design is targeted for investigative analysis where maximized peak capacity, throughput, and sensitivity are critical and is perfectly suited for running any MS-based applications. The system is comprised of a Binary Solvent Manager (BSM) and a Sample Manager with Flow-Through Needle (SM-FTN-I); this configuration offers excellent carryover performance.

ACQUITY UPLC I-CLASS PLUS SYSTEM FEATURES (WITH SM-FTN-I)

| | |
|---|---|
| Total system bandspread [†] 5 σ | $\leq 9 \mu\text{L}$ (default configuration) |
| Dwell volume (total system) [†] | $\leq 100 \mu\text{L}$ (default configuration) |
| Gradient delay volume [†] | $\leq 80 \mu\text{L}$ |
| Integrated leak management | Leak sensors, as standard, and safe leak handling |
| System synchronization | Injection synchronization between both pumps and the sample manager enhances retention time reproducibility |
| Operating flow rate range | 0.001 to 2.000 mL/min, in 0.001 mL increments (firmware version 1.71 and later) |
| Maximum operating pressure | 18,000 psi up to 1 mL/min, 12,000 psi up to 2 mL/min |
| pH range [†] | 1 to 12.5 |
| Unattended operation | Leak sensors, full 96-hour diagnostic data display through console software |
| Cycle time | $\leq 15 \text{ s}$ inject to inject, with load ahead enabled |

BINARY SOLVENT MANAGER (BSM)

| | |
|------------------------------------|--|
| Number of solvents | Up to four, in combination of two, A1 or A2 and B1 or B2 |
| Solvent conditioning | Integrated vacuum degassing, six lines with two allocated for the injector needlewash/purge solvents |
| Gradient formation | High pressure mixing, binary gradient |
| Gradient profiles | 11 gradient curves (including linear, step [2], concave [4], and convex [4]) |
| Primary check valves | Intelligent Intake Valves ($i^2\text{Valve}$) |
| Flow accuracy [†] | $\pm 1.0\%$ of set flow rate at 0.500 mL/min, as per Empower™ SystemsQT™ |
| Flow precision [†] | $\leq 0.075\%$ RSD or 0.01 min SD, (0.2 to 2.0 mL/min), whichever is greater using premixed solvent |
| Composition ripple [†] | $\leq 1.0 \text{ mAU}$ |
| Composition precision [†] | $\leq 0.15\%$ RSD or ± 0.01 min SD, whichever is greater |
| Composition accuracy [†] | $\pm 0.5\%$ absolute from 5% to 95%, 0.2 to 2.0 mL/min |

[INSTRUMENT SPECIFICATIONS]

| | |
|---------------------------------|--|
| Pressure pulsation [†] | ≤0.4% or 25 psi, whichever is greater |
| Compressibility compensation | Automatic, no user intervention required |
| Priming | Wet priming runs at a flow rate of 4 mL/min |
| Pump seal wash | Equipped with a programmable active wash system to flush the rear of the high pressure seals and the plungers |
| Flow ramping | Automatic |
| Primary wetted materials | 316L stainless steel, UHMWPE blend, MP35N, titanium alloy, gold, sapphire, ruby, zirconia, Nitronic 60, DLC, fluoropolymer, PEEK, PEEK blend |
| Mixing options | Standard: 50 µL Optional: 100 µL and 380 µL |

SAMPLE MANAGER-FTN (SM-FTN-I)

| | |
|--------------------------------------|--|
| Injection volume range | 0.1 to 10.0 µL as standard configuration Up to 1000.0 µL with optional extension loop |
| Accuracy | ±0.2 µL, measured by fluid weight removed from vial with 10.0 µL injections averaged over 20 injections using standard 100-µL syringe |
| Precision [†] | ≤0.25%, 5 to 50 µL |
| Linearity [†] | ≥0.999 |
| Maximum sample capacity | Any two of the following: <ul style="list-style-type: none">▪ 96 and 384 microtiter plates▪ 48 position 2.00-mL vial plates▪ 48 position 0.65-mL micro-centrifuge tube plates▪ 24 position 1.50-mL micro-centrifuge tube plates |
| Sample compartment temperature range | 4.0 to 40.0 °C, settable in 0.1 °C increments; maintains 19 °C below ambient with a tolerance range between -2 and +4 °C |
| Temperature accuracy | ±0.5 °C at sensor |
| Temperature stability | ±1.0 °C at sensor |
| Sample manager heat time | ≤30 min ambient-40 °C |
| Sample manager cool time | ≤60 min ambient-4 °C |
| Injection needle wash | Integrated, active, programmable |
| Minimum sample required | 3 µL residual, using Waters Total Recovery 2-mL Vials (zero offset) |
| Sample carryover [†] | ≤0.001% caffeine (UV) ≤0.001% sulphadimethoxine (MS) |
| Advanced sample manager capabilities | Auto-dilution and auto-addition |
| Primary wetted materials | 316L stainless steel, polyimide, PEEK blend, DLC, PPS |

[INSTRUMENT SPECIFICATIONS]

COLUMN HEATERS (CH-A)

| | |
|--|---|
| Column capacity | CH-A: Single column, up to 4.6 mm internal diameter (I.D.), up to 150 mm in length with filter or guard column. Mounting extends out for use with MS-based detector |
| Fittings | 18,000 psi, low dispersion, with reusable column inlet fittings |
| Column compartment temperature range | Settable from 20.0 to 90.0 °C, settable in 0.1 °C increments |
| Column compartment temperature accuracy | ±0.5 °C at sensor |
| Column compartment temperature stability | ±0.3 °C at sensor |
| Column compartment heat time | ≤15 min ambient-60 °C |
| Solvent conditioning | Active pre-heating as standard; passive pre-heating (for legacy method support) |
| Column tracking | eCord™ Technology column information management tracks and archives column usage history |

COLUMN MANAGEMENT (CM-A)

| | |
|---|---|
| Column capacity | CM-A: Two columns, as standard (maximum length of 150 mm with filter or guard column) up to 4.6 mm internal diameter (I.D.) |
| Switching valves | Two nine-port, eight-position valves (CM-A only); provides programmable access switching, waste and bypass positions for rapid solvent changeover |
| Column compartment(s) temperature range | 4.0 to 90.0 °C, settable in 0.1 °C increments; two independent heat/cool zones |
| Column compartment(s) temperature accuracy | ±0.5 °C at sensor |
| Column compartment(s) temperature stability | ±0.3 °C at sensor |
| Column compartment heat time | ≤15 min ambient -60 °C |
| Column compartment cool time | ≤15 min 60–20 °C |
| Solvent conditioning | Active pre-heating as standard; passive pre-heating (for legacy method support) |
| Fittings | 18,000 psi, low dispersion, with reusable column inlet fittings |
| Column tracking | eCord Technology column information management tracks and archives column usage history |
| 2D support | Optional |

[INSTRUMENT SPECIFICATIONS]

SAMPLE ORGANIZER

| | |
|--------------------------------------|--|
| Sample plate capacity | Sample plate capacity is configured based on the types and combinations of plates being used: <ul style="list-style-type: none">▪ Maximum of 19 standard microtiter plates, up to 15.5 mm high, or▪ Maximum of 9 intermediate height plates (or 2-mL vial holders), up to 40.0 mm high, or▪ Maximum of 6 deep well plates (or 4-mL vial holders), up to 47.0 mm high |
| Maximum sample capacity | Maximum of 7296 samples in nineteen 384-well plates |
| Sample compartment temperature range | 4.0 to 40.0 °C, settable in 0.1 °C increments with a tolerance range between -2 and +4 °C |
| Temperature accuracy | ±1 °C at the sensor |
| Temperature stability | ±1 °C at the sensor |

BASED INSTRUMENTAL CONTROL

| | |
|-------------------------|--|
| External control | Empower Software, MassLynx™ Software, UNIFI™ Scientific Information System, or standalone through console software |
| External communications | Ethernet interfacing via RJ45 connection to host PC |
| Event inputs/outputs | Rear panel contact closure and/or TTL inputs/outputs |
| Connections INSIGHT™ | Provides real-time monitoring and automatic notification of instrument performance and diagnostic information, allowing for quicker problem resolution |

ENVIRONMENTAL SPECIFICATIONS

| | |
|-----------------------------|----------------------------|
| Acoustic noise | ≤62 dBA, system |
| Humidity operating | 20% to 80%, non-condensing |
| Operating temperature range | 4 to 40 °C |

ELECTRICAL SPECIFICATIONS

| | |
|--------------------|---|
| Power requirements | 100 to 240 VAC |
| Line frequency | 50 to 60 Hz |
| Power consumption | BSM: 360 VAC FTN: 400 VAC CM-A: 400 VAC |

[INSTRUMENT SPECIFICATIONS]

PHYSICAL SPECIFICATIONS

ACQUITY UPLC I-Class PLUS System: Width: 34.3 cm (13.5 in.)

BSM, SM-FTN-I, and CH-A Height: 71.1 cm (28.0 in.)

Depth: 71.2 cm (28.0 in.)

ACQUITY UPLC I-Class PLUS System: Width: 34.3 cm (13.5 in.)

BSM, SM-FTN-I, and CM-A Height: 79.6 cm (31.4 in.)

Depth: 71.2 cm (28.0 in.)

Sample Organizer Width: 25.4 cm (10.0 in.)

Height: 96.5 cm (38.0 in.)

Depth: 71.1 cm (28.0 in.)

[†] For specific test conditions, contact your Waters sales representative.

Waters

THE SCIENCE OF WHAT'S POSSIBLE.™

Waters, The Science of What's Possible, ACQUITY, UPLC, eCord, Empower, MassLynx, SystemsQT, UNIFI, and Connections INSIGHT are trademarks of Waters Corporation. All other trademarks are the property of their respective owners.

©2020 Waters Corporation. Produced in the U.S.A. November 2020 720006149EN TC-PDF

Waters Corporation

34 Maple Street

Milford, MA 01757 U.S.A.

T: 1 508 478 2000

F: 1 508 872 1990

www.waters.com